

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-15077-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 30.06.2022

Date of issue 30.06.2022

Holder of certificate:

Kolb & Baumann GmbH & Co. KG
Daimlerstraße 24, 63741 Aschaffenburg

Calibration in the fields:

Dimensional quantities

Length

- Gauge blocks
- Length measuring instruments
- Length measuring devices ^{a)}
- Diameter
- Form error

^{a)} also on-site calibration

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.

The certificate together with the annex reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de/en/accredited-bodies-search.html>.

Abbreviations used: see last page

Page 1 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the accreditation certificate D-K-15077-01-00
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Length Gauge blocks made of steel according to DIN EN ISO 3650:1999	0.5 mm to 100 mm must be of the same nominal length	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ for the deviation f_o and f_u from the central length: $0.05 \mu\text{m}$	l = gauge block length Quality of the measuring surface according to the commitments in the Laboratory Quality Manual resp. in the calibration procedures
	0.5 mm to 100 mm Unusual nominal length using a laser interferometric probe, type SIOS LM 20, with a reference gauge block which nominal size does not differ more than 10 mm from the test specimen	Measurement of the deviations f_o and f_u from the central length by 5 points comparison measurement	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ for the deviation f_o and f_u from the central length: $0.05 \mu\text{m}$	
Gauge blocks made of ceramics or tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm must be of the same nominal length		For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$ for the deviation f_o and f_u from the central length: $0.05 \mu\text{m}$	
	0.5 mm to 100 mm Unusual nominal length using a laser interferometric probe, type SIOS LM 20, with a reference gauge block which nominal size does not differ more than 10 mm from the test specimen		For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$ for the deviation f_o and f_u from the central length: $0.05 \mu\text{m}$	
Gauge blocks made of steel according to DIN EN ISO 3650:1999	100 mm to 1000 mm must be of the same nominal length	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement Interferential measurement of the deviations f_o and f_u from the central length	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ for the deviation f_o and f_u from the central length: $0.05 \mu\text{m}$	

Date of issue: 30.06.2022

Valid from: 30.06.2022

Page 2 of 5

Annex to the accreditation certificate D-K-15077-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Gauge blocks made of steel or tungsten carbide according to DIN EN ISO 3650:1999	0.1 mm to < 0.5 mm must be of the same nominal length	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement For the smallest measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the mean size: 0.15 μm	
Pairs of gauge blocks made of steel or tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the difference of the central lengths for pairs of gauge blocks with the same nominal length, respectively of difference of the central lengths up to 10 μm Measurement of the deviations f_0 and f_n from the central length by 5 points comparison measurement	For the difference of the central length of the pairs: 0.03 μm For the deviation from the central length: 0.03 μm (only for the nominal values 1.005 mm and 1.01 mm) else 0.05 μm	
Step-height Step-height-standard consisting of a plane-parallel plate with wrung gauge blocks made of tungsten carbide according to DIN EN ISO 3650:1999	0 mm to 25 mm	QM-APA 8.5.1.001.019: 2021-11 Option 1: The step height h is determined from the difference between measurements for the centre length Option 2: The step height h is determined by absolute measurement using a laser interferometric probe, type SIOS LM 20, referring to the reference plane	For the step-height: 0.10 μm	Ensuring the wringing of the gauge blocks on the base plate is done by interferential measurement of parallelism of the steps relating to the base plate
Gauge block comparator for calibration of gauge blocks	0.5 mm to 100 mm	DKD-R 4-1:2018	0.03 μm + 0.002 · D	$D \leq 10 \mu\text{m}$, for nominal length difference

Date of issue: 30.06.2022

Valid from: 30.06.2022

Annex to the accreditation certificate D-K-15077-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks	
Micrometers	to 300 mm	VDI/VDE/DGQ 2618 Part 10.1:2001	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length	
	> 300 mm to 1000 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$		
Dial gauges	to 100 mm	VDI/VDE/DGQ/DKD 2618 Part 11.1:2021	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	over 30 mm in the horizontal position	
Dial gauges	to 3 mm	VDI/VDE/DGQ 2618 Part 11.2:2002	0.6 μm		
Lever gauges	to 1.6 mm	VDI/VDE/DGQ 2618 Part 11.3:2002	0.8 μm		
Digital indicator	bis 100 mm	VDI/VDE/DGQ/DKD 2618 Part 11.4:2020	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	larger 30 mm in horizontal position only	
Plug gauges Diameter	0.5 mm to 500 mm	VDI/VDE/DGQ 2618 Part 4.1:2006 Option 3 und 4	$2 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	d = measured diameter	
Ring gauges Diameter	2 mm to 250 mm				
Optical flats and optical parallels	\emptyset 10 mm to \emptyset 200 mm				
Length	0.5 mm to 100 mm	QM-APA 8.5.1.002.013 2020-07 comparison measurement	$0.1 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$	l = measured length with gauge block comparator	
deviation from parallelism	to 5 μm		0.05 μm		Parallelism measurement only up to \emptyset 80 mm
deviation from flatness	to 5 μm		QM-APA 8.5.1.002.010 2020-07 interferometric		
Plane-parallel length standards not in accordance with DIN EN ISO 3650:1999	(7 x 7) mm ² to (75 x 75) mm ² \emptyset 8 mm to \emptyset 100 mm			l = measured length with gauge block comparator	
Length	0.5 mm to 100 mm	QM-APA 8.5.1.002.013 2020-07 comparison measurement	$0.1 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$	Parallelism measurement only up to \emptyset 80 mm	
deviation from parallelism			0.05 μm		

Date of issue: 30.06.2022

Valid from: 30.06.2022

Page 4 of 5

- Translation -

Annex to the accreditation certificate D-K-15077-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Length Gauge block comparator for calibration of gauge blocks	0.5 mm to 100 mm	DKD-R 4-1:2018	$0.03 \mu\text{m} + 0.002 \cdot D$	$D \leq 10 \mu\text{m}$, for nominal length difference

Abbreviations used:

CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
DGQ	Deutsche Gesellschaft für Qualität e.V.
DKD	Deutscher Kalibrierdienst
DKD-R	Guideline of Deutschen Kalibrierdienstes (DKD), published by the Physikalisch-Technischen Bundesanstalt
QM-APA	In-house method of the Kolb & Baumann GmbH & Co. KG
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.