



## Deutsche Akkreditierungsstelle

### Annex to the Partial Accreditation Certificate D-K-15086-01-01 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 09.06.2023

Date of issue: 20.07.2023

This annex is a part of the accreditation certificate D-K-15086-01-00.

Holder of partial accreditation certificate:

**MeßTechnikNord GmbH  
Industriestraße 29, 22880 Wedel**

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

*This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.*

Abbreviations used: see last page

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

**Page 1 of 13**

Calibration in the fields:

**Electrical quantities**

**DC and low frequency quantities**

- **DC voltage**
- **AC voltage**
- **DC current**
- **AC current**
- **DC resistance**
- **AC/DC transfer**
- **Power factor**
- **Inductance**
- **Capacitance**
- **Phase angle**
- **High voltage quantities**

**Time and frequency**

- **Time interval**
- **Frequency**

**Dimensional quantities**

**Length**

- **Length measuring instruments**

Within the measurands/calibration items marked with \*) the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
DC voltage	0 V to 120 mV > 120 mV to 1 V 10 V > 1 V to 12 V > 12 V to 1000 V > 1000 V to 10 kV > 10 kV to 50 kV		$3 \cdot 10^{-6} \cdot U + 0,5 \mu\text{V}$ $4 \cdot 10^{-6}$ $0,4 \cdot 10^{-6}$ $2,5 \cdot 10^{-6}$ $3 \cdot 10^{-6}$ $0,1 \cdot 10^{-3}$ $0,5 \cdot 10^{-3}$	$U = \text{Measured value}$
DC current	1 $\mu\text{A}$ to 10 $\mu\text{A}$ > 10 $\mu\text{A}$ to 100 mA > 100 mA to 1 A > 1 A to 10 A > 10 A to 30 A > 30 A to 100 A		$10 \cdot 10^{-6}$ $4 \cdot 10^{-6}$ $7 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $40 \cdot 10^{-6}$ $60 \cdot 10^{-6}$	
Clamp-on current meter	0,01 A to 1000 A		$5 \cdot 10^{-3}$	with current coil
DC resistance	0,1 m $\Omega$ to 0,2 m $\Omega$ > 0,2 m $\Omega$ to 2 m $\Omega$ > 2 m $\Omega$ to 40 m $\Omega$ > 40 m $\Omega$ to 0,4 $\Omega$ > 0,4 $\Omega$ to 2 $\Omega$ > 2 $\Omega$ to 20 $\Omega$ > 20 $\Omega$ to 200 $\Omega$ > 200 $\Omega$ to 2 k $\Omega$ > 2 k $\Omega$ to 20 k $\Omega$ > 20 k $\Omega$ to 200 k $\Omega$ > 200 k $\Omega$ to 2 M $\Omega$ > 2 M $\Omega$ to 20 M $\Omega$ > 20 M $\Omega$ to 200 M $\Omega$ > 200 M $\Omega$ to 2 G $\Omega$		$40 \cdot 10^{-6} \cdot R + 2 \cdot 10^{-9} \Omega$ $30 \cdot 10^{-6} \cdot R + 2 \cdot 10^{-9} \Omega$ $5 \cdot 10^{-6} \cdot R + 10 \cdot 10^{-9} \Omega$ $4 \cdot 10^{-6} \cdot R + 80 \cdot 10^{-9} \Omega$ $3 \cdot 10^{-6} \cdot R + 0,8 \cdot 10^{-6} \Omega$ $3 \cdot 10^{-6} \cdot R + 15 \cdot 10^{-6} \Omega$ $1 \cdot 10^{-6} \cdot R + 25 \cdot 10^{-6} \Omega$ $1,3 \cdot 10^{-6} \cdot R + 0,22 \cdot 10^{-3} \Omega$ $1,1 \cdot 10^{-6} \cdot R + 2,5 \cdot 10^{-3} \Omega$ $1,2 \cdot 10^{-6} \cdot R + 22 \cdot 10^{-3} \Omega$ $2 \cdot 10^{-6} \cdot R + 1 \Omega$ $2 \cdot 10^{-6} \cdot R + 12 \Omega$ $10 \cdot 10^{-6} \cdot R + 60 \Omega$ $0,1 \cdot 10^{-3} \cdot R + 100 \text{ k}\Omega$	$R = \text{Measured value}$

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 3 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage AC/DC transfer	0,1 V; 0,2 V	10 Hz 20 Hz 30 Hz; 40 Hz; 500 Hz 1 kHz; 20 kHz 50 kHz 100 kHz 500 kHz 700 kHz 1 MHz	0,10 · 10 <sup>-3</sup> 80 · 10 <sup>-6</sup> 50 · 10 <sup>-6</sup> 50 · 10 <sup>-6</sup> 60 · 10 <sup>-6</sup> 80 · 10 <sup>-6</sup> 0,10 · 10 <sup>-3</sup> 0,15 · 10 <sup>-3</sup> 0,20 · 10 <sup>-3</sup>	AC/DC transfer at fixed measured voltages and frequencies
	0,3 V; 0,5 V	10 Hz 20 Hz 30 Hz; 40 Hz; 500 Hz 1 kHz; 20 kHz 50 kHz; 100 kHz 500 kHz 700 kHz 1 MHz	0,10 · 10 <sup>-3</sup> 60 · 10 <sup>-6</sup> 30 · 10 <sup>-6</sup> 30 · 10 <sup>-6</sup> 50 · 10 <sup>-6</sup> 80 · 10 <sup>-6</sup> 0,10 · 10 <sup>-3</sup> 0,15 · 10 <sup>-3</sup>	
	1 V	10 Hz 20 Hz; 30 Hz; 40 Hz; 500 Hz 1 kHz; 10 kHz 20 kHz 50 kHz 70 kHz; 100 kHz 200 kHz; 500 kHz 700 kHz; 1 MHz	30 · 10 <sup>-6</sup> 20 · 10 <sup>-6</sup> 20 · 10 <sup>-6</sup> 30 · 10 <sup>-6</sup> 40 · 10 <sup>-6</sup> 50 · 10 <sup>-6</sup> 90 · 10 <sup>-6</sup> 0,15 · 10 <sup>-3</sup>	
	2 V; 3 V; 4 V; 6 V	10 Hz 20 Hz; 30 Hz; 40 Hz 500 Hz; 1 kHz 10 kHz; 20 kHz 50 kHz 70 kHz; 100 kHz 200 kHz; 500 kHz 700 kHz; 1 MHz	30 · 10 <sup>-6</sup> 20 · 10 <sup>-6</sup> 20 · 10 <sup>-6</sup> 20 · 10 <sup>-6</sup> 30 · 10 <sup>-6</sup> 50 · 10 <sup>-6</sup> 90 · 10 <sup>-6</sup> 0,15 · 10 <sup>-3</sup>	

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 4 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage AC/DC transfer	10 V; 12 V	10 Hz	$30 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$20 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$20 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$20 \cdot 10^{-6}$	
		50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6}$	
		200 kHz	$90 \cdot 10^{-6}$	
		500 kHz	$0,12 \cdot 10^{-3}$	
		700 kHz	$0,15 \cdot 10^{-3}$	
		1 MHz	$0,20 \cdot 10^{-3}$	
	20 V	10 Hz	$30 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$20 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$20 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$20 \cdot 10^{-6}$	
		50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6}$	
		200 kHz	$90 \cdot 10^{-6}$	
		500 kHz	$0,15 \cdot 10^{-3}$	
		700 kHz	$0,20 \cdot 10^{-3}$	
		1 MHz	$0,30 \cdot 10^{-3}$	
	30 V; 40 V	10 Hz	$30 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$20 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$20 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$20 \cdot 10^{-6}$	
		50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6}$	
		200 kHz	$90 \cdot 10^{-6}$	
		500 kHz	$0,15 \cdot 10^{-3}$	
		10 Hz	$30 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$20 \cdot 10^{-6}$	
	60 V	500 Hz; 1 kHz	$20 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$20 \cdot 10^{-6}$	
		50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6}$	
		200 kHz	$90 \cdot 10^{-6}$	

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 5 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage AC/DC transfer	100 V; 120 V	10 Hz	$30 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$20 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$20 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$20 \cdot 10^{-6}$	
		50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6}$	
	200 V	10 Hz	$55 \cdot 10^{-6}$	
	300 V; 400 V; 500 V	20 Hz; 30 Hz; 40 Hz	$50 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$50 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 50 kHz	$50 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$70 \cdot 10^{-6}$	
		40 Hz	$50 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$50 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 50 kHz	$70 \cdot 10^{-6}$	
	600 V; 700 V	70 kHz	$90 \cdot 10^{-6}$	
		100 kHz	$0,15 \cdot 10^{-3}$	
		40 Hz	$70 \cdot 10^{-6}$	
		500 Hz; 1 kHz	$70 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$70 \cdot 10^{-6}$	
		40 Hz; 500 Hz	$0,10 \cdot 10^{-3}$	
	800 V	1 kHz; 10 kHz; 20 kHz	$0,10 \cdot 10^{-3}$	
	1000 V	40 Hz; 500 Hz	$0,10 \cdot 10^{-3}$	
		1 kHz; 10 kHz; 20 kHz	$0,10 \cdot 10^{-3}$	

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 6 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Permanent Laboratory**

Measurement quantity / Calibration item	Range	Calibration and Measurement Capabilities (CMC)					Remarks
		Measurement conditions / procedure		Expanded uncertainty of measurement			
AC voltage	1 mV to 2,2 mV	10 Hz	to 20 Hz	0,19 · 10 <sup>-3</sup> · U + 1,3 µV			with Fluke 5790A U = Measured value
		> 20 Hz	to 40 Hz	0,17 · 10 <sup>-3</sup> · U + 1,3 µV			
		> 40 Hz	to 20 kHz	0,12 · 10 <sup>-3</sup> · U + 1,3 µV			
		> 20 kHz	to 50 kHz	0,12 · 10 <sup>-3</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	0,18 · 10 <sup>-3</sup> · U + 2,5 µV			
	> 2,2 mV to 7 mV	10 Hz	to 20 Hz	0,12 · 10 <sup>-3</sup> · U + 1,3 µV			
		> 20 Hz	to 40 Hz	72 · 10 <sup>-6</sup> · U + 1,3 µV			
		> 40 Hz	to 20 kHz	70 · 10 <sup>-6</sup> · U + 1,3 µV			
		> 20 kHz	to 50 kHz	72 · 10 <sup>-6</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	83 · 10 <sup>-6</sup> · U + 2,5 µV			
	> 7 mV to 22 mV	10 Hz	to 20 Hz	91 · 10 <sup>-6</sup> · U + 1,3 µV			
		> 20 Hz	to 40 Hz	53 · 10 <sup>-6</sup> · U + 1,3 µV			
		> 40 Hz	to 20 kHz	47 · 10 <sup>-6</sup> · U + 1,3 µV			
		> 20 kHz	to 50 kHz	57 · 10 <sup>-6</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	70 · 10 <sup>-6</sup> · U + 2,5 µV			
	> 22 mV to 70 mV	10 Hz	to 20 Hz	70 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 20 Hz	to 40 Hz	39 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 40 Hz	to 20 kHz	35 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 20 kHz	to 50 kHz	44 · 10 <sup>-6</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	62 · 10 <sup>-6</sup> · U + 2,5 µV			
	> 70 mV to 220 mV	10 Hz	to 20 Hz	0,12 · 10 <sup>-3</sup> · U + 1,5 µV			
		> 20 Hz	to 40 Hz	84 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 40 Hz	to 20 kHz	55 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 20 kHz	to 50 kHz	55 · 10 <sup>-6</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	79 · 10 <sup>-6</sup> · U + 2,5 µV			
		> 100 kHz	to 300 kHz	0,13 · 10 <sup>-3</sup> · U + 4 µV			
		> 300 kHz	to 500 kHz	0,28 · 10 <sup>-3</sup> · U + 6 µV			
		> 500 kHz	to 1 MHz	0,92 · 10 <sup>-3</sup> · U + 6 µV			
	> 220 mV to 700 mV	10 Hz	to 20 Hz	0,12 · 10 <sup>-3</sup> · U + 1,5 µV			
		> 20 Hz	to 40 Hz	66 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 40 Hz	to 20 kHz	37 · 10 <sup>-6</sup> · U + 1,5 µV			
		> 20 kHz	to 50 kHz	55 · 10 <sup>-6</sup> · U + 2,0 µV			
		> 50 kHz	to 100 kHz	71 · 10 <sup>-6</sup> · U + 2,5 µV			
		> 100 kHz	to 300 kHz	0,13 · 10 <sup>-3</sup> · U + 4 µV			
		> 300 kHz	to 500 kHz	0,28 · 10 <sup>-3</sup> · U + 6 µV			
		> 500 kHz	to 1 MHz	0,91 · 10 <sup>-3</sup> · U + 6 µV			

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 7 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage	> 700 mV to 2,2 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	67 · 10 <sup>-6</sup> 33 · 10 <sup>-6</sup> 36 · 10 <sup>-6</sup> 45 · 10 <sup>-6</sup> 70 · 10 <sup>-6</sup> 0,13 · 10 <sup>-3</sup> 0,24 · 10 <sup>-3</sup> 0,85 · 10 <sup>-3</sup>	
	> 2,2 V to 7 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	67 · 10 <sup>-6</sup> 34 · 10 <sup>-6</sup> 28 · 10 <sup>-6</sup> 38 · 10 <sup>-6</sup> 80 · 10 <sup>-6</sup> 0,15 · 10 <sup>-3</sup> 0,38 · 10 <sup>-3</sup> 1,11 · 10 <sup>-3</sup>	
	> 7 V to 22 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	67 · 10 <sup>-6</sup> 34 · 10 <sup>-6</sup> 28 · 10 <sup>-6</sup> 38 · 10 <sup>-6</sup> 80 · 10 <sup>-6</sup> 0,17 · 10 <sup>-3</sup> 0,39 · 10 <sup>-3</sup> 1,14 · 10 <sup>-3</sup>	
	> 22 V to 70 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	67 · 10 <sup>-6</sup> 35 · 10 <sup>-6</sup> 30 · 10 <sup>-6</sup> 39 · 10 <sup>-6</sup> 81 · 10 <sup>-6</sup>	
	> 70 V to 220 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	82 · 10 <sup>-6</sup> 58 · 10 <sup>-6</sup> 55 · 10 <sup>-6</sup> 58 · 10 <sup>-6</sup> 96 · 10 <sup>-6</sup>	
	> 220 V to 700 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	94 · 10 <sup>-6</sup> 77 · 10 <sup>-6</sup> 74 · 10 <sup>-6</sup> 0,13 · 10 <sup>-3</sup> 0,42 · 10 <sup>-3</sup>	
	> 700 V to 1000 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz	0,12 · 10 <sup>-3</sup> 0,11 · 10 <sup>-3</sup> 0,11 · 10 <sup>-3</sup>	

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 8 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
High voltage	> 1 kV to 10 kV	40 Hz to 60 Hz	$5 \cdot 10^{-3} \cdot U + 2 \text{ V}$	$U = \text{Measured value}$
AC current	190 µA to 800 µA	10 Hz to 20 Hz	$0,26 \cdot 10^{-3} \cdot I + 20 \text{ nA}$	$I = \text{Measured value}$
		> 20 Hz to 40 Hz	$0,15 \cdot 10^{-3} \cdot I + 20 \text{ nA}$	
		> 40 Hz to 10 kHz	$0,10 \cdot 10^{-3} \cdot I + 20 \text{ nA}$	
	> 800 µA to 8 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3} \cdot I + 20 \text{ nA}$	
		> 20 Hz to 40 Hz	$90 \cdot 10^{-6} \cdot I + 20 \text{ nA}$	
		> 40 Hz to 10 kHz	$60 \cdot 10^{-6} \cdot I + 20 \text{ nA}$	
	> 8 mA to 10 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 10 mA to 20 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 20 mA to 30 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 30 mA to 50 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 50 mA to 100 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 100 mA to 200 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 200 mA to 300 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 300 mA to 500 mA	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	
	> 500 mA to 1 A	10 Hz to 20 Hz	$0,25 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$95 \cdot 10^{-6}$	
		> 40 Hz to 10 kHz	$50 \cdot 10^{-6}$	

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 9 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC current	> 1 A to 2 A	10 Hz to 20 Hz	$0,3 \cdot 10^{-3}$	with current coil
		> 20 Hz to 40 Hz	$0,15 \cdot 10^{-3}$	
		> 40 Hz to 10 kHz	$85 \cdot 10^{-6}$	
	> 2 A to 3 A	10 Hz to 20 Hz	$0,3 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0,15 \cdot 10^{-3}$	
		> 40 Hz to 10 kHz	$70 \cdot 10^{-6}$	
	> 3 A to 5 A	10 Hz to 20 Hz	$0,3 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0,2 \cdot 10^{-3}$	
		> 40 Hz to 10 kHz	$0,18 \cdot 10^{-3}$	
	> 5 A to 10 A	10 Hz to 20 Hz	$0,3 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0,15 \cdot 10^{-3}$	
		> 40 Hz to 10 kHz	$0,13 \cdot 10^{-3}$	
	> 10 A to 20 A	10 Hz to 20 Hz	$0,3 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0,15 \cdot 10^{-3}$	
		> 40 Hz to 10 kHz	$0,1 \cdot 10^{-3}$	
Clamp-on current meter	0,01 A to 1000 A	45 Hz to 65 Hz	$8 \cdot 10^{-3}$	with current coil
	0,01 A to 150 A	> 65 Hz to 400 Hz	$6 \cdot 10^{-3}$	
Capacitance Standards	10 pF to 100 pF	500 Hz to 4 kHz	$3,5 \cdot 10^{-3} \cdot C + 0,05 \text{ pF}$	C = Measured value
	> 100 pF to 1 nF	500 Hz to 4 kHz	$2 \cdot 10^{-3}$	
	> 1 nF to 10 nF	50 Hz to 800 Hz	$2 \cdot 10^{-3}$	
		> 800 Hz to 4 kHz	$1 \cdot 10^{-3}$	
	> 10 nF to 10 $\mu\text{F}$	50 Hz to 10 kHz	$1 \cdot 10^{-3}$	
	> 10 $\mu\text{F}$ to 100 $\mu\text{F}$	50 Hz to 4 kHz	$1 \cdot 10^{-3}$	
	> 100 $\mu\text{F}$ to 1 mF	20 Hz to 600 Hz	$2 \cdot 10^{-3}$	
Capacitance Measuring instruments	10 pF	1 kHz	$30 \cdot 10^{-6}$	
	100 pF	1 kHz	$30 \cdot 10^{-6}$	
	1 nF	100 Hz; 400 Hz	$0,2 \cdot 10^{-3}$	
		1 kHz	$20 \cdot 10^{-6}$	
		10 kHz	$0,2 \cdot 10^{-3}$	
	10 nF	100 Hz; 400 Hz; 1 kHz	$0,2 \cdot 10^{-3}$	
		10 kHz	$0,3 \cdot 10^{-3}$	
	100 nF	100 Hz; 400 Hz; 1 kHz	$0,2 \cdot 10^{-3}$	
		10 kHz	$0,3 \cdot 10^{-3}$	
	1 $\mu\text{F}$	100 Hz; 400 Hz; 1 kHz	$0,2 \cdot 10^{-3}$	
		10 kHz	$0,3 \cdot 10^{-3}$	

Valid from: 09.06.2023

Date of issue: 20.07.2023

Page 10 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Inductance Standards	100 µH to 1 mH	50 Hz to 500 Hz > 500 Hz to 5 kHz > 5 kHz to 10 kHz	4 · 10 <sup>-3</sup> · L + 0,25 µH 2 · 10 <sup>-3</sup> 1 · 10 <sup>-3</sup>	$L = \text{Measured value}$
	> 1 mH to 10 mH	50 Hz to 500 Hz > 500 Hz to 10 kHz	1,5 · 10 <sup>-3</sup> 0,8 · 10 <sup>-3</sup>	
	> 10 mH to 100 mH	50 Hz to 10 kHz	0,8 · 10 <sup>-3</sup>	
	> 100 mH to 10 H	50 Hz to 1 kHz	0,8 · 10 <sup>-3</sup>	
Inductance Measuring instruments	100 µH	100 Hz; 400 Hz 1 kHz; 10 kHz	0,3 · 10 <sup>-3</sup>	$R = \text{ratio of larger to lower voltage } (U_{\text{ref}} / U_{\text{var}})$
	1 mH	100 Hz; 400 Hz	0,3 · 10 <sup>-3</sup>	
		1 kHz	0,2 · 10 <sup>-3</sup>	
		10 kHz	0,3 · 10 <sup>-3</sup>	
	10 mH	100 Hz; 400 Hz; 1 kHz	0,2 · 10 <sup>-3</sup>	
		10 kHz	0,3 · 10 <sup>-3</sup>	
	100 mH	100 Hz; 400 Hz; 1 kHz	0,2 · 10 <sup>-3</sup>	
		10 kHz	0,3 · 10 <sup>-3</sup>	
	1 H	100 Hz; 400 Hz; 1 kHz	0,2 · 10 <sup>-3</sup>	
Phase angle between voltages	0° to 360°	10 Hz to 2 kHz $U_{\text{ref}}$   $U_{\text{var}}$		$R = \text{ratio of larger to lower voltage } (U_{\text{ref}} / U_{\text{var}})$
		5 V	5 V	
		50 mV to 100 V	(0,003 + 0,0005 · R)°	
		> 100 V to 120 V	(0,005 + 0,0001 · R)°	
Phase angle between voltage and current Measuring instruments	0° to 360°	10 Hz to 2 kHz $U_{\text{ref}}$   $I_{\text{var}}$		$R = \text{ratio of larger to lower voltage } (U_{\text{ref}} / U_{\text{var}})$ $I_{\text{var}} = U_{\text{var}} \cdot 0,1 \text{ A/V}$
		5 V	0,5 A; 5 A	
		50 mV to 100 V	(0,005 + 0,0005 · R)°	
		> 100 V to 120 V	(0,005 + 0,0001 · R)°	
		50 mV to 100 V	(0,005 + 0,0005 · R)°	
		> 100 V to 120 V	(0,005 + 0,0001 · R)°	

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Phase angle between voltages Sources	0° to 360°	10 Hz to 2 kHz 50 mV to 120 V	0,05°	
Power factor	-1 to 1	10 Hz to 2 kHz $U_{\text{ref}}$   $I_{\text{var}}$		
		5 V	0,5 A; 5 A	0,005°
		50 mV to 100 V	10 mA to 0,5 A	$(0,005 + 0,0005 \cdot R)^\circ$
		> 100 V to 120 V	10 mA to 0,5 A	$(0,005 + 0,0001 \cdot R)^\circ$
		50 mV to 100 V	> 0,5 A to 5 A	$(0,005 + 0,0005 \cdot R)^\circ$
		> 100 V to 120 V	> 0,5 A to 5 A	$(0,005 + 0,0001 \cdot R)^\circ$

Valid from: 09.06.2023  
 Date of issue: 20.07.2023

**Page 12 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15086-01-01**

**Permanent laboratory**

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Frequency	1 MHz; 5 MHz; 10 MHz		$2 \cdot 10^{-12}$	Phase-time difference measurement
	1 μHz to 3 GHz		$7 \cdot 10^{-12} + U_{Tf}$	Trigger uncertainty $U_{Tf}$ has to be determined from signal to noise ratio and signal rising time.
	> 3 GHz to 18 GHz		$2 \cdot 10^{-10} + U_{Tf}$	
Time interval $t$	5 ns to 100000 s		$[(1 \text{ ns})^2 + (5 \cdot 10^{-11} \cdot t)^2]^{1/2}$	
Length Calipers for external, internal and depth dimensions *	0 mm to 300 mm > 300 mm to 1000 mm	VDI/VDE/DGQ 2618, Blatt 9.1:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$ $50 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	$l$ = measured length
Micrometers *	0 mm to 100 mm	VDI/VDE/DGQ 2618, Blatt 10.1:2001	$3 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Reference gauges for micrometers with plane-parallel or spherical measuring faces, spherical-end gauges and internal calipers *	to 100 mm	VDI/VDE/DGQ 2618, Blatt 4.4:2009	$2 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	
Mechanical dial gauges *	0 mm to 10 mm	VDI/VDE/DGQ/DKD 2618, Blatt 11.1:2021	$4 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l$ = measured length at $l > 10 \text{ mm}$ in horizontal position
	> 10 mm to 50 mm		$5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DGQ	Deutsche Gesellschaft für Qualität e.V.
DIN	Deutsches Institut für Normung e.V.
DKD	Deutscher Kalibrierdienst (German Calibration Service)
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.

Valid from: 09.06.2023

Date of issue: 20.07.2023

**Page 13 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.