

Deutsche Akkreditierungsstelle

Annex to the Partial Accreditation Certificate

D-K-15089-01-01

according to DIN EN ISO/IEC 17025:2018

Valid from: 09.08.2023

Date of issue: 27.05.2024

This annex is a part of the accreditation certificate D-K-15089-00.

Holder of partial accreditation certificate:

Perschmann Calibration GmbH
Hauptstr. 46d, 38110 Braunschweig

with the location

Perschmann Calibration GmbH
Hauptstr. 46d, 38110 Braunschweig

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Abbreviations used: see last page

Page 1 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Calibration in the areas:

Dimensional quantities

Length

- Gauge blocks
- Length measuring instruments
- Line scales, distances
- Length measuring devices ^{a)}
- Diameter
- Form error
- Flatness ^{a)}
- Straightness ^{a)}
- Thread

Coordinate measuring technology

- Coordinate measuring machines ^{b)}

Electrical quantities

DC and low frequency

- DC voltage
 - AC voltage
 - DC current
 - AC current
 - DC resistance
- Time and frequency**
- Frequency and speed

^{a)} also on-site calibration

^{b)} only on-site calibration

Within the accreditation areas marked with the *, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards/equivalent calibration procedures within the flexible scope of accreditation.

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 2 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Length Gauge blocks * made of steel according to DIN EN ISO 3650:1999	0,5 mm to 100 mm featuring the nominal values of the standard made of steel	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_x from the nominal value l_n by comparison measurement	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.05 \mu\text{m}$	Measuring surfaces quality as stated in QMH rsp. In the test specifications l = gauge block length
	> 100 mm to 150 mm featuring the nominal values of the standard made of steel	Measurement of the deviations f_o and f_u from the central length by 5 points comparison For the smallest	For the central length: $0.05 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.07 \mu\text{m}$	
Gauge blocks * made of ceramics according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standard made of steel	measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.05 \mu\text{m}$	
	> 100 mm to 150 mm featuring the nominal values of the standard made of steel		For the central length: $0.07 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.07 \mu\text{m}$	

Valid from: 09.08.2023

Date of issue: 27.05.2024

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Gauge blocks * made of steel with special cross section (round or square), also with drilling in the middle	0.5 mm to 100 mm featuring the nominal values of the standard made of steel	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_x from the nominal value l_n by comparison measurement	For the central length: $0.1 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.1 \mu\text{m}$	At square gauge blocks with drilling the mean size is substituted by ANSI-ASME B89.1.9M measured between hole and front side
Gauge blocks * made of tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standard made of steel	Measurement of the deviations f_o and f_u from the central length by 5 points comparison For the smallest measurement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.08 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$ For the deviations f_o and f_u from the central length: $0.05 \mu\text{m}$	
Gauge blocks * made of steel	> 150 mm to 1000 mm in the nominal dimensions, which differ of the standard with a max. of 50 mm	VDI/VDE/DGQ 2618 part 3.1:2004 Measurement of the deviation of the central length l_x from the nominal value l_n by comparison measurement	For the central length: $0.2 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$	
Setting ring gauges made of steel * diameter	2 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.1:2006	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	The measurement uncertainty applies to the complete calibration of diameter, roundness, straightness and parallelism. For the calibration of the diameter without form measurement, the best measurement uncertainty increases by $0.2 \mu\text{m}$. d = measured diameter
Setting plug gauges made of steel * diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 part 4.1: 2006	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	
Measuring pins made of steel * diameter	0.17 mm to 20 mm	VDI/VDE/DGQ 2618 part 4.2: 2006	$0.4 \mu\text{m}$	
Roundness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 40 μm	VDI/VDE/DGQ 2618 part 4.1: 2006	$0.2 \mu\text{m} + 1 \cdot 10^{-2} \cdot RONt$	Diameter: 2 mm to 200 mm
Straightness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 10 μm	VDI/VDE/DGQ 2618 part 4.1: 2006	$0.5 \mu\text{m}$	axial length: to 30 mm

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 4 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Parallelism * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 10 µm	VDI/VDE/DGQ 2618 part 4.1: 2006	0.7 µm	axial length: to 30 mm
Setting dimension *	25 mm to 900 mm	VDI/VDE/DGQ 2618 part 4.4:2009	$0.7 \mu\text{m} + 1.5 \cdot 10^{-6} \cdot l$	$l =$ measured length
Caliper gauge *	5 mm to 170 mm	VDI/VDE/DGQ 2618 part 4.7:2005	$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l =$ measured length
Reference and setting gauge	to 12 mm	Annex F/43:2017-11	3 µm	Feeler gauge, gap gauge of plastic, delrin, teflon, brass or steel
Radius gauge	to 40 mm	Annex F/42:2018-01	3 µm	to 40 mm radii
Angel meter *	0° to 360°	VDI/VDE/DGQ 2618 part 7.2:2008	1' 30"	
Graduator	0° to 180°	Annex F/46:2017-11	12'	
Measuring tape Circumference tape measure	0 m to 50 m	Annex F/47-1:2017-12 Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	$l =$ measured length
Diameter tape measure	0 m to 10 m	Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Rules	0 m to 10 m	Annex F/47-3:2017-12 Annex F/47-4:2017-11	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	$l =$ measured length Graduated metal rules, reference- and plotting scale, rules, folding rules
Calipers for external, internal and depth dimensions *	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 9.1:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	$l =$ measured length
Depth calipers *	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 9.2:2006	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	$l =$ measured length
Height gauge *	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 9.3:2006	$20 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with contact help
Indicating caliper micrometers *	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 10.1:2001	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	600 mm = final value of the measuring range
Indicating caliper gap gauge *	0 mm to 100 mm	VDI/VDE/DGQ 2618 part 10.3:2002	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	$l =$ measured length
Dial indicator snap gauge	0 mm to 100 mm	Annex F/39:2017-12	$2 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Depth micrometers *	0 mm to 300 mm	VDI/VDE/DGQ 2618 part 10.5:2010	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with interchangeable contact points
Internal micrometers with two-point contact *	25 mm to 950 mm	VDI/VDE/DGQ 2618 part 10.7:2010	$3.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	$d =$ measured diameter
Internal micrometers with jaws	5 mm to 100 mm	Annex F/37:2021-09	$5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Internal micrometers with three-point contact *	3 mm to 200 mm	VDI/VDE/DGQ 2618 part 10.8:2002	$3 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	$d =$ measured diameter

Valid from: 09.08.2023

Date of issue: 27.05.2024

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Dial gauges with scales *scale interval > 1 µm	to 100 mm	DKD-R 4-3 part 11.1:2018	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Dial gauges with scales * scale interval 1 µm	to 5 mm		1.5 µm	error of measurement y_i
			2 µm	Deviation span f_e, f_{ges}, f_u, f_t and f_w
Dial gauges with scales * scale interval > 1 µm	to 100 mm	VDI/VDE/DGQ 2618 part 11.1:2014	$3 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Dial gauges with scale * scale interval 1 µm	to 5 mm		1.5 µm	error of measurement y_i
			2 µm	deviation span $MPE_r, MPE_e, MPE_{ges},$ $MPE_{1/1}, MPE_{1/2},$ $MPE_{1/10}, MPE_u$
Dial gauges with digital display numerical interval 0.1 µm	to 25 mm	Annex F/04-2:2014-12	$0.6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$0.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_t and f_w
Dial gauges with digital display numerical interval 1 µm	to 100 mm		$1 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_t and f_w
Dial gauges with digital display * numerical interval 0.1 µm	to 25 mm	VDI/VDE/DGQ/DKD 2618 part 11.4:2020	$0.6 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$0.8 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span $MPE_R, MPE_H, MPE_E,$ MPE_P
Dial gauges with digital display * numerical interval 1 µm	to 100 mm		$1 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
			$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span $MPE_R, MPE_H, MPE_E,$ MPE_P
Dial indicator * scale interval > 0.5 µm	to 3 mm	VDI/VDE/DGQ 2618 part 11.2:2002	0.6 µm	
Lever gauges *	to 1,6 mm	VDI/VDE/DGQ 2618 part 11.3:2002	1 µm	
Lever gauges for external measurements *	0 mm to 70 mm	VDI/VDE/DGQ 2618 part 12.1:2005	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	up to a probe length of 200 mm
Lever gauges for internal measurements *	2,5 mm to 80 mm	VDI/VDE/DGQ 2618 part 13.1:2005	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Thickness gauges	0 mm to 30 mm	Annex F/17:2011-12	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length up to a measuring depth of 200 mm

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 6 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Bore gauges with 2-point contact *	1 mm to 3 mm	VDI/VDE/DGQ 2618 part 13.2:2005 (image 1)	0.8 μm	Range of application: with gauge slider $d = 1.75 \text{ mm}$ to $d = 25 \text{ mm}$
		VDI/VDE/DGQ 2618 part 13.2:2005 (image 2)	0.8 μm	Range of application: to $d = 300 \text{ mm}$
			1.2 μm	Range of application: $d > 300 \text{ mm}$ to $d = 600 \text{ mm}$
		VDI/VDE/DGQ 2618 part 13.2:2005 (image 3)	0.8 μm	Range of application: pluge gauge to $d = 100 \text{ mm}$
Height gauges *	0 mm to 1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.5 \mu\text{m} + 3 \cdot 10^{-6} \cdot L$	$L =$ measured length
Deviation from straightness and perpendicularity	to 30 μm	to 800 mm lead length	$2.5 \mu\text{m} + 1 \cdot 10^{-6} \cdot l_z$	$l_z =$ lead length
90° Squares Perpendicularity	to 30 μm	Annex F/12:2017-02 to 750 mm leg length	$2 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	$l_z =$ length of form respectively position embodiment
Deviation from straightness and parallelism	to 30 μm	Annex F/13:2020-10 to 750 mm length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	
Flatness deviation	to 30 μm	Annex F/13:2020-10 to 750 mm edge length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	
Deviation from flatness Horizontal flatness standard, e.g., surface plates as per DIN 876:1984	to 50 μm	Annex F/18:2020-10 to 2 m edge length electronic inclination measuring	$0.9 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$	$l =$ longest edge length of the measuring standard For calibration in the permanent calibration laboratory, the uncertainty increases starting by an edge length $l > 1 \text{ m}$ by a factor of 1.2
Deviation from straightness Horizontal flatness standard, e.g., surface plates as per DIN 876:1984	bis 50 μm	Annex F/18:2020-10 to 3 m edge length electronic inclination measuring	$2.2 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$	

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 7 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Thread gauges * (single-start and multi-start cylindrical and conical external and internal threads with straight flanks, symmetrical and asymmetrical profile)				
External thread	3 mm to 150 mm	Scanning method VDI/VDE/DGQ 2618 part 4.8:2006, Option 1 to Option 4 (Specifying the thread angle α)	$2.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	d = measured diameter l_F = side length
Simple pitch diameter	Nominal diameter		2 μm	
Outside diameter			5 μm	
Core diameter or recess diameter			1 μm	
Lead or pitch	0.5 mm to 8 mm		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at 6'	
Thread angle α	$\geq 27^\circ$			
Internal thread	3 mm to 160 mm	Scanning method VDI/VDE/DGQ 2618 part 4.9:2006, Option 1 to Option 4 (Specifying the thread angle α)	$2.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Simple pitch diameter	Nominal diameter		5 μm	
Outside diameter or recess diameter			2 μm	
Core diameter			1 μm	
Lead or pitch	0.5 mm to 8 mm		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at 6'	
Thread angle α	$\geq 27^\circ$			
Stand off	3 mm to 150 mm	Annex F/09-3:2021-09	50 μm	
Thread plug gauge * Simple pitch diameter	1.4 mm to 150 mm	VDI/VDE/DGQ 2618 part 4.8:2006, Option 1 (Three wire procedure)	$2.5 \mu\text{m} + 7,5 \cdot 10^{-6} \cdot d$	d = measured diameter P_h = lead $P_h \geq 0.3 \text{ mm}$ to $\leq 6 \text{ mm}$

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 8 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
DC and low frequency quantities				
DC voltage measuring instruments	0 mV to 220 mV > 0.22 V to 2.2 V > 2.2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1100 V	Annex F/23-1:2020-05	$0.65 \mu\text{V} + 6.6 \cdot 10^{-6} \cdot U$ $1.2 \mu\text{V} + 4.7 \cdot 10^{-6} \cdot U$ $2.6 \mu\text{V} + 3.5 \cdot 10^{-6} \cdot U$ $4 \mu\text{V} + 3.5 \cdot 10^{-6} \cdot U$ $5 \mu\text{V} + 5 \cdot 10^{-6} \cdot U$ $0.4 \text{ mV} + 6.5 \cdot 10^{-6} \cdot U$	U = measuring value
DC voltage sources	0 V to 0.2 V > 2.2 V to 2 V > 2 V to 20 V > 20 V to 200 V > 200 V to 1000 V	Annex F/29-1:2020-05	$0.53 \mu\text{V} + 2.9 \cdot 10^{-6} \cdot U$ $1 \mu\text{V} + 2.7 \cdot 10^{-6} \cdot U$ $4.1 \mu\text{V} + 3.0 \cdot 10^{-6} \cdot U$ $40 \mu\text{V} + 4.5 \cdot 10^{-6} \cdot U$ $0.5 \text{ mV} + 4.5 \cdot 10^{-6} \cdot U$	U = measuring value
DC current Measuring instruments	0 μA to 220 μA > 0.22 mA to 2.2 mA mA to 22 mA > 2.2 mA to 220 mA > 22 mA to 2.2 A > 0.22 A to 3 A > 2.2 A to 11 A > 3 A to 20.5 A > 11 A	Annex F/23-2:2020-05	$6 \text{ nA} + 40 \cdot 10^{-6} \cdot I$ $9.3 \text{ nA} + 34 \cdot 10^{-6} \cdot I$ $40 \text{ nA} + 35 \cdot 10^{-6} \cdot I$ $0.7 \mu\text{A} + 45 \cdot 10^{-6} \cdot I$ $13 \mu\text{A} + 79 \cdot 10^{-6} \cdot I$ $31 \mu\text{A} + 0.29 \cdot 10^{-3} \cdot I$ $0.39 \text{ mA} + 0.39 \cdot 10^{-3} \cdot I$ $0.58 \text{ mA} + 0.78 \cdot 10^{-3} \cdot I$	I = measuring value
DC current sources	10 μA to 200 μA > 0.2 mA to 2 mA > 2 mA to 20 mA > 20 mA to 200 mA > 0.2 A to 2.0 A > 2 A to 20 A	Annex F/29-2:2020-05	$0.4 \text{ nA} + 12 \cdot 10^{-6} \cdot I$ $8.9 \text{ nA} + 10 \cdot 10^{-6} \cdot I$ $41 \text{ nA} + 13 \cdot 10^{-6} \cdot I$ $0.8 \mu\text{A} + 36 \cdot 10^{-6} \cdot I$ $17 \mu\text{A} + 0.17 \cdot 10^{-3} \cdot I$ $0.4 \text{ mA} + 0.38 \cdot 10^{-3} \cdot I$	
DC current clamps	0.2 A to < 10 A 10 A to 100 A > 100 A to 1000 A	Annex F/23-2:2020-05	$10 \text{ mA} + 2 \cdot 10^{-3} \cdot I$ $0.1 \text{ A} + 2 \cdot 10^{-3} \cdot I$ $0.8 \text{ A} + 2.5 \cdot 10^{-3} \cdot I$	with current coil with 2, 10 and 50 windings

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 9 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
AC voltage AC voltage measuring instruments	0.02 V to 0.22 V	Annex F/23-3:2020-05 40 Hz to 20 kHz	$4 \mu\text{V} + 80 \cdot 10^{-6} \cdot U$	U = measuring value
	> 0.22 V to 2.2 V	40 Hz to 20 kHz	$8 \mu\text{V} + 42 \cdot 10^{-6} \cdot U$	
	> 2.2 V to 22 V	40 Hz to 20 kHz	$50 \mu\text{V} + 42 \cdot 10^{-6} \cdot U$	
	> 22 V to 220 V	40 Hz to 20 kHz	$0.64 \text{ mV} + 60 \cdot 10^{-6} \cdot U$	
	> 220 V to 1100 V	50 Hz to 1 kHz	$3.5 \text{ mV} + 70 \cdot 10^{-6} \cdot U$	
AC voltage sources	10 mV to 200 mV	Annex F/29-3:2020-05 40 Hz to 100 Hz > 100 Hz to 2 kHz Hz to 10kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz	$4 \mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ $2 \mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ $4 \mu\text{V} + 0.13 \cdot 10^{-3} \cdot U$ $8 \mu\text{V} + 0.34 \cdot 10^{-3} \cdot U$	
	> 0.2 V to 2.0 V	40 Hz to 100 Hz > 100 Hz to 2 kHz Hz to 10 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz	$20 \mu\text{V} + 90 \cdot 10^{-6} \cdot U$ $20 \mu\text{V} + 75 \cdot 10^{-6} \cdot U$ $20 \mu\text{V} + 0.11 \cdot 10^{-3} \cdot U$ $40 \mu\text{V} + 0.22 \cdot 10^{-3} \cdot U$	
	> 2.0 V to 20 V	40 Hz to 100 Hz > 100 Hz to 2 kHz Hz to 10 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz	$0.2 \text{ mV} + 90 \cdot 10^{-6} \cdot U$ $0.2 \text{ mV} + 75 \cdot 10^{-6} \cdot U$ $0.2 \text{ mV} + 0.11 \cdot 10^{-3} \cdot U$ $0.4 \text{ mV} + 0.22 \cdot 10^{-3} \cdot U$	
	> 20 V to 200 V	40 Hz to 100 Hz > 100 Hz to 2 kHz Hz to 10kHz > 2 kHz to 30 kHz > 10 kHz	$2 \text{ mV} + 90 \cdot 10^{-6} \cdot U$ $2 \text{ mV} + 75 \cdot 10^{-6} \cdot U$ $2 \text{ mV} + 0.11 \cdot 10^{-3} \cdot U$ $4 \text{ mV} + 0.22 \cdot 10^{-3} \cdot U$	
	> 200 V to 1000 V	40 kHz to 10 kHz > 10 kHz to 30 kHz	$20 \text{ mV} + 0.11 \cdot 10^{-3} \cdot U$ $40 \text{ mV} + 0.22 \cdot 10^{-3} \cdot U$	
AC current Measuring instruments	20 μA to 220 μA	Annex F/23-4:2020-05 40 Hz to 1 kHz	$8 \text{ nA} + 0.12 \cdot 10^{-3} \cdot I$	I = measuring value
	> 0.22 mA to 2.2 mA	40 Hz to 1 kHz	$36 \text{ nA} + 0.1 \cdot 10^{-3} \cdot I$	
	> 2.2 mA to 22 mA	40 Hz to 1 kHz	$0.35 \mu\text{A} + 0.1 \cdot 10^{-3} \cdot I$	
	> 22 mA to 220 mA	40 Hz to 1 kHz	$2.5 \mu\text{A} + 0.1 \cdot 10^{-3} \cdot I$	
	> 0.22 A to 2.2 A	40 Hz to 1 kHz	$36 \mu\text{A} + 0.24 \cdot 10^{-3} \cdot I$	
	> 2.2 A to 3 A	45 Hz to 1 kHz	$78 \mu\text{A} + 0.47 \cdot 10^{-3} \cdot I$	
	> 3 A to 11 A	45 Hz to 100 Hz	$1.6 \text{ mA} + 0.78 \cdot 10^{-3} \cdot I$	
> 11 A to 20.5 A	45 Hz to 100 Hz	$3.9 \text{ mA} + 1.2 \cdot 10^{-3} \cdot I$		

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 10 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
AC current clamps	0.2 A to < 10 A 10 A to 100 A > 100 A to 500 A > 500 A to 1000 A	Annex F/23-4:2020-05 50 Hz 50 Hz 50 Hz 50 Hz	10 mA + 2 · 10 ⁻³ · I 0.1 A + 2 · 10 ⁻³ · I 0.4 A + 2.5 · 10 ⁻³ · I 0.8 A + 2.5 · 10 ⁻³ · I	I = measuring value with current coil with 2, 10 and 50 windings
AC current sources	0.2 mA to 2.0 mA > 2.0 mA to 20 mA > 20 mA to 200 mA > 0.2 A to 2.0 A > 2.0 A to 20 A	Annex F/29-4:2020-05 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 2 kHz 10 Hz to 2 kHz	0.2 μA + 0.3 · 10 ⁻³ · I 2 μA + 0.3 · 10 ⁻³ · I 20 μA + 0.29 · 10 ⁻³ · I 0.2 mA + 0.62 · 10 ⁻³ · I 2 mA + 0.82 · 10 ⁻³ · I	I = measuring value
DC resistance Measuring instruments 4-wire connection	0 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ	Annex F/23-5:2020-05	2.3 mΩ 2.4 mΩ 2.8 mΩ 4.8 mΩ 37 mΩ 0.37 Ω 3.7 Ω	Fixed resistors
DC resistances 4-wire connection	0 Ω to 2 Ω > 2 Ω to 20 Ω > 20 Ω to 200 Ω > 200 Ω to 2 kΩ > 2 kΩ to 20 kΩ	Annex F/30:2020-05	0.11 mΩ + 3 · 10 ⁻⁶ · R 0.1 mΩ + 6.1 · 10 ⁻⁶ · R 91 μΩ + 7.8 · 10 ⁻⁶ · R 0.12 Ω + 0.59 · 10 ⁻³ · R 0.11 Ω + 4.4 · 10 ⁻⁶ · R	R = measuring value
Measuring instruments 2-wire connection	1 MΩ 10 MΩ 100 MΩ	Annex F/23-5:2020-05	0.06 kΩ 1.8 kΩ 0.24 MΩ	Fixed resistors
DC resistance DC resistances 2-wire connection	> 20 kΩ to 200 kΩ > 200 kΩ to 2 MΩ > 2 MΩ to 20 MΩ > 20 MΩ to 200 MΩ > 200 to 2 GΩ MΩ	Annex F/30:2020-05	91 mΩ + 7.8 · 10 ⁻⁶ · R 0.12 kΩ + 0.77 · 10 ⁻³ · R 0.14 kΩ + 0.81 · 10 ⁻³ · R 10 kΩ + 0.12 · 10 ⁻³ · R 1 MΩ + 0.51 · 10 ⁻³ · R	R = measuring value
Time and Frequency frequency measuring instruments	0.01 Hz to 120 Hz > 120 Hz to 1.2 kHz > 1.2 kHz to 12 kHz > 12 kHz to 120 kHz > 120 to 1.2 MHz kHz	Annex F/23-6:2020-05	12 mHz + 50 · 10 ⁻⁶ · F 0.12 Hz + 50 · 10 ⁻⁶ · F 1.2 Hz + 50 · 10 ⁻⁶ · F 12 Hz + 50 · 10 ⁻⁶ · F 0.12 kHz + 50 · 10 ⁻⁶ · F	F = measuring value
Revolution speed revolution counter - optical	120 min ⁻¹ to 100000 min ⁻¹	Annex F/24:2020-05	0.05 min ⁻¹ + 18 · 10 ⁻⁶ · n	Direct optical excitation n = measuring value

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 11 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

On-site calibration

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range	Measurement conditions/procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Length Deviation from flatness Horizontal flatness standard, e.g., surface plates as per DIN 876:1984	to 50 µm	Annex F/18:2020-10 to 2 m edge length Electronic inclination measuring	$0.9 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$	l = longest edge length of the measuring standard
Deviation from straightness Horizontal flatness standard, e.g., surface plates as per DIN 876:1984	to 50 µm	Annex F/18:2020-10 to 3 m edge length Electronic inclination measuring	$2.2 \mu\text{m} + 1.7 \cdot 10^{-6} \cdot l$	
Height gauges *	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$2.5 \mu\text{m} + 5 \cdot 10^{-6} \cdot l$	l = measured length
Coordinate measuring technology * Measuring projectors Measuring microscopes	Devices featuring a measuring surface with a face diagonal ≤ 530 mm	Calibration of metrological characteristics according to DKD-R 4-3 part 18.1:2018, and the following standards and guidelines DIN EN ISO 10360 VDI/VDE 2617		
Determination of probing error $PS-ID(OT)$ with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007		0.4 µm	Measuring projectors and measuring microscopes with visual probing with crosshairs or electronic edge detection	
The error of indication for size measurement $E-ID(OT)$ und $E-2D(OT)$ is determined with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007		$0.5 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot L$	L = measured length	

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 12 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15089-01-01

Abbreviations used:

Annex F	Calibration Guide of Perschmann Calibration GmbH
CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
DGQ	Deutsche Gesellschaft für Qualität e.V.
DIN	Deutsches Institut für Normung e.V.
DKD	Deutscher Kalibrierdienst
DKD-R	Guideline of Deutscher Kalibrierdienst, published by Physikalisch-Technischen Bundesanstalt
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.

Valid from: 09.08.2023

Date of issue: 27.05.2024

Page 13 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.