



Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-15123-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: **21.06.2023**

Date of issue: 21.06.2023

Holder of accreditation certificate:

Fluke Deutschland GmbH
Heinrich-Pesch-Straße 11, 50739 Köln

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Abbreviations used: see last page

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Page 1 of 31

Calibrations in the fields:

Electrical quantities

DC and low frequency quantities

- **DC voltage** ^{a)}
- **AC voltage** ^{a)}
- **DC current** ^{a)}
- **AC current** ^{a)}
- **DC resistance** ^{a)}
- **AC/DC Transfer**
- **Inductance**
- **Capacitance**

Time and frequency

- **Frequency**

High frequency and radiation quantities

High frequency quantities

- **Oscilloscope quantities**
- **Rise time**

^{a)} also as on-site-calibration

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
DC voltage reference standard	1 V		0.4 · 10 ⁻⁶	
	1.018 V		0.4 · 10 ⁻⁶	
measurement instruments and sources	10 V		0.2 · 10 ⁻⁶	
	> 10 V to 100 V		· [0.7 + (1.1 V/U) ²] ^½	
DC current measurement instruments	> 100 V to 1000 V		0.3 · 10 ⁻⁶	
	> 10 A to 1000 A		0.6 · 10 ⁻⁶	
sources	0.1 μA		35 · 10 ⁻⁶	
	> 0.1 μA to < 1 μA		8.0 · 10 ⁻⁶	
	> 1 μA to < 10 μA		8.0 · 10 ⁻⁶	
	10 μA to 2 A		6.0 · 10 ⁻⁶	
	> 2 A to 10 A		12 · 10 ⁻⁶	
	> 10 A to 20 A		14 · 10 ⁻⁶	
	> 20 A to 100 A		20 · 10 ⁻⁶	
current clamps	0.1 μA		35 · 10 ⁻⁶	
	> 0.1 μA to < 1 μA		6.0 · 10 ⁻⁶	
	> 1 μA to < 10 μA		6.0 · 10 ⁻⁶	
	10 μA to 2 A		4.0 · 10 ⁻⁶	
	> 2 A to 10 A		15 · 10 ⁻⁶	
	> 10 A to 20 A		20 · 10 ⁻⁶	
	> 20 A to 100 A		30 · 10 ⁻⁶	
DC resistance measurement instruments and reference standard	0 A to 1000 A		5.0 · 10 ⁻³	
	10 μΩ to < 100 μΩ		22 · 10 ⁻⁶	
	100 μΩ to < 1 mΩ		4.6 · 10 ⁻⁶	
	1 mΩ to < 10 mΩ		2.3 · 10 ⁻⁶	
	10 mΩ to < 1 Ω		1.2 · 10 ⁻⁶	
	1 Ω to < 10 kΩ		3.0 · 10 ⁻⁷	
	10 kΩ to 1 MΩ		5.0 · 10 ⁻⁷	
	> 1 MΩ to 100 MΩ		2.0 · 10 ⁻⁶	
	> 100 MΩ to 1000 MΩ		8.0 · 10 ⁻⁶	
	> 1 GΩ to 10 GΩ		1.1 · 10 ⁻³	
measurement instruments	> 1 GΩ to 10 GΩ		4.0 · 10 ⁻⁴	
	10 mW to 20 kW		27 · 10 ⁻⁶	
AC voltage AC Voltage- measurement instruments	1 mV		10 Hz 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	Voltage range: 2.2 mV AC Measurement Standard Fluke 5790A/5790B
			0.45 · 10 ⁻³ 0.42 · 10 ⁻³ 0.45 · 10 ⁻³ 0.45 · 10 ⁻³ 0.50 · 10 ⁻³ 0.60 · 10 ⁻³	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instruments	2 mV	10 Hz	$0.22 \cdot 10^{-3}$	Voltage range: 2.2 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz	$0.20 \cdot 10^{-3}$	
		40 Hz; 55 Hz; 60 Hz	$0.20 \cdot 10^{-3}$	
		120 Hz; 300 Hz; 400 Hz	$0.20 \cdot 10^{-3}$	
		500 Hz; 1 kHz; 10 kHz	$0.20 \cdot 10^{-3}$	
		20 kHz; 30 kHz; 50 kHz	$0.20 \cdot 10^{-3}$	
		70 kHz; 100 kHz	$0.20 \cdot 10^{-3}$	
		200 kHz; 300 kHz	$0.22 \cdot 10^{-3}$	
		500 kHz	$0.22 \cdot 10^{-3}$	
		700 kHz; 800 kHz	$0.25 \cdot 10^{-3}$	
		1 MHz	$0.30 \cdot 10^{-3}$	
	2 mV	10 Hz	$0.20 \cdot 10^{-3}$	Voltage range: 7 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz	$0.18 \cdot 10^{-3}$	
		30 Hz; 40 Hz; 55 Hz; 60 Hz	$0.17 \cdot 10^{-3}$	
		120 Hz; 300 Hz; 400 Hz	$0.17 \cdot 10^{-3}$	
		500 Hz; 1 kHz; 10 kHz	$0.17 \cdot 10^{-3}$	
		20 kHz; 30 kHz; 50 kHz	$0.17 \cdot 10^{-3}$	
		70 kHz	$0.17 \cdot 10^{-3}$	
		100 kHz	$0.18 \cdot 10^{-3}$	
		200 kHz; 300 kHz; 500 kHz	$0.19 \cdot 10^{-3}$	
		700 kHz; 800 kHz	$0.22 \cdot 10^{-3}$	
		1 MHz	$0.25 \cdot 10^{-3}$	
	6 mV	10 Hz	$0.12 \cdot 10^{-3}$	Voltage range: 22 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz	$0.10 \cdot 10^{-3}$	
		40 Hz; 55 Hz; 60 Hz	$0.10 \cdot 10^{-3}$	
		120 Hz; 300 Hz; 400 Hz	$0.10 \cdot 10^{-3}$	
		500 Hz; 1 kHz; 10 kHz	$0.10 \cdot 10^{-3}$	
		20 kHz; 30 kHz; 50 kHz	$0.10 \cdot 10^{-3}$	
		70 kHz; 100 kHz	$0.10 \cdot 10^{-3}$	
		200 kHz; 300 kHz; 500 kHz	$0.12 \cdot 10^{-3}$	
		700 kHz; 800 kHz	$0.14 \cdot 10^{-3}$	
		1 MHz	$0.17 \cdot 10^{-3}$	
	2 mV	10 Hz	$0.19 \cdot 10^{-3}$	Voltage range: 22 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz	$0.16 \cdot 10^{-3}$	
		40 Hz; 55 Hz; 60 Hz	$0.16 \cdot 10^{-3}$	
		120 Hz; 300 Hz; 400 Hz	$0.16 \cdot 10^{-3}$	
		500 Hz; 1 kHz; 10 kHz	$0.16 \cdot 10^{-3}$	
		20 kHz; 30 kHz	$0.16 \cdot 10^{-3}$	
		50 kHz; 70 kHz; 100 kHz	$0.17 \cdot 10^{-3}$	
		200 kHz; 300 kHz; 500 kHz	$0.18 \cdot 10^{-3}$	
		700 kHz	$0.20 \cdot 10^{-3}$	
		800 kHz	$0.21 \cdot 10^{-3}$	
		1 MHz	$0.25 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	6 mV	10 Hz	$1.0 \cdot 10^{-3}$	Voltage range: 22 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$80 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$80 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$80 \cdot 10^{-6}$	
	10 mV	1 kHz; 10 kHz	$80 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$75 \cdot 10^{-6}$	
		70 kHz	$75 \cdot 10^{-6}$	
		100 kHz	$80 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$90 \cdot 10^{-6}$	
		500 kHz	$0.10 \cdot 10^{-3}$	
		700 kHz; 800 kHz	$0.14 \cdot 10^{-3}$	
		1 MHz	$0.16 \cdot 10^{-3}$	
		10 Hz	$80 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$60 \cdot 10^{-6}$	
	20 mV	60 Hz; 120 Hz; 300 Hz	$60 \cdot 10^{-6}$	Voltage range: 70 mV
		400 Hz; 500 Hz	$60 \cdot 10^{-6}$	
		1 kHz	$60 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 30 kHz	$55 \cdot 10^{-6}$	
		50 kHz; 70 kHz; 100 kHz	$35 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$45 \cdot 10^{-6}$	
		500 kHz	$50 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$70 \cdot 10^{-6}$	
		1 MHz	$80 \cdot 10^{-6}$	
		10 Hz	$90 \cdot 10^{-6}$	
	20 mV	20 Hz; 30 Hz; 40 Hz; 55 Hz	$65 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$65 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$60 \cdot 10^{-6}$	
		1 kHz; 10 kHz	$60 \cdot 10^{-6}$	
		20 kHz; 30 kHz	$55 \cdot 10^{-6}$	
		50 kHz; 70 kHz	$55 \cdot 10^{-6}$	
		100 kHz	$65 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$80 \cdot 10^{-6}$	
		500 kHz	$90 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$0.12 \cdot 10^{-3}$	
		1 MHz	$0.14 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	40 mV	10 Hz	$75 \cdot 10^{-6}$	Voltage range: 70 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$55 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$55 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$50 \cdot 10^{-6}$	
	60 mV	1 kHz; 10 kHz	$50 \cdot 10^{-6}$	
		20 kHz; 30 kHz	$45 \cdot 10^{-6}$	
		50 kHz; 70 kHz	$45 \cdot 10^{-6}$	
		100 kHz	$55 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$65 \cdot 10^{-6}$	
		500 kHz	$75 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$0.10 \cdot 10^{-3}$	
		1 MHz	$0.12 \cdot 10^{-3}$	
		10 Hz	$60 \cdot 10^{-6}$	
		20 Hz; 30 Hz	$40 \cdot 10^{-6}$	
	60 mV	60 Hz; 120 Hz; 300 Hz	$40 \cdot 10^{-6}$	Voltage range: 220 mV
		400 Hz; 500 Hz	$35 \cdot 10^{-6}$	
		1 kHz; 10 kHz	$35 \cdot 10^{-6}$	
		20 kHz; 30 kHz	$35 \cdot 10^{-6}$	
		50 kHz; 70 kHz; 100 kHz	$30 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$40 \cdot 10^{-6}$	
		500 kHz	$50 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$70 \cdot 10^{-6}$	
		1 MHz	$80 \cdot 10^{-6}$	
		10 Hz	$40 \cdot 10^{-6}$	
	100 mV; 200 mV	20 Hz; 30 Hz	$30 \cdot 10^{-6}$	
		55 Hz; 40 Hz; 60 Hz	$30 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$30 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$30 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$30 \cdot 10^{-6}$	
		70 kHz; 100 kHz	$30 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$40 \cdot 10^{-6}$	
		500 kHz	$75 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$0.10 \cdot 10^{-3}$	
		1 MHz	$0.12 \cdot 10^{-3}$	
		10 Hz	$15 \cdot 10^{-6}$	
		20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$8.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz; 20 kHz	$8.0 \cdot 10^{-6}$	
		30 kHz; 50 kHz; 70 kHz	$8.0 \cdot 10^{-6}$	
		100 kHz	$9.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	200 mV	10 Hz	$15 \cdot 10^{-6}$	Voltage range: 700 mV AC Measurement Standard Fluke 5790A/5790B
		20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$8.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz; 20 kHz	$8.0 \cdot 10^{-6}$	
		30 kHz; 50 kHz; 70 kHz	$8.0 \cdot 10^{-6}$	
		100 kHz	$9.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
	300 mV; 400 mV	700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	
		10 Hz; 20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
	500 mV; 600 mV; 700 mV	60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	Voltage range: 2.2 V
		400 Hz; 500 Hz	$5.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz; 20 kHz	$5.0 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$5.0 \cdot 10^{-6}$	
		70 kHz	$6.0 \cdot 10^{-6}$	
		100 kHz	$9.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	
	600 mV	10 Hz; 20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$5.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz; 20 kHz	$5.0 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$5.0 \cdot 10^{-6}$	
		70 kHz	$6.0 \cdot 10^{-6}$	
		100 kHz	$9.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	1 V; 2 V	10 Hz	$5.0 \cdot 10^{-6}$	Voltage range: 2.2 V AC Measurement Standard Fluke 5790A/5790B
		20 Hz; 30 Hz; 40 Hz	$4.0 \cdot 10^{-6}$	
		55 Hz; 60 Hz; 120 Hz	$4.0 \cdot 10^{-6}$	
		300 Hz; 400 Hz; 500 Hz	$4.0 \cdot 10^{-6}$	
	2 V; 3 V; 4 V; 5 V; 6 V; 7 V	1 kHz; 10 kHz; 20 kHz	$4.0 \cdot 10^{-6}$	Voltage range: 7 V
		30 kHz; 50 kHz; 70 kHz	$4.0 \cdot 10^{-6}$	
		100 kHz; 200 kHz	$6.0 \cdot 10^{-6}$	
		300 kHz; 500 kHz	$8.0 \cdot 10^{-6}$	
	6 V; 8 V; 10 V	700 kHz; 800 kHz	$10 \cdot 10^{-6}$	Voltage range: 22 V
		1 MHz	$20 \cdot 10^{-6}$	
		10 Hz	$5.0 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz	$4.0 \cdot 10^{-6}$	
	20 V	55 Hz; 60 Hz; 120 Hz	$4.0 \cdot 10^{-6}$	
		300 Hz; 400 Hz	$4.0 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$4.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$4.0 \cdot 10^{-6}$	
		70 kHz	$4.0 \cdot 10^{-6}$	
		100 kHz; 200 kHz	$6.0 \cdot 10^{-6}$	
		300 kHz; 500 kHz	$8.0 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$10 \cdot 10^{-6}$	
		1 MHz	$20 \cdot 10^{-6}$	
		10 Hz; 20 Hz	$6.0 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz; 60 Hz	$6.0 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$6.0 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$6.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$6.0 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	20 V	10 Hz; 20 Hz	$7.0 \cdot 10^{-6}$	Voltage range: 70 V AC Measurement Standard Fluke 5790A/5790B
		30 Hz; 40 Hz; 55 Hz; 60 Hz	$7.0 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$7.0 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$7.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$7.0 \cdot 10^{-6}$	
	30 V	70 kHz	$7.0 \cdot 10^{-6}$	
		100 kHz	$8.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$10 \cdot 10^{-6}$	
		500 kHz	$12 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$25 \cdot 10^{-6}$	
	40 V; 50 V; 60 V; 70 V	1 MHz	$35 \cdot 10^{-6}$	
		10 Hz	$8.0 \cdot 10^{-6}$	
		20 Hz; 30 Hz	$7.0 \cdot 10^{-6}$	
		40 Hz; 55 Hz; 60 Hz	$7.0 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$7.0 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$7.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$7.0 \cdot 10^{-6}$	
		70 kHz	$7.0 \cdot 10^{-6}$	
		100 kHz	$9.0 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$12 \cdot 10^{-6}$	
	60 V	500 kHz	$15 \cdot 10^{-6}$	Voltage range: 220 V
		700 kHz; 800 kHz	$25 \cdot 10^{-6}$	
		1 MHz	$35 \cdot 10^{-6}$	
		10 Hz	$9.0 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$8.0 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$8.0 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$7.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz	$7.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$7.0 \cdot 10^{-6}$	
		70 kHz	$9.0 \cdot 10^{-6}$	
	100 V; 200 V	100 kHz	$12 \cdot 10^{-6}$	
		10 Hz	$9.0 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$8.0 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$8.0 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$7.0 \cdot 10^{-6}$	
		1 kHz; 10 kHz	$7.0 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$7.0 \cdot 10^{-6}$	
		70 kHz	$9.0 \cdot 10^{-6}$	
		100 kHz	$12 \cdot 10^{-6}$	
		10 Hz	$14 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage AC voltage- measurement instrument	200 V	10 Hz 20 Hz; 30 Hz; 40 Hz 55 Hz; 60 Hz 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$14 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $18 \cdot 10^{-6}$ $30 \cdot 10^{-6}$	Voltage range: 700 V AC Measurement Standard Fluke 5790A/5790B
	300 V; 400 V	10 Hz 20 Hz; 30 Hz; 40 Hz 55 Hz; 60 Hz 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$16 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $18 \cdot 10^{-6}$ $30 \cdot 10^{-6}$	
	500 V; 600 V; 700 V	10 Hz 20 Hz; 30 Hz; 40 Hz 55 Hz; 60 Hz 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$16 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $35 \cdot 10^{-6}$	
	200 V	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz 120 Hz; 300 Hz 400 Hz; 500 Hz; 1 kHz 10 kHz; 20 kHz; 30 kHz 50 kHz 70 kHz 100 kHz	$14 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $18 \cdot 10^{-6}$ $30 \cdot 10^{-6}$	Voltage range: 1000 V
	300 V; 500 V	10 Hz 20 Hz 30 Hz; 40 Hz; 55 Hz; 60 Hz 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$16 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $18 \cdot 10^{-6}$ $30 \cdot 10^{-6}$	
	600 V; 800 V; 1000 V	10 Hz 20 Hz; 30 Hz; 40 Hz 55 Hz; 60 Hz; 120 Hz 300 Hz; 400 Hz; 500 Hz 1 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz 100 kHz	$18 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $35 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage measurement instruments and sources	0.6 mV to 2.2 mV	10 Hz to < 20 Hz	$1.1 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 2.2 mV
		20 Hz to < 40 Hz	$0.49 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		40 Hz to 20 kHz	$0.28 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		> 20 kHz to 50 kHz	$0.54 \cdot 10^{-3} \cdot U + 1.3 \mu\text{V}$	
		> 50 kHz to 100 kHz	$0.80 \cdot 10^{-3} \cdot U + 1.7 \mu\text{V}$	
		> 100 kHz to 300 kHz	$1.5 \cdot 10^{-3} \cdot U + 2.7 \mu\text{V}$	
	1.9 mV to 7 mV	> 300 kHz to 500 kHz	$1.6 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 7 mV
		> 500 kHz to 1 MHz	$2.1 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		10 Hz to < 20 Hz	$0.57 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		20 Hz to < 40 Hz	$0.25 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		40 Hz to 20 kHz	$0.14 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		> 20 kHz to 50 kHz	$0.27 \cdot 10^{-3} \cdot U + 1.3 \mu\text{V}$	
	6 mV to 22 mV	> 50 kHz to 100 kHz	$0.40 \cdot 10^{-3} \cdot U + 1.7 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 22 mV
		> 100 kHz to 300 kHz	$0.80 \cdot 10^{-3} \cdot U + 2.7 \mu\text{V}$	
		> 300 kHz to 500 kHz	$0.87 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		> 500 kHz to 1 MHz	$1.3 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		10 Hz to < 20 Hz	$0.19 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
		20 Hz to < 40 Hz	$0.12 \cdot 10^{-3} \cdot U + 0.9 \mu\text{V}$	
	19 mV to 70 mV	40 Hz to 20 kHz	$73 \cdot 10^{-6} \cdot U + 0.9 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 70 mV
		> 20 kHz to 50 kHz	$0.14 \cdot 10^{-3} \cdot U + 1.3 \mu\text{V}$	
		> 50 kHz to 100 kHz	$0.21 \cdot 10^{-3} \cdot U + 1.7 \mu\text{V}$	
		> 100 kHz to 300 kHz	$0.54 \cdot 10^{-3} \cdot U + 2.7 \mu\text{V}$	
		> 300 kHz to 500 kHz	$0.57 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		> 500 kHz to 1 MHz	$0.93 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
	60 mV to 220 mV	9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U + 1.0 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 220 mV
		10 Hz to < 20 Hz	$0.16 \cdot 10^{-3} \cdot U + 1.0 \mu\text{V}$	
		20 Hz to < 40 Hz	$80 \cdot 10^{-6} \cdot U + 1.0 \mu\text{V}$	
		40 Hz to 20 kHz	$43 \cdot 10^{-6} \cdot U + 1.0 \mu\text{V}$	
		> 20 kHz to 50 kHz	$80 \cdot 10^{-6} \cdot U + 1.3 \mu\text{V}$	
		> 50 kHz to 100 kHz	$0.17 \cdot 10^{-3} \cdot U + 1.7 \mu\text{V}$	
	190 mV to 700 mV	> 100 kHz to 300 kHz	$0.34 \cdot 10^{-3} \cdot U + 2.7 \mu\text{V}$	<i>U</i> : measured value 5790A/B Range: 700 mV
		> 300 kHz to 500 kHz	$0.44 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		> 500 kHz to 1 MHz	$0.73 \cdot 10^{-3} \cdot U + 4.0 \mu\text{V}$	
		9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U + 1.0 \mu\text{V}$	
		10 Hz to < 20 Hz	$0.14 \cdot 10^{-3} \cdot U + 1.0 \mu\text{V}$	
		20 Hz to < 40 Hz	$50 \cdot 10^{-6} \cdot U + 1.0 \mu\text{V}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage measurement instruments and sources	600 mV to 2.2 V	9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U$	<i>U</i> : measured value 5790A/B Range: 2.2 V
		10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	
		20 Hz to < 40 Hz	$43 \cdot 10^{-6} \cdot U$	
		40 Hz to 20 kHz	$15 \cdot 10^{-6} \cdot U$	
		> 20 kHz to 50 kHz	$30 \cdot 10^{-6} \cdot U$	
		> 50 kHz to 100 kHz	$47 \cdot 10^{-6} \cdot U$	
		> 100 kHz to 300 kHz	$0.10 \cdot 10^{-3} \cdot U$	
	1.9 V to 7 V	> 300 kHz to 500 kHz	$0.17 \cdot 10^{-3} \cdot U$	<i>U</i> : measured value 5790A/B Range: 7 V
		> 500 kHz to 1 MHz	$0.56 \cdot 10^{-3} \cdot U$	
		9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U$	
		10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	
		20 Hz to < 40 Hz	$44 \cdot 10^{-6} \cdot U$	
		40 Hz to 20 kHz	$15 \cdot 10^{-6} \cdot U$	
		> 20 kHz to 50 kHz	$31 \cdot 10^{-6} \cdot U$	
	6 V to 22 V	> 50 kHz to 100 kHz	$53 \cdot 10^{-6} \cdot U$	<i>U</i> : measured value 5790A/B Range: 22 V
		> 100 kHz to 300 kHz	$0.12 \cdot 10^{-3} \cdot U$	
		> 300 kHz to 500 kHz	$0.25 \cdot 10^{-3} \cdot U$	
		> 500 kHz to 1 MHz	$0.73 \cdot 10^{-3} \cdot U$	
		9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U$	
		10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	
		20 Hz to < 40 Hz	$44 \cdot 10^{-6} \cdot U$	
	19 V to 70 V	40 Hz to 20 kHz	$17 \cdot 10^{-6} \cdot U$	<i>U</i> : measured value 5790A/B Range: 70 V
		> 20 kHz to 50 kHz	$31 \cdot 10^{-6} \cdot U$	
		> 50 kHz to 100 kHz	$53 \cdot 10^{-6} \cdot U$	
		> 100 kHz to 300 kHz	$0.12 \cdot 10^{-3} \cdot U$	
		> 300 kHz to 500 kHz	$0.25 \cdot 10^{-3} \cdot U$	
		> 500 kHz to 1 MHz	$0.73 \cdot 10^{-3} \cdot U$	
		9.5 Hz to < 10 Hz	$0.67 \cdot 10^{-3} \cdot U$	
	60 V to 220 V	10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	<i>U</i> : measured value 5790A/B Range: 220 V
		20 Hz to < 40 Hz	$45 \cdot 10^{-6} \cdot U$	
		40 Hz to 20 kHz	$19 \cdot 10^{-6} \cdot U$	
		> 20 kHz to 50 kHz	$45 \cdot 10^{-6} \cdot U$	
		> 50 kHz to 100 kHz	$64 \cdot 10^{-6} \cdot U$	
		> 100 kHz to 300 kHz	$0.14 \cdot 10^{-3} \cdot U$	
		> 300 kHz to 500 kHz	$0.29 \cdot 10^{-3} \cdot U$	
	190 V to 700 V	10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	<i>U</i> : measured value 5790A/B Range: 700 V
		20 Hz to < 40 Hz	$64 \cdot 10^{-6} \cdot U$	
		40 Hz to 20 kHz	$26 \cdot 10^{-6} \cdot U$	
		> 20 kHz to 50 kHz	$80 \cdot 10^{-6} \cdot U$	
		> 50 kHz to 100 kHz	$0.27 \cdot 10^{-3} \cdot U$	
	600 V to 1050 V	10 Hz to < 20 Hz	$0.13 \cdot 10^{-3} \cdot U$	<i>U</i> : measured value 5790A/B Range: 1100 V
		20 Hz to < 40 Hz	$64 \cdot 10^{-6} \cdot U$	
		40 Hz to 20 kHz	$25 \cdot 10^{-6} \cdot U$	
		> 20 kHz to 50 kHz	$80 \cdot 10^{-6} \cdot U$	
		> 50 kHz to 100 kHz	$0.27 \cdot 10^{-3} \cdot U$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks	
AC/DC voltage- transfer	2 mV	10 Hz	$0.11 \cdot 10^{-3}$	Voltage range: 22 mV Thermal Transfer Standards Fluke 792A	
		20 Hz	$90 \cdot 10^{-6}$		
		30 Hz; 40 Hz; 55 Hz	$85 \cdot 10^{-6}$		
		60 Hz; 120 Hz; 300 Hz	$85 \cdot 10^{-6}$		
AC/DC voltage- transfer		400 Hz; 500 Hz; 1 kHz	$85 \cdot 10^{-6}$	Voltage range: 22 mV Thermal Transfer Standards Fluke 792A	
		10 kHz; 20 kHz	$85 \cdot 10^{-6}$		
		30 kHz; 50 kHz; 70 kHz	$85 \cdot 10^{-6}$		
		100 kHz	$90 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$0.11 \cdot 10^{-3}$		
		500 kHz	$0.12 \cdot 10^{-3}$		
		700 kHz; 800 kHz	$0.15 \cdot 10^{-3}$		
		1 MHz	$0.19 \cdot 10^{-3}$		
		10 Hz	$90 \cdot 10^{-6}$		
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$65 \cdot 10^{-6}$		
AC/DC voltage- transfer	6 mV	60 Hz; 120 Hz; 300 Hz	$65 \cdot 10^{-6}$	Voltage range: 22 mV Thermal Transfer Standards Fluke 792A	
		400 Hz; 500 Hz	$60 \cdot 10^{-6}$		
		1 kHz; 10 kHz	$60 \cdot 10^{-6}$		
		20 kHz; 30 kHz	$55 \cdot 10^{-6}$		
		50 kHz; 70 kHz	$55 \cdot 10^{-6}$		
		100 kHz	$60 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$75 \cdot 10^{-6}$		
		500 kHz	$90 \cdot 10^{-6}$		
		700 kHz; 800 kHz	$0.12 \cdot 10^{-3}$		
		1 MHz	$0.14 \cdot 10^{-3}$		
AC/DC voltage- transfer	10 mV	10 Hz	$75 \cdot 10^{-6}$	Voltage range: 22 mV Thermal Transfer Standards Fluke 792A	
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$55 \cdot 10^{-6}$		
		60 Hz; 120 Hz; 300 Hz	$55 \cdot 10^{-6}$		
		400 Hz; 500 Hz	$50 \cdot 10^{-6}$		
		1 kHz; 10 kHz	$50 \cdot 10^{-6}$		
		20 kHz; 30 kHz	$45 \cdot 10^{-6}$		
		50 kHz; 70 kHz	$45 \cdot 10^{-6}$		
		100 kHz	$55 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$65 \cdot 10^{-6}$		
		500 kHz	$75 \cdot 10^{-6}$		
AC/DC voltage- transfer		700 kHz; 800 kHz	$95 \cdot 10^{-6}$	Voltage range: 22 mV Thermal Transfer Standards Fluke 792A	
		1 MHz	$0.12 \cdot 10^{-3}$		
20 mV	10 Hz	$60 \cdot 10^{-6}$			
	20 Hz; 30 Hz; 40 Hz; 55 Hz	$40 \cdot 10^{-6}$			
	60 Hz; 120 Hz; 300 Hz	$40 \cdot 10^{-6}$			
	400 Hz; 500 Hz	$35 \cdot 10^{-6}$			
	1 kHz; 10 kHz; 20 kHz	$35 \cdot 10^{-6}$			
	30 kHz; 50 kHz	$30 \cdot 10^{-6}$			
	70 kHz; 100 kHz	$30 \cdot 10^{-6}$			
	200 kHz; 300 kHz	$40 \cdot 10^{-6}$			
	500 kHz	$50 \cdot 10^{-6}$			
	700 kHz; 800 kHz	$70 \cdot 10^{-6}$			
	1 MHz	$80 \cdot 10^{-6}$			

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC/DC voltage- transfer	60 mV	10 Hz	$35 \cdot 10^{-6}$	Voltage range: 220 mV Thermal Transfer Standards Fluke 792A
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$25 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$25 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$25 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 30 kHz	$25 \cdot 10^{-6}$	
		50 kHz; 70 kHz; 100 kHz	$25 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$40 \cdot 10^{-6}$	
		500 kHz	$50 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$70 \cdot 10^{-6}$	
		1 MHz	$80 \cdot 10^{-6}$	
AC/DC voltage- transfer	100 mV	10 Hz	$15 \cdot 10^{-6}$	Voltage range: 220 mV Thermal Transfer Standards Fluke 792A
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$8 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 30 kHz	$8 \cdot 10^{-6}$	
		50 kHz; 70 kHz; 100 kHz	$8 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	
AC/DC voltage- transfer	200 mV	10 Hz	$15 \cdot 10^{-6}$	Voltage range: 220 mV Thermal Transfer Standards Fluke 792A
		20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$11 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$11 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$7 \cdot 10^{-6}$	
		1 kHz; 10 kHz; 20 kHz	$7 \cdot 10^{-6}$	
		30 kHz; 50 kHz; 70 kHz	$7 \cdot 10^{-6}$	
		100 kHz	$8 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$20 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	
		10 Hz	$12 \cdot 10^{-6}$	Voltage range: 700 mV
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$11 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$11 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$7 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 30 kHz	$7 \cdot 10^{-6}$	
		50 kHz; 70 kHz	$7 \cdot 10^{-6}$	
		100 kHz	$8 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$18 \cdot 10^{-6}$	
		500 kHz	$30 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$40 \cdot 10^{-6}$	
		1 MHz	$45 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC /DC voltage- transfer	300 mV; 400 mV; 500 mV; 600 mV; 700 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz 120 Hz; 300 Hz 400 Hz; 500 Hz; 1 kHz 10 kHz; 20 kHz; 30 kHz 50 kHz; 70 kHz 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	11 · 10 ⁻⁶ 11 · 10 ⁻⁶ 11 · 10 ⁻⁶ 7 · 10 ⁻⁶ 7 · 10 ⁻⁶ 7 · 10 ⁻⁶ 8 · 10 ⁻⁶ 18 · 10 ⁻⁶ 30 · 10 ⁻⁶ 40 · 10 ⁻⁶ 45 · 10 ⁻⁶	Voltage range: 700 mV Thermal Transfer Standards Fluke 792A
	600 mV	10 Hz; 20 Hz 30 Hz; 40 Hz; 55 Hz 60 Hz; 120 Hz; 300 Hz 400 Hz; 500 Hz 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	11 · 10 ⁻⁶ 11 · 10 ⁻⁶ 11 · 10 ⁻⁶ 4 · 10 ⁻⁶ 4 · 10 ⁻⁶ 5 · 10 ⁻⁶ 6 · 10 ⁻⁶ 8 · 10 ⁻⁶ 18 · 10 ⁻⁶ 30 · 10 ⁻⁶ 40 · 10 ⁻⁶ 45 · 10 ⁻⁶	Voltage range: 2.2 V
	1 V; 2 V	10 Hz 20 Hz 30 Hz; 40 Hz; 55 Hz 60 Hz; 120 Hz 300 Hz ; 400 Hz; 500 Hz 1 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	4 · 10 ⁻⁶ 3 · 10 ⁻⁶ 2 · 10 ⁻⁶ 2 · 10 ⁻⁶ 2 · 10 ⁻⁶ 2 · 10 ⁻⁶ 3 · 10 ⁻⁶ 4 · 10 ⁻⁶ 5 · 10 ⁻⁶ 7 · 10 ⁻⁶ 9 · 10 ⁻⁶ 15 · 10 ⁻⁶	
	2 V; 3 V; 4 V; 5 V; 6 V; 7 V	10 Hz 20 Hz 30 Hz; 40 Hz; 55 Hz 60 Hz; 120 Hz 300 Hz ; 400 Hz; 500 Hz 1 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz 800 kHz 1 MHz	4 · 10 ⁻⁶ 3 · 10 ⁻⁶ 2 · 10 ⁻⁶ 4 · 10 ⁻⁶ 5 · 10 ⁻⁶ 6 · 10 ⁻⁶ 8 · 10 ⁻⁶ 9 · 10 ⁻⁶ 15 · 10 ⁻⁶	Voltage range: 7 V

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks	
AC / DC voltage- transfer	6 V; 8 V	10 Hz	$4 \cdot 10^{-6}$	Voltage range: 22 V Thermal Transfer Standards Fluke 792A	
		20 Hz	$3 \cdot 10^{-6}$		
		30 Hz; 40 Hz; 55 Hz	$2 \cdot 10^{-6}$		
		60 Hz; 120 Hz	$2 \cdot 10^{-6}$		
AC / DC voltage- transfer		300 Hz ; 400 Hz; 500 Hz	$2 \cdot 10^{-6}$	Voltage range: 22 V Thermal Transfer Standards Fluke 792A	
		1 kHz; 10 kHz; 20 kHz	$2 \cdot 10^{-6}$		
		30 kHz; 50 kHz; 70 kHz	$2 \cdot 10^{-6}$		
		100 kHz	$4 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$5 \cdot 10^{-6}$		
		500 kHz	$7 \cdot 10^{-6}$		
		700 kHz; 800 kHz	$9 \cdot 10^{-6}$		
		1 MHz	$15 \cdot 10^{-6}$		
		10 Hz	$4 \cdot 10^{-6}$		
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$3 \cdot 10^{-6}$		
AC / DC voltage- transfer	10 V	60 Hz; 120 Hz	$3 \cdot 10^{-6}$	Voltage range: 22 V Thermal Transfer Standards Fluke 792A	
		300 Hz; 400 Hz; 500 Hz	$3 \cdot 10^{-6}$		
		1 kHz; 10 kHz; 20 kHz	$3 \cdot 10^{-6}$		
		30 kHz; 50 kHz; 70 kHz	$3 \cdot 10^{-6}$		
		100 kHz	$4 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$5 \cdot 10^{-6}$		
		500 kHz	$7 \cdot 10^{-6}$		
		700 kHz; 800 kHz	$9 \cdot 10^{-6}$		
		1 MHz	$15 \cdot 10^{-6}$		
		10 Hz; 20 Hz; 30 Hz; 40 Hz	$6 \cdot 10^{-6}$		
AC / DC voltage- transfer	20 V	55 Hz; 60 Hz; 120 Hz	$6 \cdot 10^{-6}$	Voltage range: 70 V	
		300 Hz ; 400 Hz; 500 Hz	$6 \cdot 10^{-6}$		
		1 kHz; 10 kHz; 20 kHz	$6 \cdot 10^{-6}$		
		30 kHz; 50 kHz; 70 kHz	$3 \cdot 10^{-6}$		
		100 kHz	$7 \cdot 10^{-6}$		
		200 kHz; 300 kHz	$9 \cdot 10^{-6}$		
		500 kHz	$12 \cdot 10^{-6}$		
		700 kHz; 800 kHz	$18 \cdot 10^{-6}$		
		1 MHz	$25 \cdot 10^{-6}$		
		10 Hz; 20 Hz; 30 Hz; 40 Hz	$6 \cdot 10^{-6}$		

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC / DC voltage- transfer	30 V	10 Hz	$7 \cdot 10^{-6}$	Voltage range: 70 V Thermal Transfer Standards Fluke 792A
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$6 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$6 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$6 \cdot 10^{-6}$	
		10 kHz; 20 kHz; 30 kHz	$6 \cdot 10^{-6}$	
		50 kHz; 70 kHz	$6 \cdot 10^{-6}$	
		100 kHz	$8 \cdot 10^{-6}$	
		200 kHz; 300 kHz	$12 \cdot 10^{-6}$	
		500 kHz	$15 \cdot 10^{-6}$	
		700 kHz; 800 kHz	$25 \cdot 10^{-6}$	
		1 MHz	$35 \cdot 10^{-6}$	
	40 V; 50 V; 60 V; 70 V	10 Hz	$8 \cdot 10^{-6}$	Voltage range: 220 V
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$7 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$7 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$6 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$6 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$6 \cdot 10^{-6}$	
		70 kHz	$8 \cdot 10^{-6}$	
		100 kHz	$12 \cdot 10^{-6}$	
		10 Hz	$8 \cdot 10^{-6}$	
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$7 \cdot 10^{-6}$	
	60 V	60 Hz; 120 Hz; 300 Hz	$7 \cdot 10^{-6}$	Voltage range: 220 V
		400 Hz; 500 Hz; 1 kHz	$6 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$6 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$6 \cdot 10^{-6}$	
		70 kHz	$8 \cdot 10^{-6}$	
		100 kHz	$12 \cdot 10^{-6}$	
	100 V	10 Hz	$12 \cdot 10^{-6}$	Voltage range: 220 V
		20 Hz; 30 Hz; 40 Hz; 55 Hz	$11 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$11 \cdot 10^{-6}$	
		400 Hz; 500 Hz; 1 kHz	$11 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$11 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$11 \cdot 10^{-6}$	
		70 kHz	$18 \cdot 10^{-6}$	
		100 kHz	$30 \cdot 10^{-6}$	
		10 Hz; 20 Hz	$12 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
	200 V	60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	Voltage range: 1000 V
		400 Hz; 500 Hz; 1 kHz	$11 \cdot 10^{-6}$	
		10 kHz; 20 kHz	$11 \cdot 10^{-6}$	
		30 kHz; 50 kHz	$11 \cdot 10^{-6}$	
		70 kHz	$18 \cdot 10^{-6}$	
		100 kHz	$30 \cdot 10^{-6}$	
	200 V	10 Hz	$15 \cdot 10^{-6}$	Voltage range: 1000 V
		20 Hz	$13 \cdot 10^{-6}$	
		30 Hz; 40 Hz; 55 Hz	$12 \cdot 10^{-6}$	
		60 Hz; 120 Hz; 300 Hz	$12 \cdot 10^{-6}$	
		400 Hz; 500 Hz	$11 \cdot 10^{-6}$	
		1 kHz; 10 kHz	$11 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$11 \cdot 10^{-6}$	
		70 kHz	$18 \cdot 10^{-6}$	
		100 kHz	$30 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC/DC voltage-transfer	300 V	10 Hz	$15 \cdot 10^{-6}$	Voltage range: 1000 V Thermal Transfer Standards Fluke 792A
		20 Hz; 30 Hz	$12 \cdot 10^{-6}$	
		40 Hz; 55 Hz; 60 Hz	$11 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$11 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$11 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$11 \cdot 10^{-6}$	
		70 kHz	$18 \cdot 10^{-6}$	
		100 kHz	$30 \cdot 10^{-6}$	
		10 Hz	$17 \cdot 10^{-6}$	
AC voltage	500 V	20 Hz; 30 Hz	$12 \cdot 10^{-6}$	Square wave voltage, Triangular voltage at 50Ω and $1\text{ M}\Omega$
		40 Hz; 55 Hz; 60 Hz	$11 \cdot 10^{-6}$	
		120 Hz; 300 Hz; 400 Hz	$11 \cdot 10^{-6}$	
		500 Hz; 1 kHz; 10 kHz	$11 \cdot 10^{-6}$	
		20 kHz; 30 kHz; 50 kHz	$11 \cdot 10^{-6}$	
		70 kHz	$18 \cdot 10^{-6}$	
		100 kHz	$35 \cdot 10^{-6}$	
	1 mV	10 Hz to 40 Hz	$0.50 \cdot 10^{-3}$	
		55 Hz to 10 kHz	$0.50 \cdot 10^{-3}$	
		10 kHz to 100 kHz	$0.50 \cdot 10^{-3}$	
		200 kHz to 700 kHz	$0.70 \cdot 10^{-3}$	
AC voltage	2 mV	700 kHz to 1 MHz	$0.90 \cdot 10^{-3}$	Square wave voltage, Triangular voltage at 50Ω and $1\text{ M}\Omega$
		10 Hz to 40 Hz	$0.30 \cdot 10^{-3}$	
		55 Hz to 10 kHz	$0.30 \cdot 10^{-3}$	
		10 kHz to 100 kHz	$0.30 \cdot 10^{-3}$	
		200 kHz to 700 kHz	$0.50 \cdot 10^{-3}$	
	6 mV; 10 mV; 20 mV	700 kHz to 1 MHz	$0.80 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$0.20 \cdot 10^{-3}$	
		55 Hz to 10 kHz	$0.20 \cdot 10^{-3}$	
		10 kHz to 100 kHz	$0.20 \cdot 10^{-3}$	
		200 kHz to 700 kHz	$0.50 \cdot 10^{-3}$	
AC voltage	40 mV	700 kHz to 1 MHz	$0.80 \cdot 10^{-3}$	Square wave voltage, Triangular voltage at 50Ω and $1\text{ M}\Omega$
		10 Hz to 40 Hz	$0.20 \cdot 10^{-3}$	
		55 Hz to 10 kHz	$0.20 \cdot 10^{-3}$	
		10 kHz to 100 kHz	$0.20 \cdot 10^{-3}$	
		200 kHz to 700 kHz	$0.50 \cdot 10^{-3}$	
	60 mV	700 kHz to 1 MHz	$0.80 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$0.15 \cdot 10^{-3}$	
		55 Hz to 10 kHz	$0.15 \cdot 10^{-3}$	
		10 kHz to 100 kHz	$0.15 \cdot 10^{-3}$	
		200 kHz to 700 kHz	$0.50 \cdot 10^{-3}$	
		700 kHz to 1 MHz	$0.80 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC voltage	100 mV to 30 V	10 Hz to 40 Hz 55 Hz to 10 kHz 10 kHz to 100 kHz 200 kHz to 700 kHz 700 kHz to 1 MHz	$0.10 \cdot 10^{-3}$ $0.10 \cdot 10^{-3}$ $0.10 \cdot 10^{-3}$ $0.45 \cdot 10^{-3}$ $0.80 \cdot 10^{-3}$	Square wave voltage, Triangular voltage at 50 Ω and 1 MΩ
AC voltage: Square wave voltage. Triangular voltage	40 V to 100 V	10 Hz to 40 Hz 55 Hz to 10 kHz 10 kHz to 100 kHz	$0.10 \cdot 10^{-3}$ $0.10 \cdot 10^{-3}$ $0.10 \cdot 10^{-3}$	Square wave voltage, Triangular voltage at 50 Ω and 1 MΩ
AC/DC current transfer	300 μA; 1 mA; 3 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$15 \cdot 10^{-6}$ $11 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $30 \cdot 10^{-6}$ $35 \cdot 10^{-6}$	
	5 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$13 \cdot 10^{-6}$ $13 \cdot 10^{-6}$ $22 \cdot 10^{-6}$ $32 \cdot 10^{-6}$ $39 \cdot 10^{-6}$ $49 \cdot 10^{-6}$ $67 \cdot 10^{-6}$	
	10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$15 \cdot 10^{-6}$ $11 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $25 \cdot 10^{-6}$	
	20 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$13 \cdot 10^{-6}$ $13 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $14 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $20 \cdot 10^{-6}$ $22 \cdot 10^{-6}$	
	30 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$13 \cdot 10^{-6}$ $13 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $18 \cdot 10^{-6}$ $22 \cdot 10^{-6}$ $29 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC/DC current transfer	50 mA	10 Hz to 40 Hz	$13 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$13 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$15 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$15 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$15 \cdot 10^{-6}$	
		> 70 kHz to 100 kHz	$18 \cdot 10^{-6}$	
	100 mA	10 Hz to 40 Hz	$14 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$14 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$14 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$15 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$16 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$16 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$16 \cdot 10^{-6}$		
	200 mA	10 Hz to 40 Hz	$15 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$15 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$15 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$16 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$17 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$17 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$17 \cdot 10^{-6}$		
	300 mA	10 Hz to 40 Hz	$13 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$13 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$13 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$13 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$17 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$24 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$28 \cdot 10^{-6}$		
	500 mA	10 Hz to 40 Hz	$13 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$13 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$13 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$13 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$14 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$17 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$17 \cdot 10^{-6}$		
	1 A	10 Hz to 40 Hz	$13 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$13 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$13 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$17 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$17 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$17 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$21 \cdot 10^{-6}$		
	2 A	10 Hz to 40 Hz	$15 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$15 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$15 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$17 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$21 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$29 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$43 \cdot 10^{-6}$		

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC/DC current transfer	3 A	10 Hz to 40 Hz	$21 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$21 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$21 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$27 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$46 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$64 \cdot 10^{-6}$	
	5 A	> 70 kHz to 100 kHz	$90 \cdot 10^{-6}$	
		10 Hz to 40 Hz	$20 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$20 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$20 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$21 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$30 \cdot 10^{-6}$	
	10 A	> 50 kHz to 70 kHz	$45 \cdot 10^{-6}$	
		> 70 kHz to 100 kHz	$65 \cdot 10^{-6}$	
		10 Hz to 40 Hz	$24 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$25 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$25 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$25 \cdot 10^{-6}$	
	20 A	> 20 kHz to 50 kHz	$53 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$64 \cdot 10^{-6}$	
		> 70 kHz to 100 kHz	$84 \cdot 10^{-6}$	
		10 Hz to 40 Hz	$34 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$34 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$43 \cdot 10^{-6}$	
	50 A	> 10 kHz to 20 kHz	$43 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$64 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$86 \cdot 10^{-6}$	
		> 70 kHz to 100 kHz	$0.11 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$45 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$45 \cdot 10^{-6}$	
	100 A	> 1 kHz to 10 kHz	$56 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$56 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$76 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$0.11 \cdot 10^{-3}$	
		> 70 kHz to 100 kHz	$0.15 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$64 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$64 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$81 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$83 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$93 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$0.14 \cdot 10^{-3}$	
		> 70 kHz to 100 kHz	$0.17 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC current calibration generators and sources	100 µA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$65 \cdot 10^{-6}$ $60 \cdot 10^{-6}$ $65 \cdot 10^{-6}$ $70 \cdot 10^{-6}$ $75 \cdot 10^{-6}$ $80 \cdot 10^{-6}$ $0.10 \cdot 10^{-3}$	
	300 µA; 1 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$25 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $35 \cdot 10^{-6}$ $45 \cdot 10^{-6}$ $50 \cdot 10^{-6}$ $80 \cdot 10^{-6}$	
	3 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$20 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $35 \cdot 10^{-6}$ $45 \cdot 10^{-6}$ $50 \cdot 10^{-6}$ $80 \cdot 10^{-6}$	
	5 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$26 \cdot 10^{-6}$ $24 \cdot 10^{-6}$ $29 \cdot 10^{-6}$ $36 \cdot 10^{-6}$ $42 \cdot 10^{-6}$ $49 \cdot 10^{-6}$ $68 \cdot 10^{-6}$	
	10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$19 \cdot 10^{-6}$ $17 \cdot 10^{-6}$ $23 \cdot 10^{-6}$ $31 \cdot 10^{-6}$ $39 \cdot 10^{-6}$ $44 \cdot 10^{-6}$ $64 \cdot 10^{-6}$	
	20 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$19 \cdot 10^{-6}$ $17 \cdot 10^{-6}$ $23 \cdot 10^{-6}$ $31 \cdot 10^{-6}$ $38 \cdot 10^{-6}$ $44 \cdot 10^{-6}$ $65 \cdot 10^{-6}$	
	30 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	$19 \cdot 10^{-6}$ $17 \cdot 10^{-6}$ $24 \cdot 10^{-6}$ $31 \cdot 10^{-6}$ $39 \cdot 10^{-6}$ $45 \cdot 10^{-6}$ $67 \cdot 10^{-6}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC current calibration generators and sources	50 mA	10 Hz to 40 Hz	$19 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$17 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$24 \cdot 10^{-6}$	
		> 10 kHz to 20 kHz	$31 \cdot 10^{-6}$	
		> 20 kHz to 50 kHz	$38 \cdot 10^{-6}$	
		> 50 kHz to 70 kHz	$42 \cdot 10^{-6}$	
		> 70 kHz to 100 kHz	$63 \cdot 10^{-6}$	
	100 mA	10 Hz to 40 Hz	$20 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$17 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$23 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$31 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$38 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$43 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$63 \cdot 10^{-6}$		
	200 mA	10 Hz to 40 Hz	$21 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$18 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$24 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$32 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$39 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$43 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$63 \cdot 10^{-6}$		
	300 mA	10 Hz to 40 Hz	$19 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$17 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$23 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$30 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$42 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$49 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$69 \cdot 10^{-6}$		
	500 mA	10 Hz to 40 Hz	$19 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$17 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$23 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$30 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$40 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$46 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$65 \cdot 10^{-6}$		
	1 A	10 Hz to 40 Hz	$20 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$18 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$23 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$33 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$42 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$46 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$66 \cdot 10^{-6}$		
	2 A	10 Hz to 40 Hz	$22 \cdot 10^{-6}$	
	> 40 Hz to 1 kHz	$18 \cdot 10^{-6}$		
	> 1 kHz to 10 kHz	$25 \cdot 10^{-6}$		
	> 10 kHz to 20 kHz	$32 \cdot 10^{-6}$		
	> 20 kHz to 50 kHz	$44 \cdot 10^{-6}$		
	> 50 kHz to 70 kHz	$52 \cdot 10^{-6}$		
	> 70 kHz to 100 kHz	$76 \cdot 10^{-6}$		

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC current calibration generators and sources	3 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	27 · 10 ⁻⁶ 25 · 10 ⁻⁶ 29 · 10 ⁻⁶ 36 · 10 ⁻⁶ 53 · 10 ⁻⁶ 67 · 10 ⁻⁶ 94 · 10 ⁻⁶	
	5 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	26 · 10 ⁻⁶ 25 · 10 ⁻⁶ 29 · 10 ⁻⁶ 36 · 10 ⁻⁶ 52 · 10 ⁻⁶ 65 · 10 ⁻⁶ 92 · 10 ⁻⁶	
	10A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	30 · 10 ⁻⁶ 29 · 10 ⁻⁶ 33 · 10 ⁻⁶ 38 · 10 ⁻⁶ 71 · 10 ⁻⁶ 81 · 10 ⁻⁶ 0.11 · 10 ⁻³	
	20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	38 · 10 ⁻⁶ 37 · 10 ⁻⁶ 48 · 10 ⁻⁶ 52 · 10 ⁻⁶ 79 · 10 ⁻⁶ 0.10 · 10 ⁻³ 0.13 · 10 ⁻³	
	30 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	61 · 10 ⁻⁶ 61 · 10 ⁻⁶ 70 · 10 ⁻⁶ 78 · 10 ⁻⁶ 98 · 10 ⁻⁶ 0.13 · 10 ⁻³ 0.17 · 10 ⁻³	
	50 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	61 · 10 ⁻⁶ 61 · 10 ⁻⁶ 70 · 10 ⁻⁶ 78 · 10 ⁻⁶ 98 · 10 ⁻⁶ 0.13 · 10 ⁻³ 0.17 · 10 ⁻³	
	80 A; 100 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	68 · 10 ⁻⁶ 68 · 10 ⁻⁶ 85 · 10 ⁻⁶ 92 · 10 ⁻⁶ 0.10 · 10 ⁻³ 0.15 · 10 ⁻³ 0.19 · 10 ⁻³	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC current measurement instruments	100 µA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	0.10 · 10 ⁻³ 95 · 10 ⁻⁶ 0.15 · 10 ⁻³ 0.15 · 10 ⁻³ 0.16 · 10 ⁻³ 0.16 · 10 ⁻³ 0.17 · 10 ⁻³	
	300 µA; 1 mA; 3 mA; 5 mA; 10 mA; 20 mA; 30 mA; 50 mA; 100 mA; 200 mA; 300 mA; 500 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	75 · 10 ⁻⁶ 65 · 10 ⁻⁶ 0.14 · 10 ⁻³ 0.14 · 10 ⁻³ 0.14 · 10 ⁻³ 0.15 · 10 ⁻³ 0.16 · 10 ⁻³	
	1 A; 2 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	75 · 10 ⁻⁶ 70 · 10 ⁻⁶ 0.14 · 10 ⁻³ 0.14 · 10 ⁻³ 0.15 · 10 ⁻³ 0.15 · 10 ⁻³ 0.16 · 10 ⁻³	
	3 A; 5 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	80 · 10 ⁻⁶ 75 · 10 ⁻⁶ 0.14 · 10 ⁻³ 0.14 · 10 ⁻³ 0.15 · 10 ⁻³ 0.16 · 10 ⁻³ 0.17 · 10 ⁻³	
	10A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	80 · 10 ⁻⁶ 75 · 10 ⁻⁶ 0.14 · 10 ⁻³ 0.15 · 10 ⁻³ 0.16 · 10 ⁻³ 0.17 · 10 ⁻³ 0.19 · 10 ⁻³	
	20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 70 kHz > 70 kHz to 100 kHz	90 · 10 ⁻⁶ 80 · 10 ⁻⁶ 0.14 · 10 ⁻³ 0.15 · 10 ⁻³ 0.17 · 10 ⁻³ 0.18 · 10 ⁻³ 0.20 · 10 ⁻³	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC current measurement instruments	30 A	10 Hz to 40 Hz	$90 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$85 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$0.15 \cdot 10^{-3}$	
		> 20 kHz to 30 kHz	$0.16 \cdot 10^{-3}$	
		50 kHz	$0.17 \cdot 10^{-3}$	
	50 A	70 kHz	$0.19 \cdot 10^{-3}$	
		100 kHz	$0.23 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$90 \cdot 10^{-6}$	
		> 40 Hz to 1 kHz	$85 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$0.15 \cdot 10^{-3}$	
Inductance measurement instruments	80 A	> 20 kHz to 30 kHz	$0.16 \cdot 10^{-3}$	
		50 kHz	$0.18 \cdot 10^{-3}$	
		70 kHz	$0.20 \cdot 10^{-3}$	
		100 kHz	$0.24 \cdot 10^{-3}$	
		10 Hz to 40 Hz	$0.10 \cdot 10^{-3}$	
	100 A	> 40 Hz to 1 kHz	$90 \cdot 10^{-6}$	
		> 1 kHz to 10 kHz	$0.16 \cdot 10^{-3}$	
		> 20 kHz to 30 kHz	$0.17 \cdot 10^{-3}$	
		50 kHz	$0.18 \cdot 10^{-3}$	
		70 kHz	$0.20 \cdot 10^{-3}$	
		100 kHz	$0.24 \cdot 10^{-3}$	
	100 µH	100 Hz; 400 Hz; 1000 Hz; 10 kHz	$0.5 \cdot 10^{-3}$	
	1 mH; 10 mH	100 Hz; 400 Hz; 1000 Hz; 10 kHz	$0.4 \cdot 10^{-3}$	
	100 mH	100 Hz; 400 Hz; 1000 Hz	$0.4 \cdot 10^{-3}$	
		10 kHz	$0.5 \cdot 10^{-3}$	
	1 H; 2 H; 5 H	100 Hz; 400 Hz; 1000 Hz	$0.4 \cdot 10^{-3}$	
	10 H	100 Hz; 400 Hz	$0.4 \cdot 10^{-3}$	
		1000 Hz	$0.6 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
Capacitance measurement instruments	1 pF	100 Hz	$0.12 \cdot 10^{-3} \cdot C + 16 \text{ fF}$	The absolute component has only to be considered for instruments with a trimming function
		1 kHz	$0.12 \cdot 10^{-3} \cdot C + 2 \text{ fF}$	
		10 kHz	$0.12 \cdot 10^{-3} \cdot C + 78 \text{ aF}$	
	10 pF	100 Hz	$70 \cdot 10^{-6} \cdot C + 16 \text{ fF}$	
		1 kHz	$20 \cdot 10^{-6} \cdot C + 2 \text{ fF}$	
		10 kHz	$80 \cdot 10^{-6} \cdot C + 78 \text{ aF}$	
		100 kHz	$0.1 \cdot 10^{-3} \cdot C$	
		1000 kHz	$0.4 \cdot 10^{-3} \cdot C$	
	100 pF	100 Hz	$70 \cdot 10^{-6} \cdot C + 16 \text{ fF}$	
		1 kHz	$20 \cdot 10^{-6} \cdot C + 2 \text{ fF}$	
		10 kHz	$80 \cdot 10^{-6} \cdot C$	
		100 kHz	$0.1 \cdot 10^{-3} \cdot C$	
		1000 kHz	$0.4 \cdot 10^{-3} \cdot C$	
	10 nF	100 Hz; 1 kHz; 10 kHz	$0.1 \cdot 10^{-3} \cdot C$	
Capacitance reference standard	100 nF..1 μF	100 Hz	$0.11 \cdot 10^{-3} \cdot C$	
		1 kHz	$0.11 \cdot 10^{-3} \cdot C$	
		10 kHz	$0.16 \cdot 10^{-3} \cdot C$	
		1 kHz; 10 kHz	$5,5 \cdot 10^{-3}$	
			$1.3 \cdot 10^{-3}$	
	> 100 nF to 110 μF		$0.33 \cdot 10^{-3}$	
			$0.26 \cdot 10^{-3}$	
			$0.35 \cdot 10^{-3}$	
Capacitance calibration generators	110 μF to 110 mF	DC	$0.3 \cdot 10^{-3}$	Calibration of capacitance Fluke 55x0A calibration generators

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
Oscilloscopes deflection vertical	5 mV to 30 mV	Square wave voltage 10 Hz to 10 kHz	0.03 %	t_0 = period of time markers
	> 30 mV to 300 mV	10 Hz to 10 kHz	0.02 %	
	> 300 mV to 200 V	10 Hz to 10 kHz	0.016 %	
	> 500 ps to 5 s	Time mark amplitude < 1 V	0.015 % · t_0 + 2 ps	
rise time	> 50 ps to 100 ps	Step amplitude < 1 V	16 ps	t_r = Rise time
	> 100 ps to 10 ms		$2 \cdot 10^{-2} \cdot t_r + 4$ ps	
Oscilloscope calibrator Deflection vertical	5 mV to 30 mV	Square wave voltage 10 Hz to 10 kHz	$0.23 \cdot 10^{-3}$	W_{tf} = relative trigger uncertainty
	> 30 mV to 300 mV	10 Hz to 10 kHz	$0.12 \cdot 10^{-3}$	
	> 300 mV to 200 V	10 Hz to 10 kHz	$40 \cdot 10^{-6}$	
	> 1 s to 5 s	Time mark amplitude > 0.5 V	$3 \cdot 10^{-11} + W_{tf}$	
horizontal	> 10 ns to 1 s		$2 \cdot 10^{-11} + W_{tf}$	
	> 500 ps to 10 ns		$3 \cdot 10^{-11}$	
	14 ps to 100 ps	Step amplitude < 1 V	3 ps	t_r = rise-time
	> 100 ps to 10 ms		$2 \cdot 10^{-2} \cdot t_r + 4$ ps	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
Frequency	1 MHz 5 MHz 10 MHz	Sinus. $U_{\text{eff}} \geq 0.5 \text{ V}$ measurement period > 24 h	$6 \cdot 10^{-12}$	Analoge phase time difference measurement
	0.01 Hz to 1 Hz	Sinus. $U_{\text{eff}} \geq 0.5 \text{ V}$ measurement period > 1000 s	$3 \cdot 10^{-11} + W_{\text{tf}}$	$W_{\text{tf}} = \text{rel. trigger uncertainty digital Frequency measurement on count basis}$
	1 Hz to 150 MHz	Sinus. $U_{\text{eff}} \geq 0.5 \text{ V}$ measurement period > 1000 s	$2 \cdot 10^{-11} + W_{\text{tf}}$	Digital Frequency measurement on count basis
	150 MHz to 4.7 GHz	Sinus. $U_{\text{eff}} \geq 0.5 \text{ V}$ measurement period > 1000 s	$3 \cdot 10^{-11}$	

On-site Calibration

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
DC voltage measurement instruments and sources	0.01 V to 0.22 V > 0.22 V to 2.2 V > 2.2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		$8 \cdot 10^{-6} \cdot U + 2 \mu\text{V}$ $5 \cdot 10^{-6} \cdot U + 2 \mu\text{V}$ $4 \cdot 10^{-6} \cdot U + 8 \mu\text{V}$ $4 \cdot 10^{-6} \cdot U + 5 \mu\text{V}$ $7 \cdot 10^{-6} \cdot U + 60 \mu\text{V}$ $8 \cdot 10^{-6} \cdot U + 0.7 \text{ mV}$	$U = \text{measured value}$
DC current	10 μA to 220 μA > 220 μA to 2.2 mA > 2.2 mA to 22 mA > 22 mA to 220 mA > 0.22 A to 2.2 A > 2.2 A to 10 A		$80 \cdot 10^{-6}$ $50 \cdot 10^{-6}$ $50 \cdot 10^{-6}$ $60 \cdot 10^{-6} \cdot I + 1 \mu\text{A}$ $80 \cdot 10^{-6} \cdot I + 25 \mu\text{A}$ $4 \cdot 10^{-4} \cdot I + 0.48 \text{ mA}$	$I = \text{measured value}$
Current clamps	0 A to 1000 A		$5 \cdot 10^{-3}$	
DC resistance measurement instruments and reference standards	0.1 Ω to 1.9 Ω > 1.9 Ω to 10 Ω > 10 Ω to 19 Ω > 19 Ω to 190 Ω > 190 Ω to 1.9 k Ω > 1.9 k Ω to 19 k Ω > 19 k Ω to 190 k Ω > 190 k Ω to 1.9 M Ω > 1.9 M Ω to 10 M Ω > 10 M Ω to 19 M Ω > 19 M Ω to 100 M Ω		$85 \cdot 10^{-6}$ $30 \cdot 10^{-6}$ $25 \cdot 10^{-6}$ $20 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $12 \cdot 10^{-6}$ $15 \cdot 10^{-6}$ $20 \cdot 10^{-6}$ $35 \cdot 10^{-6}$ $70 \cdot 10^{-6}$ $0.12 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Valid from: 21.06.2023

Date of issue: 21.06.2023

Page 29 of 31

On-site Calibration

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC Voltage measurement instruments and sources	60 mV to 220 mV	10 Hz to 20 Hz	$0.84 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0.38 \cdot 10^{-3}$	
		> 40 Hz to 20 kHz	$0.27 \cdot 10^{-3}$	
		> 20 kHz to 50 kHz	$0.48 \cdot 10^{-3}$	
		> 50 kHz to 100 kHz	$1.4 \cdot 10^{-3}$	
	> 220 mV to 2.2 V	> 100 kHz to 300 kHz	$1.7 \cdot 10^{-3}$	
		> 300 kHz to 500 kHz	$2.5 \cdot 10^{-3}$	
		> 500 kHz to 1 MHz	$5.2 \cdot 10^{-3}$	
		10 Hz to 20 Hz	$0.95 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0.3 \cdot 10^{-3}$	
AC Voltage measurement instruments and sources	> 2.2 V to 22 V	> 40 Hz to 20 kHz	$0.11 \cdot 10^{-3}$	
		> 20 kHz to 50 kHz	$0.22 \cdot 10^{-3}$	
		> 50 kHz to 100 kHz	$0.62 \cdot 10^{-3}$	
		> 100 kHz to 300 kHz	$1.2 \cdot 10^{-3}$	
		> 300 kHz to 500 kHz	$3.0 \cdot 10^{-3}$	
	> 22 V to 220 V	> 500 kHz to 1 MHz	$6.7 \cdot 10^{-3}$	
		10 Hz to 20 Hz	$0.95 \cdot 10^{-3}$	
		> 20 Hz to 40 Hz	$0.3 \cdot 10^{-3}$	
		> 40 Hz to 20 kHz	$0.13 \cdot 10^{-3}$	
		> 20 kHz to 50 kHz	$0.41 \cdot 10^{-3}$	
	> 220 V to 1100 V	> 50 kHz to 100 kHz	$0.95 \cdot 10^{-3}$	
		10 Hz to 50 Hz	$0.50 \cdot 10^{-3}$	
		> 50 Hz to 1 kHz	$0.15 \cdot 10^{-3}$	
		> 1 kHz to 20 kHz	$0.20 \cdot 10^{-3}$	
	> 220 V to 750 V	> 20 kHz to 30 kHz	$0.50 \cdot 10^{-3}$	
		30 kHz to 50 kHz	$0.5 \cdot 10^{-3}$	
		> 50 kHz to 100 kHz	$1.8 \cdot 10^{-3}$	
		30 kHz to 50 kHz	$0.5 \cdot 10^{-3}$	
		> 50 kHz to 100 kHz	$1.8 \cdot 10^{-3}$	
		30 kHz to 50 kHz	$0.5 \cdot 10^{-3}$	

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

On-site Calibration

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
AC Current measurement instruments and sources	10 µA to 220 µA	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	0.7 · 10 ⁻³ 0.36 · 10 ⁻³ 0.14 · 10 ⁻³ 0.59 · 10 ⁻³ 1.7 · 10 ⁻³	I = measured value
	> 220 µA to 2.2 mA	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	0.7 · 10 ⁻³ 0.36 · 10 ⁻³ 0.14 · 10 ⁻³ $0.59 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$ $1.7 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	
	> 2.2 mA to 22 mA	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	0.7 · 10 ⁻³ · $I + 1 \mu\text{A}$ 0.36 · 10 ⁻³ · $I + 1 \mu\text{A}$ 0.14 · 10 ⁻³ · $I + 1 \mu\text{A}$ $0.59 \cdot 10^{-3} \cdot I + 5 \mu\text{A}$ $1.7 \cdot 10^{-3} \cdot I + 10 \mu\text{A}$	
	> 22 mA to 220 mA	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	0.7 · 10 ⁻³ · $I + 5 \mu\text{A}$ 0.36 · 10 ⁻³ · $I + 5 \mu\text{A}$ 0.14 · 10 ⁻³ · $I + 5 \mu\text{A}$ $0.59 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$ $1.7 \cdot 10^{-3} \cdot I + 0.1 \text{ mA}$	
	> 220 mA to 2.2 A	20 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.64 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$ $0.76 \cdot 10^{-3} \cdot I + 0.1 \text{ mA}$ $8.7 \cdot 10^{-3} \cdot I + 0.2 \text{ mA}$	
	> 2.2 A to 11 A	40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.47 \cdot 10^{-3} \cdot I + 0.2 \text{ mA}$ $1 \cdot 10^{-3} \cdot I + 0.5 \text{ mA}$ $3.8 \cdot 10^{-3} \cdot I + 0.9 \text{ mA}$	

Abbreviations used:

DIN Deutsches Institut für Normung e.V. – German institute for standardization

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.