

# Deutsche Akkreditierungsstelle

## Annex to the Partial Accreditation Certificate D-K-15219-01-02 according to DIN EN ISO/IEC 17025:2018

Valid from: 12.01.2024

Date of issue: 12.01.2024

This annex is a part of the accreditation certificate D-K-15219-01-00.

Holder of partial accreditation certificate:

**imetrologie GmbH**  
**Institut für Metrologie und Prozesstechnologie**  
**Luitpoldstraße 3, 97264 Helmstadt**

with the location

**imetrologie GmbH**  
**Institut für Metrologie und Prozesstechnologie**  
**Luitpoldstraße 3, 97264 Helmstadt**

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

*This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.*

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

Calibration in the fields:

**Thermodynamic quantities**

**Temperature quantities**

- Resistance thermometers <sup>b)</sup>
- Thermocouples <sup>b)</sup>
- Direct reading thermometers <sup>b)</sup>
- Temperature indicators and simulators <sup>b)</sup>
- Temperature transmitters, data loggers
- Fixed-point cells
- Liquid-in-glass thermometers
- Temperature block calibrators
- Climatic chambers (temperature) <sup>a)</sup>
- Calibration baths <sup>b)</sup>

**Humidity quantities**

- Devices for absolute humidity
- Devices for relative humidity
- Humidity generators and calibrators <sup>b)</sup>
- Climatic chambers (humidity) <sup>a)</sup>

**Mechanical quantities**

- Pressure <sup>b)</sup>

**Material testing machines (MTM)**

- Force (MTM) <sup>a)</sup>
- Extension (MTM) <sup>a)</sup>

<sup>a)</sup> only on-site calibration

<sup>b)</sup> also on-site calibration

Within the measurands/calibration items marked with <sup>\*</sup>), the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards/equivalent calibration procedures within the flexible scope of accreditation

Valid from: 12.01.2024

Date of issue: 12.01.2024

**Page 2 of 13**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range        | Measurement conditions / procedure                     | Expanded uncertainty of measurement | Remarks   |
|---|--------------|--|-------------------------------------|---|
| <b>Temperature</b><br>Fixed-point cells *)  | 0.01 °C      | G-ITS-90, Part 2.2:2018<br>triple point of water       | 0.5 mK                              | Comparison with reference fixed-point cells with standard resistance thermometers |
| Standard platinum resistance thermometers (SPRT), direct reading thermometers with SPRT sensor    | -196 °C      | QMV 7.2/13<br>liquid nitrogen                          | 15 mK                               | Comparison with reference thermometers  |
|   | -100 °C      | QMV 7.2/13<br>in liquid baths                          | 5.0 mK                              |   |
| Standard platinum resistance thermometers (SPRT), direct reading thermometers with SPRT sensor *) | -189.3442 °C | G-ITS-90, Part 2.3:2021<br>triple point of argon       | 4.0 mK                              | Calibration at fixed-point temperatures of ITS 90                                 |
|   | -38.8344 °C  | G-ITS-90, Part 2.4:2021<br>triple point of mercury     | 1.5 mK                              |   |
|   | 0.01 °C      | G-ITS-90, Part 2.2:2018<br>triple point of water       | 0.8 mK                              |   |
|   | 29.7646 °C   | G-ITS-90, Part 2.4:2021<br>melting point of gallium    | 1.5 mK                              |   |
|   | 156.5985 °C  | G-ITS-90, Part 2.4:2021<br>freezing point of indium    | 2.5 mK                              |   |
|   | 231.928 °C   | G-ITS-90, Part 2.4:2021<br>freezing point of tin       | 3.0 mK                              |   |
|   | 419.527 °C   | G-ITS-90, Part 2.4:2021<br>freezing point of zinc      | 3.0 mK                              |   |
|   | 660.323 °C   | G-ITS-90, Part 2.4:2021<br>freezing point of aluminium | 7.0 mK                              |   |

Valid from: 12.01.2024

Date of issue: 12.01.2024

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range                     | Measurement conditions / procedure                        | Expanded uncertainty of measurement | Remarks  |
|---|---------------------------|---|-------------------------------------|--|
| Standard platinum resistance thermometers (SPRT), direct reading thermometers with SPRT sensor    | -196 °C to 0 °C           | QMV 7.2/30 in liquid bath and at fixed-point temperatures | 20 mK                               | Calibration with determination of the characteristic   |
|   | -100 °C to 0 °C           |   | 7.0 mK                              | The measurement uncertainty is the uncertainty of the characteristic in the specified range          |
| Standard platinum resistance thermometers (SPRT), direct reading thermometers with SPRT sensor *) | -196 °C to -189.3442 °C   | EURAMET TG 01:2017  | 8.0 mK                              | Extrapolation  |
|   | -189.3442 °C to 0.01 °C   | G-ITS-90, Part 5:2021 fixed-points: Ar, Hg, TPW           | 6.0 mK                              | Calibration at fixed-point temperatures with determination of the characteristic according to ITS-90 |
|   | -38.8344 °C to 29.7646 °C | G-ITS-90, Part 5:2021 fixed-points: Hg, TPW, Ga           | 2.0 mK                              |  |
|   | 0 °C to 156.5985 °C       | G-ITS-90, Part 5:2021 fixed-points: TPW, In               | 3.5 mK                              | The measurement uncertainty is the uncertainty of the characteristic in the specified range          |
|   | 0 °C to 231.928 °C        | G-ITS-90, Part 5:2021 fixed-points: TPW, In, Sn           | 4.0 mK                              |  |
|   | 0 °C to 419.527 °C        | G-ITS-90, Part 5:2021 fixed-points: TPW, Sn, Zn           | 4.5 mK                              |  |
|   | 0 °C to 660.323 °C        | G-ITS-90, Part 5:2021 fixed-points: TPW, Sn, Zn, Al       | 8.0 mK                              |  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Annex to the Partial Accreditation Certificate D-K-15219-01-02

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item  | Range  | Measurement conditions / procedure                                  | Expanded uncertainty of measurement | Remarks                                |
|--|--|---|-------------------------------------|--|
| Resistance thermometers, direct reading thermometers with resistance sensor *) | -196 °C  | DKD-R 5-1:2018 liquid nitrogen                                      | 20 mK                               | Comparison with reference thermometers |
|  | -38.8344 °C  | DKD-R 5-1:2018 triple point of mercury                              | 5 mK                                |  |
|  | 0 °C   | DKD-R 5-1:2018 ice point  | 5 mK                                |  |
|  | 0.01 °C  | DKD-R 5-1:2018 triple point of water                                | 5 mK                                |  |
|  | 29.7646 °C   | DKD-R 5-1:2018 melting point of gallium                             | 5 mK                                |  |
|  | 156.5985 °C  | DKD-R 5-1:2018 freezing point of indium                             | 5 mK                                |  |
|  | -100 °C to -80 °C  | DKD-R 5-1:2018 in liquid baths                                      | 6 mK                                |  |
|  | > -80 °C to 100 °C   |   | 5 mK                                |  |
|  | > 100 °C to 160 °C   |   | 6 mK                                |  |
|  | > 160 °C to 250 °C   |   | 10 mK                               |  |
|  | > 250 °C to 550 °C   |   | 20 mK                               |  |
|  | > 550 °C to 660 °C   | DKD-R 5-1:2018 in furnaces mit with metal insert (sodium heat pipe) | 50 mK                               | Comparison with reference thermometers |
|  | > 660 °C to 960 °C   | 1 K   |                                     |  |
|  | -100 °C to 150 °C  | DKD-R 5-1:2018 in dry block calibrators                             | 15 mK                               |  |
|  | > 150 °C to 200 °C   |   | 20 mK                               |  |
|  | > 200 °C to 300 °C   |   | 30 mK                               |  |
|  | > 300 °C to 400 °C   |   | 40 mK                               |  |
|  | > 400 °C to 500 °C   |   | 50 mK                               |  |
|  | > 500 °C to 600 °C   |   | 85 mK                               |  |
| > 600 °C to 700 °C   | 0.10 K   | Comparison with reference thermometers                              |                                     |  |
| -80 °C to 100 °C   | DKD-R 5-1:2018 in climatic chambers or humidity generators |   | 0.1 K                               |  |
| > 100 °C to 180 °C   |  | 0.15 K  |                                     |  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Page 5 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item  | Range  | Measurement conditions / procedure                         | Expanded uncertainty of measurement                            | Remarks                                 |
|--|--|--|--|---|
| Noble metal thermocouples, direct reading thermometers with noble metal thermocouple sensor *) | -50 °C to 420 °C   | DKD-R 5-3:2018 in liquid baths                             | 0.4 K  | Comparison with reference thermometers  |
|  | 400 °C to 660 °C   | DKD-R 5-3:2018 in furnaces mit with metal insert           | 0.5 K  |   |
|  | > 660 °C to 1000 °C  |  | 0.8 K  |   |
|  | > 1000 °C to 1085 °C   |  | 1.2 K  |   |
|  | > 1085 °C to 1200 °C   |  | 1.5 K  |   |
|  | > 1200 °C to 1324 °C   |  | 2.0 K  |   |
|  | > 1324 °C to 1400 °C   |  | 3.0 K  |   |
| Base metal thermocouples, direct reading thermometers with base metal thermocouple sensor *)   | -196 °C  | DKD-R 5-3:2018 in liquid nitrogen                          | 0.3 K  | Comparison with reference thermometers  |
|  | -100 °C to 200 °C  | DKD-R 5-3:2018 in liquid baths                             | 0.2 K  |   |
|  | > 200 °C to 300 °C   |  | 0.2 K  |   |
|  | > 300 °C to 400 °C   |  | 0.3 K  |   |
|  | > 400 °C to 550 °C   |  | 0.6 K  |   |
|  | > 400 °C to 660 °C   |  | 0.6 K  |   |
|  | > 660 °C to 1100 °C  | DKD-R 5-3:2018 in furnaces mit with metal insert           | 1.5 K  |   |
|  | > 1100 °C to 1200 °C   |  | 3.0 K  |   |
|  | > 1200 °C to 1300 °C   |  | 4.0 K  |   |
|  | Comparison with reference thermometers   |  | -100 °C to 300 °C  | DKD-R 5-3:2018 in dry block calibrators |
|  |  | > 300 °C to 400 °C   | 0.4 K  |   |
|  |  | > 400 °C to 660 °C   | 0.7 K  |   |
|  |  | > 660 °C to 1100 °C  | 1.7 K  |   |
|  |  | > 1100 °C to 1200 °C                                       | 3.2 K  |   |
|  | Extension cables, compensation cables and reference junctions for thermocouples *) | -10 °C to 40 °C  | DKD-R 5-3:2018 in liquid baths and at fixed-point temperatures | 50 mK                                   |
| direct reading thermometers with thermocouple sensor *)  | -80 °C to 180 °C   | DKD-R 5-3:2018 in climatic chambers or humidity generators | 0.4 K  | Comparison with reference thermometers  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Annex to the Partial Accreditation Certificate D-K-15219-01-02

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item                              | Range              | Measurement conditions / procedure                                | Expanded uncertainty of measurement | Remarks   |
|--|--------------------|---|-------------------------------------|---|
| Liquid-in glass thermometers *)                                      | -80 °C to -58 °C   | PTB testing instruction<br>Volume 2:2003<br>in liquid bath        | 0.20 K                              | Comparison with reference thermometers  |
|  | > -58 °C to 110 °C |   | 10 mK                               |   |
|  | > 110 °C to 205 °C |   | 20 mK                               |   |
|  | > 205 °C to 420 °C |   | 40 mK                               |   |
|  | > 420 °C to 550 °C |   | 0.20 K                              |   |
| Temperature block calibrators *)                                     | -100 °C to 155 °C  | DKD-R 5-4:2018  | 0.03 K                              | Comparison with reference thermometers  |
|  | > 155 °C to 200 °C |   | 0.05 K                              |   |
|  | > 200 °C to 300 °C |   | 0.06 K                              |   |
|  | > 300 °C to 400 °C |   | 0.08 K                              |   |
|  | > 400 °C to 500 °C |   | 0.10 K                              |   |
|  | > 500 °C to 600 °C |   | 0.12 K                              |   |
|  | > 600 °C to 700 °C |   | 0.15 K                              |   |
|  | > 700 °C to 960 °C |   | 4 K                                 |   |
| > 960 °C to 1200 °C  | 5 K                |   |                                     |   |
| Temperature transmitters with resistance sensor *)                   | -200 °C to 660 °C  | see resistance thermometers                                       | $U_{PRT} + 0.1 \text{ K}$           | $U_{PRT}$ resp. $U_{TC}$ is the expanded uncertainty of measurement of the calibration of the resistance sensor resp. thermocouple sensor |
| Temperature transmitters with thermocouple sensor *)                 | -200 °C to 1400 °C | see thermocouples   | $U_{TC} + 0.5 \text{ K}$            |   |
| Calibration baths  | -100 °C to 300 °C  | QMV 7.2/23  | 30 mK                               | Comparison with reference thermometers  |
| Temperature indicators and simulators for resistance thermometers *) | -200 °C to 850 °C  | DKD-R 5-5:2018  | 2 mK                                | Characteristics according to DIN EN IEC 60751:2023  |
| for base metal thermocouples *)                                      | -200 °C to 1300 °C | DKD-R 5-5:2018<br>with or without reference junction compensation | 0.1 K                               | Characteristics according to DIN EN 60584:2014  |
| for thermocouples type S, R *)                                       | 0 °C to 1768 °C    |   | 0.2 K                               |   |
| for thermocouples type B *)  | 600 °C to 1820 °C  |   | 0.2 K                               |   |
| Humidity generators  | -10 °C to 95 °C    | QMV 7.2/22  | 0.05 K                              | Comparison with reference thermometers  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Page 7 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Partial Accreditation Certificate D-K-15219-01-02

**Permanent Laboratory**

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item  | Range            | Measurement conditions / procedure  | Expanded uncertainty of measurement | Remarks   |
|--|------------------|---|-------------------------------------|---|
| <b>Dew point temperature</b><br>Dew point mirrors,<br>dew point meters,<br>measuring transducers | -20 °C to 60 °C  | QMV 7.2/15  | 65 mK                               | Comparison with<br>reference dew point<br>mirror  |
|  | > 60 °C to 90 °C |   | 85 mK                               |   |
| Humidity generators,<br>dew point production<br>devices  | -20 °C to 60 °C  | QMV 7.2/22  | 50 mK                               | Comparison with<br>reference dew point<br>mirror  |
|  | > 60 °C to 90 °C |   | 70 mK                               |   |
| <b>Relative humidity</b><br>Hygrometers,<br>hygrographs,<br>measuring<br>transducers *)          | 10 % to 95 %     | DKD-R 5-8<br>in humidity generators<br>air temperature:<br>-10 °C to 70 °C  | $0.1\% + 0.0065 \cdot rH$           | Comparison with<br>reference thermo-<br>meter and reference<br>dew point mirror<br><i>rH</i> = measured value<br>Measurement<br>uncertainty<br>expressed as<br>absolute value of the<br>relative humidity |
|  | 5 % to 98 %      | DKD-R 5-8<br>in climatic chambers<br>air temperature:<br>5 °C to 95 °C      | $0.2\% + 0.008 \cdot rH$            |   |
| Psychrometers  | 10 % to 95 %     | QMV 7.2/15<br>in humidity generators<br>air temperature:<br>-10 °C to 70 °C | $0.1\% + 0.0065 \cdot rH$           |   |
|  | 5 % to 98 %      | QMV 7.2/15<br>in climatic chambers<br>air temperature:<br>5 °C to 95 °C     | $0.2\% + 0.008 \cdot rH$            |   |
| Humidity generators,<br>devices for generation<br>of relative humidity                           | 5 % to 98 %      | QMV 7.2/22<br>air temperature:<br>5 °C to 95 °C                             | $0.2\% + 0.006 \cdot rH$            |   |

Valid from: 12.01.2024

Date of issue: 12.01.2024



**Annex to the Partial Accreditation Certificate D-K-15219-01-02**
**Permanent Laboratory**
**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item                | Range                     | Measurement conditions / procedure  | Expanded uncertainty of measurement                   | Remarks  |   |
|--|---------------------------|---|---|--|---|
| <b>Pressure</b><br>Absolute pressure<br>$p_{abs}^{*)}$ | 0.015 bar to 1.8 bar      | DKD-R 6-1:2014<br><br>> 70 bar:<br>method of calibration<br>$p_{abs} = p_e + p_{amb}$ | $2.2 \cdot 10^{-5} \cdot p_{abs} + 3.0 \mu\text{bar}$ | Pressure medium : gas<br>The measurement uncertainty of the vacuum gauge ( $U_{res}$ ) has to be taken into account. |   |
|  | > 1.8 bar to 7 bar        |   | $2.3 \cdot 10^{-5} \cdot p_{abs} + 7.1 \mu\text{bar}$ |  |   |
|  | > 7 bar to 70 bar         |   | $3.1 \cdot 10^{-5} \cdot p_{abs} + 71 \mu\text{bar}$  |  |   |
|  | > 70 bar to 101 bar       |   |   | $5.4 \cdot 10^{-5} \cdot p_{abs} + 0.70 \text{ mbar}$  | Pressure medium : gas<br>With gas/oil separation device.<br>The measurement uncertainty of the barometer ( $U_{Baro}$ ) has to be taken into account. |
|  | > 101 bar to 201 bar      |   |   | $5.7 \cdot 10^{-5} \cdot p_{abs} + 1.4 \text{ mbar}$   |   |
|  | > 201 bar to 1001 bar     |   |   | $5.7 \cdot 10^{-5} \cdot p_{abs} + 3.5 \text{ mbar}$   |   |
| Absolute pressure<br>$p_{abs}^{*)}$                    | 1 bar; 2 bar to 101 bar   | DKD-R 6-1:2014<br><br>method of calibration<br>$p_{abs} = p_e + p_{amb}$              | $5.4 \cdot 10^{-5} \cdot p_{abs} + 0.70 \text{ mbar}$ | Pressure medium: oil<br>The measurement uncertainty of the barometer ( $U_{Baro}$ ) has to be taken into account.    |   |
|  | > 101 bar to 201 bar      |   | $5.7 \cdot 10^{-5} \cdot p_{abs} + 1.4 \text{ mbar}$  |  |   |
|  | > 201 bar to 1501 bar     |   | $5.7 \cdot 10^{-5} \cdot p_{abs} + 3.5 \text{ mbar}$  |  |   |
| Negative and positive gauge pressure $p_e^{*)}$        | -1.0 bar to -0.015 bar    | DKD-R 6-1:2014  | $4.9 \cdot 10^{-5} \cdot  p_e  + 5.3 \mu\text{bar}$   | Pressure medium : gas  |   |
|  | > -0.015 bar to 0.015 bar |   | 7.5 $\mu\text{bar}$                                   |  |   |
|  | > 0.015 bar to 1.8 bar    |   | $2.0 \cdot 10^{-5} \cdot p_e + 3.0 \mu\text{bar}$     |  |   |
|  | > 1.8 bar to 7.0 bar      |   | $2.0 \cdot 10^{-5} \cdot p_e + 7.1 \mu\text{bar}$     |  |   |
|  | > 7.0 bar to 70 bar       |   |   | $3.0 \cdot 10^{-5} \cdot p_e + 71 \mu\text{bar}$   | Pressure medium : gas<br>With gas/oil separation device.  |
|  | > 70 bar to 100 bar       |   |   | $5.4 \cdot 10^{-5} \cdot p_e + 0.70 \text{ mbar}$  |   |
|  | > 100 bar to 200 bar      |   |   | $5.7 \cdot 10^{-5} \cdot p_e + 1.4 \text{ mbar}$   |   |
|  | > 200 bar to 1000 bar     |   |   | $5.7 \cdot 10^{-5} \cdot p_e + 3.5 \text{ mbar}$   |   |
| Positive gauge pressure $p_e^{*)}$                     | 0 bar; 1 bar to 100 bar   | DKD-R 6-1:2014  | $5.4 \cdot 10^{-5} \cdot p_e + 0.70 \text{ mbar}$     | Pressure medium: oil   |   |
|  | > 100 bar to 200 bar      |   | $5.7 \cdot 10^{-5} \cdot p_e + 1.4 \text{ mbar}$      |  |   |
|  | > 200 bar to 1500 bar     |   | $5.7 \cdot 10^{-5} \cdot p_e + 3.5 \text{ mbar}$      |  |   |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Page 9 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

**On-site Calibration**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range                | Measurement conditions / procedure                     | Expanded uncertainty of measurement | Remarks                                |
|---|----------------------|--|-------------------------------------|--|
| <b>Temperatur</b><br>Resistance thermometers, direct reading thermometers with resistance sensor *) | -100 °C to 150 °C    | DKD-R 5-1:2018<br>in dry block calibrators             | 65 mK                               | Comparison with reference thermometers |
|   | > 150 °C to 200 °C   |  | 70 mK                               |  |
|   | > 200 °C to 300 °C   |  | 80 mK                               |  |
|   | > 300 °C to 400 °C   |  | 90 mK                               |  |
|   | > 400 °C to 500 °C   |  | 0.10 K                              |  |
|   | > 500 °C to 600 °C   |  | 0.14 K                              |  |
|   | > 600 °C to 700 °C   |  | 0.16 K                              |  |
| thermocouples, direct reading thermometers with thermocouple sensor *)                              | -100 °C to 300 °C    | DKD-R 5-3:2018<br>in dry block calibrators             | 0.4 K                               | Comparison with reference thermometers |
|   | > 300 °C to 400 °C   |  | 0.5 K                               |  |
|   | > 400 °C to 660 °C   |  | 0.8 K                               |  |
|   | > 660 °C to 1100 °C  |  | 1.7 K                               |  |
|   | > 1100 °C to 1200 °C |  | 3.2 K                               |  |
| Measuring locations in climatic chambers with air circulation *)                                    | -90 °C to -50 °C     | DKD-R 5-7:2018<br>method C<br>measurement in air       | 0.15 K                              | Comparison with reference thermometers |
|   | > -50 °C to < 0 °C   |  | 0.12 K                              |  |
|   | 0 °C to 100 °C       |  | 0.08 K                              |  |
|   | > 100 °C to 150 °C   |  | 0.13 K                              |  |
|   | > 150 °C to 200 °C   |  | 0.20 K                              |  |
|   | > 200 °C to 350 °C   |  | 0.33 K                              |  |
|   | > 350 °C to 500 °C   |  | 0.50 K                              |  |
| Climatic chambers with air circulation *)   | -90 °C to < 0 °C     | DKD-R 5-7:2018<br>method A and B<br>measurement in air | 0.4 K                               |  |
|   | 0 °C to 100 °C       |  | 0.2 K                               |  |
|   | > 100 °C to 150 °C   |  | 0.4 K                               |  |
|   | > 150 °C to 400 °C   |  | 0.6 K                               |  |
|   | > 400 °C to 500 °C   |  | 1.0 K                               |  |
| Measuring locations in climatic chambers without air circulation *)                                 | -90 °C to < 0 °C     | DKD-R 5-7:2018<br>method C<br>measurement in air       | 0.4 K                               |  |
|   | 0 °C to 100 °C       |  | 0.3 K                               |  |
|   | > 100 °C to 150 °C   |  | 0.4 K                               |  |
|   | > 150 °C to 200 °C   |  | 0.5 K                               |  |
|   | > 200 °C to 350 °C   |  | 0.8 K                               |  |
| Climatic chambers without air circulation *)  | -90 °C to 150 °C     | DKD-R 5-7:2018<br>method A and B<br>measurement in air | 0.6 K                               |  |
|   | > 150 °C to 350 °C   |  | 0.8 K                               |  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

**Annex to the Partial Accreditation Certificate D-K-15219-01-02**

**On-site Calibration**

**Calibration and Measurement Capabilities (CMC)**

| Measurement quantity / Calibration item   | Range               | Measurement conditions / procedure                                    | Expanded uncertainty of measurement | Remarks  |
|---|---------------------|---|-------------------------------------|--|
| Humidity generators   | 5 °C to 95 °C       | QMV 7.2/22  | 0.05 K                              | Comparison with reference thermometers   |
| Calibration baths   | -100 °C to 300 °C   | QMV 7.2/23  | 30 mK                               | Comparison with reference thermometers   |
| Temperature indicators and simulators<br>for resistance thermometers *)                     | -200 °C to 850 °C   | DKD-R 5-5:2018  | 0.1 K                               | Characteristics according to DIN EN 60751:2023   |
| for noble metal thermocouples *)  | 0 °C to 1820 °C     | DKD-R 5-5:2018<br>with or without reference junction compensation     | 0.6 K                               | Characteristics according to DIN EN 60584:2014   |
| for base metal thermocouples *)   | -200 °C to 500 °C   |   | 0.3 K                               |  |
|   | > 500 °C to 1370 °C |   | 0.5 K                               |  |
| <b>Relative Feuchte</b><br>Humidity generators, devices for generation of relative humidity | 5 % to 98 %         | QMV 7.2/22<br>air temperature: 5 °C to 95 °C                          | 0.2 % + 0.006 · rH                  | VComparison with reference thermometer and reference dew point mirror<br>rH = measured value<br>Measurement uncertainty expressed as absolute value of the relative humidity |
| Climatic chambers with air circulation *)   | 5 % to 30 %         | DKD-R 5-7:2018<br>method A and B<br>air temperature: 5 °C to 70 °C    | 0.8 %                               | Measurement uncertainty expressed as absolute value of the relative humidity   |
|   | > 30 % to 60 %      |   | 1.2 %                               |  |
|   | > 60 % to 95 %      |   | 1.6 %                               |  |
|   | 5 % to 95 %         | DKD-R 5-7:2018<br>method A and B<br>air temperature: > 70 °C to 95 °C | 2.1 %                               |  |
| Measuring locations in climatic chambers with air circulation *)                            | 5 % to 30 %         | DKD-R 5-7:2018<br>method C<br>air temperature: 5 °C to 70 °C          | 0.6 %                               |  |
|   | > 30 % to 60 %      |   | 0.8 %                               |  |
|   | > 60 % to 95 %      |   | 1.0 %                               |  |
|   | 5 % to 95 %         | DKD-R 5-7:2018<br>method C<br>air temperature: > 70 °C to 95 °C       | 1.8 %                               |  |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Annex to the Partial Accreditation Certificate D-K-15219-01-02

**On-site Calibration**

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item   | Range                     | Measurement conditions / procedure  | Expanded uncertainty of measurement                   | Remarks   |
|---|---------------------------|---|---|---|
| <b>Force (MTM)</b><br>Force measuring devices for Material Testing Machines according to DIN 51220:2003 *)              | 1 N to 500 N              | DIN EN ISO 7500-1:2018  | 0.10 %  | Mass stacks (compression and tensile)   |
|   | 50 N to 200 kN            | Sheet 1:1999<br>Sheet 2:1999<br>Sheet 3:1999  | 0.12 %  | Force transducer (compression and tensile)  |
| <b>Extension (MTM)</b><br>Lengt variation measuring device for Material Testing Machines according to DIN 51220:2003 *) | 20 mm to 1200 mm          | DIN EN ISO 9513:2013  | $2.0 \cdot 10^{-3} \cdot l$                           | Measuring principle: incremental<br>$l$ = measured extension  |
| <b>Pressure</b><br>Absolute pressure $p_{abs}$ *)   | 0.015 bar to 1.8 bar      | DKD-R 6-1:2014<br><br>> 70 bar:<br>method of calibration<br>$p_{abs} = p_e + p_{amb}$ | $2.3 \cdot 10^{-5} \cdot p_{abs} + 3.1 \mu\text{bar}$ | Pressure medium : gas<br>The measurement uncertainty of the vacuum gauge ( $U_{res}$ ) has to be taken into account.                                  |
|   | > 1.8 bar to 7 bar        |   | $2.4 \cdot 10^{-5} \cdot p_{abs} + 7.3 \mu\text{bar}$ |   |
|   | > 7 bar to 70 bar         |   | $3.2 \cdot 10^{-5} \cdot p_{abs} + 73 \mu\text{bar}$  |   |
|   | > 70 bar to 101 bar       |   | $6.0 \cdot 10^{-5} \cdot p_{abs} + 0.77 \text{ mbar}$ | Pressure medium : gas<br>With gas/oil separation device.<br>The measurement uncertainty of the barometer ( $U_{Baro}$ ) has to be taken into account. |
|   | > 101 bar to 201 bar      |   | $6.3 \cdot 10^{-5} \cdot p_{abs} + 1.5 \text{ mbar}$  |   |
|   | > 201 bar to 1001 bar     |   | $6.3 \cdot 10^{-5} \cdot p_{abs} + 3.9 \text{ mbar}$  |   |
| Absolute pressure $p_{abs}$ *)  | 1 bar; 2 bar to 101 bar   | DKD-R 6-1:2014<br><br>method of calibration<br>$p_{abs} = p_e + p_{amb}$              | $6.0 \cdot 10^{-5} \cdot p_{abs} + 0.77 \text{ mbar}$ | Pressure medium : gas<br>The measurement uncertainty of the barometer ( $U_{Baro}$ ) has to be taken into account.                                    |
|   | > 101 bar to 201 bar      |   | $6.3 \cdot 10^{-5} \cdot p_{abs} + 1.5 \text{ mbar}$  |   |
|   | > 201 bar to 1501 bar     |   | $6.3 \cdot 10^{-5} \cdot p_{abs} + 3.9 \text{ mbar}$  |   |
| Negative and positive gauge pressure $p_e$ *)   | -1.0 bar to -0.015 bar    | DKD-R 6-1:2014  | $5.1 \cdot 10^{-5} \cdot  p_e  + 5.4 \mu\text{bar}$   | Pressure medium : gas   |
|   | > -0.015 bar to 0.015 bar |   | 7.5 $\mu\text{bar}$                                   |   |
|   | > 0.015 bar to 1.8 bar    |   | $2.1 \cdot 10^{-5} \cdot p_e + 3.1 \mu\text{bar}$     |   |
|   | > 1.8 bar to 7.0 bar      |   | $2.1 \cdot 10^{-5} \cdot p_e + 7.3 \mu\text{bar}$     |   |
|   | > 7.0 bar to 70 bar       |   | $3.1 \cdot 10^{-5} \cdot p_e + 73 \mu\text{bar}$      | Pressure medium : gas<br>With gas/oil separation device.  |
|   | > 70 bar to 100 bar       |   | $6.0 \cdot 10^{-5} \cdot p_e + 0.77 \text{ mbar}$     |   |
|   | > 100 bar to 200 bar      |   | $6.3 \cdot 10^{-5} \cdot p_e + 1.5 \text{ mbar}$      |   |
|   | > 200 bar to 1000 bar     |   | $6.3 \cdot 10^{-5} \cdot p_e + 3.9 \text{ mbar}$      |   |
| Positive gauge pressure $p_e$ *)  | 0 bar; 1 bar to 100 bar   | DKD-R 6-1:2014  | $6.0 \cdot 10^{-5} \cdot p_e + 0.77 \text{ mbar}$     | Pressure medium: oil  |
|   | > 100 bar to 200 bar      |   | $6.3 \cdot 10^{-5} \cdot p_e + 1.5 \text{ mbar}$      |   |
|   | > 200 bar to 1500 bar     |   | $6.3 \cdot 10^{-5} \cdot p_e + 3.9 \text{ mbar}$      |   |

Valid from: 12.01.2024

Date of issue: 12.01.2024

Page 12 of 13

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Abbreviations used:**

|                    |  |
|--------------------|--|
| CMC                | Calibration and measurement capabilities   |
| DIN                | Deutsches Institut für Normung e.V. (German Institute for Standardization)                       |
| DKD-R              | Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt |
| EN                 | Europäische Norm (European Standard)   |
| EURAMET            | European Association of National Metrology Institutes  |
| QMV                | Calibration procedure of imetrologie GmbH Institut für Metrologie und Prozesstechnologie         |
| G-ITS-90, Part 2.2 | BIPM-Guide to the Realization of the ITS-90, Triple Point of Water                               |
| G-ITS-90, Part 2.3 | BIPM-Guide to the Realization of the ITS-90, Cryogenic Fixed Points                              |
| G-ITS-90, Part 2.4 | BIPM-Guide to the Realization of the ITS-90, Metal Fixed Points for Contact Thermometry          |
| G-ITS-90, Part 5   | BIPM-Guide to the Realization of the ITS-90, Platinum Resistance Thermometry                     |
| IEC                | International Electrotechnical Commission  |
| ISO                | International Organization for Standardization   |