

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-17127-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 01.09.2021

Date of issue: 01.09.2021

Holder of certificate:

Decom Prüflabor GmbH & Co. KG
Barbarastraße 2a, 24376 Kappeln

Calibration in the fields:

Dimensional quantities

Length

– **Thread**

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.

*The certificate together with the annex reflects the status as indicated by the date of issue.
The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de/en/accredited-bodies-search.html>.*

Abbreviations used: see last page

Page 1 of 5

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the accreditation certificate D-K-17127-01-00
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks	
Length					
Thread gauges (single and multi-start cylindrical external and internal threads with straight flanks, symmetrical and asymmetrical profile)					
External thread	Nominal diameter 1 mm to 300 mm	EURAMET cg-10 v. 2.1, category 1 to 3 VDI/VDE/DGQ 2618 Sheet 4.8:2006-04, Option 1 to Option 5 Three-wire-method (perpendicular to thread axis)		<i>l_F</i> = flank length in mm	
Pitch diameter			2.5 μm		
Major diameter			2 μm		
Minor diameter resp. recess diameter			5 μm		
Lead resp. pitch			0.25 mm to 12 mm		1 μm
Flank angle			≥ 3°		(1.2 + 1 mm / <i>l_F</i>); but not smaller than 3'
Internal thread	Nominal diameter 3 mm to 300 mm	EURAMET cg-10 v. 2.1, category 1 to 3 VDI/VDE/DGQ 2618 Sheet 4.9:2006-04, Option 1 to Option 5 Two-ball-method (perpendicular to thread axis)			
Pitch diameter			2.5 μm		
Major diameter resp. recess diameter			7 μm		
Minor diameter			3.5 μm		
Lead resp. pitch			0.5 mm to 12 mm		1 μm
Flank angle			≥ 3°		(1.2 + 1 mm / <i>l_F</i>); but not smaller than 5'

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-17127-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks	
Thread gauges (single and multi-start tapered external and internal threads with straight flanks, symmetrical and asymmetrical profile)					
External thread	Nominal diameter 1 mm to 300 mm	T.9:2019-08 Two-wire-method (inclined to the thread axis)			
Pitch diameter			2.5 µm		
Major diameter			2 µm		
Minor diameter resp. recess diameter			5 µm		
Lead resp. pitch			0.25 mm to 12 mm		1 µm
Flank angle			≥ 3°		(1.2 + 1 mm / l_F)'; but not smaller than 3'
Internal thread	Nominal diameter 3 mm to 300 mm	T.10:2019-08 Four-ball-method (perpendicular to thread generatrix)			
Pitch diameter			2.5 µm		
Major diameter resp. recess diameter			7 µm		
Minor diameter			3.5 µm		
Lead resp. pitch			0.5 mm to 12 mm		1 µm
Flank angle			≥ 3°		(1.2 + 1 mm / l_F)'; but not smaller than 5'

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-17127-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks	
External and internal threads single- and multi-start (1 to 10 threads), cylindrical and tapered, symmetrical and asymmetrical threads within the axial section straight flanks)					
External thread and internal threads	Nominal diameter	T.11:2021-01 Spatial Thread Measuring Method (single-ball- or multi-ball-method in respect to a setting mandrel or setting ring using the scanning method, self-centering with a three-coordinate measuring system and turntable).		The individual characteristics of the thread are recorded with a large number of measuring points and the quantities are computed from the measuring points using the "FUNKE" software.	
Major diameter for External thread	3 mm to 450 mm		2.0 μm		
Minor diameter resp. recess diameter for External thread			5.0 μm		
Minor diameter for internal thread			2.0 μm		
Major diameter resp. recess diameter for internal thread			5.0 μm		
Pitch diameter			2.5 μm		Evaluation with „FUNKE“
Lead Ph	0.35 mm to 100 mm		1.0 μm		
Total lead deviation $\Delta Ph\Sigma$	0.35 mm to 100 mm		1.0 μm		
Pitch P	0.70 mm to 50 mm		1.0 μm		
Flank angle β	-15° to 80°				0.87 mrad (3')
Flank angle γ	-15° to 80°			0.87 mrad (3')	
Virtual pitch diameter	Nominal diameter 3 mm to 450 mm			3.5 μm	The virtual pitch diameter simulates a pairing test, similar to the test with Go Thread gauges
Stand-off measurement				10 μm	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-17127-01-00

Abbreviations used:

CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
DGQ	Deutsche Gesellschaft für Qualität e.V.
EURAMET	European Association of National Metrology Institutes
T	calibration instruction of the Decom Prüflabor GmbH & Co. KG
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
VDI	Verein Deutscher Ingenieure e.V.

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.