

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-17601-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 29.06.2023

Date of issue: 29.06.2023

Holder of accreditation certificate:

H+H High Voltage Technology GmbH
Im Kurzen Busch 15, 58640 Iserlohn

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Abbreviations used: see last page

Page 1 of 4

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-17601-01-00

Calibration in the fields:

Electrical quantities

DC and low frequency quantities

- High voltage quantities *)
- High voltage impulse quantities *)
- Impuls charge

High frequency and radiation quantities

High frequency quantities

- Rise time
- Waveform quantities

*) also on site calibration

Within the scope of accreditation marked with **), the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range

Valid from: 29.06.2023

Date of issue: 29.06.2023

Annex to the Accreditation Certificate D-K-17601-01-00
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks
DC voltage	-1 kV to -300 kV 1 kV to +300 kV	DIN EN 60060-2:2011-10	0.7 %	
AC voltage	1 kV to 100 kV 50 kV to 300 kV	DIN EN 60060-2:2011-10 50 Hz 50 Hz	0.4 % 0.8 %	
Lightning impulse voltage (LI) Amplitude	-10 kV to -500 kV 10 kV to 500 kV	DIN EN 60060-2:2011-10	0.8 %	LI = Lightning impulse T ₁ = Front slope time T ₂ = Back slope- Half lifetime
Time parameter T ₁ T ₂	0.8 μs to 1.6 μs 40 μs to 60 μs		2.2 %	
Switching impulse voltage (SI) Amplitude	-10 kV to -500 kV +10 kV to +500 kV	DIN EN 60060-2:2011-10	1.1 %	SI = Switching impulse T _P = Scheitelzeit T ₂ = Back slope- Half lifetime
Time parameter T _P T ₂	200 μs to 300 μs 1000 to 4000 μs		2.0 %	
Impulse charge **)	1 pC to 100 nC	DIN EN 60270-2:2011-10	0.03 · q + 0.5 pC	q: charge quantity
Rise time t _r	3 ns to 1.0 μs		4,1 %	
Electrostatic discharge (ESD) **)		DIN EN 61000-4-2:2009-12		I _p = first burst current peak I ₃₀ = current at 30 ns I ₆₀ = current at 60 ns U _L = free state-voltage
Current pulse I _p	1 A to 35 A		3.1 %	
Basic values				
Current pulse I ₃₀	1 A to 35 A		1.9 %	
Current pulse I ₆₀	1 A to 35 A			
Rise time t _r	0.6 ns to 1 ns	3.9 %		
DC voltage U _L	1 kV to 100 kV	0.7 %		
EFT/B Burst **)		DIN EN 61000-4-4:2013-04		R _L = load resistor on R _L = 50 Ω on R _L = 1000 Ω
Voltage pulse	100 V to 2200 V 100 V to 4400 V		1.7 % 2.1 %	
Rise time and Impulse width	3 ns to 1 μs		4.1 %	
Burst duration and burst periode	100 ns to 1 s	2.0 %		
Surge **)		DIN EN 61000-4-5:2019-03		
DC amplitude	500 V to 12 kV		1.9 %	
Current amplitude	1 A to 120 kA		3.0 %	
Rise time and Impulse width	0.8 μs to 1.0 ms	1.0 %		
Dips **)		DIN EN 61000-4-11:2021-10 (16 2/3) Hz / 50 Hz / 60 Hz		
Voltage amplitude	1 V to 700 V		1.3 %	
Time interval	100 ns to 1 s	0.6 %		

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range

Annex to the Accreditation Certificate D-K-17601-01-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)							
Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement ¹	Remarks		
DC voltage	-1 kV 1 kV	to to	-300 kV +300 kV	DIN EN 60060-2:2011-10	0.7 %		
AC voltage	1 kV 50 kV	to to	100 kV 300 kV	DIN EN 60060-2:2011-10 50 Hz 50 Hz	0.4 % 0.8 %		
Lightning impulse voltage (LI) Amplitude	-10 kV 10 kV	to to	-500 kV 500 kV	DIN EN 60060-2:2011-10	0.8 %	LI = Lightning impulse T ₁ = Front slope time T ₂ = Back slope-half lifetime	
Time parameter T ₁	0.8 μs	to	1.6 μs				2.2 %
T ₂	40 μs	to	60 μs				
Switching impulse voltage (SI) Amplitude	-10 kV +10 kV	to to	-500 kV +500 kV	DIN EN 60060-2:2011-10	1.1 %	SI = Switching impulse T _p = Scheitelzeit T ₂ = Back slope-Half lifetime	
Time parameter T _p	200 μs	to	300 μs				2.0 %
T ₂	1000 μs	to	4000 μs				

Abbreviations used:

CMC Calibration and measurement capabilities
DIN Deutsches Institut für Normung e.V.
EN Europäische Norm

¹ Unless otherwise specified, the unit of a variable corresponds to the unit of the measuring range