

## Deutsche Akkreditierungsstelle

### Annex to the Accreditation Certificate D-K-17602-01-00 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 19.04.2024

**Date of issue:** 29.05.2024

Holder of accreditation certificate:

**Tektronix GmbH**  
**Heinrich-Pesch-Straße 11, 50739 Köln**

with the location

**Tektronix GmbH**  
**Heinrich-Pesch-Straße 11, 50739 Köln**

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

*This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.*

**Annex to the Accreditation Certificate D-K-17602-01-00**

Calibrations in the following areas:

**Electrical quantities**

**DC and low frequency quantities**

- DC voltage <sup>a)</sup>
- DC current <sup>a)</sup>
- DC resistance
- AC voltage

**Time and frequency**

- Frequency

**High frequency quantities**

- Oscilloscope quantities <sup>a)</sup>
- Rise time <sup>a)</sup>
- Band width

<sup>a)</sup> also on-site calibrations

Valid from: 19.04.2024

Date of issue: 29.05.2024

**Page 2 of 10**

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

Annex to the Accreditation Certificate D-K-17602-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
DC voltage measurement instruments	10 mV to 220 mV		$4 \cdot 10^{-6} \cdot U + 2 \mu\text{V}$	$U = \text{measured value}$
	> 220 mV to 2.2 V		$5 \cdot 10^{-6} \cdot U + 3 \mu\text{V}$	
sources	> 2.2 V to 22 V		$5 \cdot 10^{-6} \cdot U + 10 \mu\text{V}$	
	> 22 V to 220 V		$5 \cdot 10^{-6} \cdot U + 0.15 \text{ mV}$	
	> 220 V to 1100 V		$6 \cdot 10^{-6} \cdot U + 1 \text{ mV}$	
	10 mV to 100 mV		$5 \cdot 10^{-6} \cdot U + 3 \mu\text{V}$	
	> 100 mV to 1 V		$4 \cdot 10^{-6} \cdot U + 9 \mu\text{V}$	
	> 1 V to 10 V		$5 \cdot 10^{-6} \cdot U + 15 \mu\text{V}$	
	> 10 V to 100 V		$5 \cdot 10^{-6} \cdot U + 0.15 \text{ mV}$	
	> 100 V to 1100 V		$7 \cdot 10^{-6} \cdot U + 1.5 \text{ mV}$	
DC current measurement instruments	1 $\mu\text{A}$ to 220 $\mu\text{A}$		$45 \cdot 10^{-6} \cdot I + 7 \text{ nA}$	$I = \text{measured value}$
	> 220 $\mu\text{A}$ to 2.2 mA		$43 \cdot 10^{-6} \cdot I + 70 \text{ nA}$	
	> 2.2 mA to 22 mA		$50 \cdot 10^{-6} \cdot I + 0.40 \mu\text{A}$	
	> 22 mA to 220 mA		$65 \cdot 10^{-6} \cdot I + 0.50 \mu\text{A}$	
current clamps	> 220 mA to 2.2 A		$0.2 \cdot 10^{-3} \cdot I + 4 \mu\text{A}$	Coil with 5 windings
	> 2.2 A to 10 A		$0.21 \cdot 10^{-3} \cdot I + 0.14 \text{ mA}$	
	110 $\mu\text{A}$ to 11 mA		$0.15 \cdot 10^{-3} \cdot I + 30 \text{ nA}$	Coil with 50 windings
	> 11 mA to 110 mA		$0.15 \cdot 10^{-3} \cdot I + 200 \text{ nA}$	
	> 110 mA to 1.1 A		$0.15 \cdot 10^{-3} \cdot I + 2 \mu\text{A}$	
	> 1.1 A to 11 A		$0.25 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$	
	> 11 A to 55 A		$0.6 \cdot 10^{-3} \cdot I + 0.7 \text{ mA}$	
	10 mA to 110 mA		$1.5 \cdot 10^{-3} \cdot I + 0.3 \mu\text{A}$	
	> 110 mA to 1.1 A		$1.5 \cdot 10^{-3} \cdot I + 2 \mu\text{A}$	
	> 1.1 A to 11 A		$1.5 \cdot 10^{-3} \cdot I + 15 \mu\text{A}$	
	> 11 A to 110 A		$2.5 \cdot 10^{-3} \cdot I + 0.4 \text{ mA}$	
	> 110 A to 550 A		$6 \cdot 10^{-3} \cdot I + 7 \text{ mA}$	
sources	1 $\mu\text{A}$ to 100 $\mu\text{A}$		$50 \cdot 10^{-6} \cdot I + 9 \text{ nA}$	
	> 100 $\mu\text{A}$ to 1 mA		$50 \cdot 10^{-6} \cdot I + 80 \text{ nA}$	
	> 1 mA to 10 mA		$60 \cdot 10^{-6} \cdot I + 0.5 \mu\text{A}$	
	> 10 mA to 100 mA		$70 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	
	> 100 mA to 1 A		$0.24 \cdot 10^{-3} \cdot I + 10 \mu\text{A}$	
	> 1 A to 10 A		$0.25 \cdot 10^{-3} \cdot I + 0.30 \text{ mA}$	
DC resistance measurement instruments	1 $\Omega$		$90 \cdot 10^{-6} \cdot R$	$R = \text{measured value}$
	1.9 $\Omega$		$60 \cdot 10^{-6} \cdot R$	
	10 $\Omega$		$20 \cdot 10^{-6} \cdot R$	
	19 $\Omega$		$20 \cdot 10^{-6} \cdot R$	
	100 $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	190 $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	1 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	1.9 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	10 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	19 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	100 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	190 k $\Omega$		$15 \cdot 10^{-6} \cdot R$	
	1 M $\Omega$		$75 \cdot 10^{-6} \cdot R$	
	1.9 M $\Omega$		$75 \cdot 10^{-6} \cdot R$	
	10 M $\Omega$		$75 \cdot 10^{-6} \cdot R$	
	19 M $\Omega$		$75 \cdot 10^{-6} \cdot R$	
	100 M $\Omega$		$0.1 \cdot 10^{-3} \cdot R$	

Valid from: 19.04.2024

Date of issue: 29.05.2024

Annex to the Accreditation Certificate D-K-17602-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
DC resistance resistance	1 Ω to 10 Ω > 10 Ω to 100 Ω > 100 Ω to 1 MΩ > 1 MΩ to 10 MΩ > 10 MΩ to 100 MΩ		$20 \cdot 10^{-6} \cdot R$ $15 \cdot 10^{-6} \cdot R$ $25 \cdot 10^{-6} \cdot R$ $40 \cdot 10^{-6} \cdot R$ $0.1 \cdot 10^{-3} \cdot R$	R = measured value
frequency sources and measuring devices	1 MHz to 10 MHz 0.1 Hz to 20 GHz		$1 \cdot 10^{-11} \cdot f$ $1 \cdot 10^{-8} \cdot f + U_{Tf}$	f = measured value 1 MH step size $U_{Tf}$ trigger uncertainty
AC voltage measuring devices and probes	1 mV to 33 mV	10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 500 kHz	$0.80 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$ $0.15 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$ $0.2 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$ $1 \cdot 10^{-3} \cdot U + 7.5 \mu\text{V}$ $3.5 \cdot 10^{-3} \cdot U + 14 \mu\text{V}$ $8 \cdot 10^{-3} \cdot U + 54 \mu\text{V}$	Fluke 5522A Uncertainties excluding standard deviation and repeatability, which are determined at the time of measurement.
	> 33 mV to 330 mV	10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 500 kHz	$0.3 \cdot 10^{-3} \cdot U + 12 \mu\text{V}$ $0.15 \cdot 10^{-3} \cdot U + 11 \mu\text{V}$ $0.17 \cdot 10^{-3} \cdot U + 11 \mu\text{V}$ $0.35 \cdot 10^{-3} \cdot U + 13 \mu\text{V}$ $0.85 \cdot 10^{-3} \cdot U + 0.8 \text{ mV}$ $2 \cdot 10^{-3} \cdot U + 2 \text{ mV}$	In the case of probes, additional measurement uncertainties are caused by the display unit, unless this is part of a measuring chain.
	> 330 mV to 3.3 V	10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 500 kHz	$0.3 \cdot 10^{-3} \cdot U + 0.2 \text{ mV}$ $0.16 \cdot 10^{-3} \cdot U + 75 \mu\text{V}$ $0.2 \cdot 10^{-3} \cdot U + 75 \mu\text{V}$ $0.3 \cdot 10^{-3} \cdot U + 90 \mu\text{V}$ $0.7 \cdot 10^{-3} \cdot U + 180 \mu\text{V}$ $2.3 \cdot 10^{-3} \cdot U + 0.9 \text{ mV}$	
	> 3.3 V to 33 V	10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	$0.3 \cdot 10^{-3} \cdot U + 0.9 \text{ mV}$ $0.16 \cdot 10^{-3} \cdot U + 0.9 \text{ mV}$ $0.25 \cdot 10^{-3} \cdot U + 0.9 \text{ mV}$ $0.35 \cdot 10^{-3} \cdot U + 1.1 \text{ mV}$ $0.9 \cdot 10^{-3} \cdot U + 2.1 \text{ mV}$	
	> 33 V to 330 V	45 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	$0.2 \cdot 10^{-3} \cdot U + 6.5 \text{ mV}$ $0.2 \cdot 10^{-3} \cdot U + 10 \text{ mV}$ $0.25 \cdot 10^{-3} \cdot U + 10 \text{ mV}$ $0.26 \cdot 10^{-3} \cdot U + 24 \text{ mV}$ $1.9 \cdot 10^{-3} \cdot U + 0.11 \text{ V}$	
	> 330 V to 1020 V	45 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.3 \cdot 10^{-3} \cdot U + 20 \text{ mV}$ $0.25 \cdot 10^{-3} \cdot U + 20 \text{ mV}$ $0.18 \cdot 10^{-3} \cdot U + 0.22 \text{ V}$	

Valid from: 19.04.2024

Date of issue: 29.05.2024

Annex to the Accreditation Certificate D-K-17602-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
sources	1 mV to 10 mV	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$0.23 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $0.3 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $0.31 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $3 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $10 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$ $20 \cdot 10^{-3} \cdot U + 15 \mu\text{V}$	Fluke 8588A 1 year absolut specification $\pm 5^\circ\text{C}$ Tcal  Characteristic impedance of the device under test: 50 $\Omega$ , 1 M $\Omega$ , 10 M $\Omega$
	> 10 mV to 100 mV	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$85 \cdot 10^{-6} \cdot U + 3.5 \mu\text{V}$ $0.13 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $0.24 \cdot 10^{-3} \cdot U + 3 \mu\text{V}$ $0.48 \cdot 10^{-3} \cdot U + 12 \mu\text{V}$ $2 \cdot 10^{-3} \cdot U + 65 \mu\text{V}$ $10 \cdot 10^{-3} \cdot U + 220 \mu\text{V}$	
	> 100 mV to 1 V	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$85 \cdot 10^{-6} \cdot U + 18 \mu\text{V}$ $0.13 \cdot 10^{-3} \cdot U + 18 \mu\text{V}$ $0.24 \cdot 10^{-3} \cdot U + 18 \mu\text{V}$ $0.48 \cdot 10^{-3} \cdot U + 120 \mu\text{V}$ $2 \cdot 10^{-3} \cdot U + 630 \mu\text{V}$ $10 \cdot 10^{-3} \cdot U + 2 \text{ mV}$	
	> 1 V to 10 V	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$90 \cdot 10^{-6} \cdot U + 160 \mu\text{V}$ $0.13 \cdot 10^{-3} \cdot U + 120 \mu\text{V}$ $0.24 \cdot 10^{-3} \cdot U + 150 \mu\text{V}$ $0.49 \cdot 10^{-3} \cdot U + 1.1 \text{ mV}$ $2 \cdot 10^{-3} \cdot U + 6 \text{ mV}$ $10 \cdot 10^{-3} \cdot U + 14 \text{ mV}$	
	> 10 V to 100 V	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$0.1 \cdot 10^{-3} \cdot U + 1.5 \text{ mV}$ $0.12 \cdot 10^{-3} \cdot U + 1.5 \text{ mV}$ $0.19 \cdot 10^{-3} \cdot U + 8 \text{ mV}$ $0.51 \cdot 10^{-3} \cdot U + 16 \text{ mV}$ $3.7 \cdot 10^{-3} \cdot U + 62 \text{ mV}$ $11 \cdot 10^{-3} \cdot U + 0.53 \text{ V}$	
	100 V to 1000 V	1 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz	$0.12 \cdot 10^{-3} \cdot U + 32 \text{ mV}$ $0.12 \cdot 10^{-3} \cdot U + 32 \text{ mV}$ $0.2 \cdot 10^{-3} \cdot U + 80 \text{ mV}$ $0.55 \cdot 10^{-3} \cdot U + 0.16 \text{ V}$	
Rise time sources	7 ps to 35 ps* 10 ps to 40ps** > 40 ps to 25ns** 18 ps to 1 ms***	50 mV to 500 mV 50 mV to 50 V 50 mV to 50 V 10 mV to 250 V	$2 \cdot 10^{-2} \cdot t_R + 0.8 \text{ ps}$ 2.3 ps $5.5 \cdot 10^{-2} \cdot t_R$ $3 \cdot 10^{-2} \cdot t_R + 4 \text{ ps}$	* Derived size of waveform (sources) ** Correction factor method *** Gauss correction  ** & ***: To 4% Pulse roof distortion, worse above.
measuring devices and probes	18 ps to 1 ns* 16 ps to 1 ns** 500 ps to 3 ns* 1.5 ns to 25 ns*	10 mV to 250 mV 10 mV to 250 mV 0.25 V to 3 V 25 V und 50 V	$3 \cdot 10^{-2} \cdot t_R + 8 \text{ ps}$ $4 \cdot 10^{-2} \cdot t_R + 4 \text{ ps}$ $2 \cdot 10^{-2} \cdot t_R + 65 \text{ ps}$ $2 \cdot 10^{-2} \cdot t_R + 120 \text{ ps}$	* Gauss correction ** Correction factor method Periodic signals, pulse amplitudes
current clamps	1.5 ns to 20 ns 50 ns to 300 ns	0.5 A und 1A 5A	$3 \cdot 10^{-2} \cdot t_R + 200 \text{ ps}$ $1.5 \cdot 10^{-2} \cdot t_R + 250 \text{ ps}$	$t_R$ = rise time

Valid from: 19.04.2024

Date of issue: 29.05.2024

Annex to the Accreditation Certificate D-K-17602-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Waveform sources	5 GHz to 70 GHz*	10 mV to 1V	$5 \cdot 10^{-2} \cdot + \delta_r(t)$ [1] $1 \cdot 10^{-4} \cdot \Delta t + 0.5$ ps [2]	Sample to sample uncertainty for coherent, periodic signals. Derived step or impulse functions.  * spectral range of the FFT [1] Amplitude in the time domain [2] in the time domain  $\delta_r(t)$ = Influence of the source reflection of the test specimen on partial areas of the measuring points
Oscilloscopes	0.0 V	Range $\pm 200$ mV	25 $\mu$ V	$U$ = measured value 50 $\Omega$ Terminating resistor
DC voltage measuring devices	1 mV to $\pm 5$ V 1 mV to $\pm 200$ V	in 50 $\Omega$ in 1 M $\Omega$	$0.19 \cdot 10^{-3} \cdot U + 19$ $\mu$ V $0.20 \cdot 10^{-3} \cdot U + 19$ $\mu$ V	with Fluke-9500(B) + Fluke-9530
frequency response	4.4 mV <sup>(4)</sup> to 5.6 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 5.6 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 3.4 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 3.4 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 2.2 V <sup>(4)</sup>	10 MHz to 100 MHz > 100 MHz to 550 MHz > 550 MHz to 1.1 GHz > 1.1 GHz to 2.5 GHz > 2.5 GHz to 3.2 GHz	0.22 dB 0.27 dB 0.35 dB 0.39 dB 0.48 dB	with Fluke 9500/9530 Reference frequency $f_{ref}$ : 50 kHz to 10 MHz  $ \Gamma  \leq 0.23$ (VSWR $\leq 1.6$ ) Peak-to-peak voltage of the incident wave set on the calibrator. The measured variable is the ratio of the RMS values of the input voltage measured by the DUT at the measuring frequency and the reference frequency. Reflection factor of the connected test specimen with $ \Gamma  = \frac{VSWR-1}{VSWR+1}$  with VSWR = Standing wave ratio. Reference impedance: 50 $\Omega$

Valid from: 19.04.2024

Date of issue: 29.05.2024

Annex to the Accreditation Certificate D-K-17602-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
frequency response	< 0.8 V	1.0 GHz	0.22 dB	with Power Meter + Power Splitter  Coaxial connection technology PC-2.92 / Typ-K
	< 0.8 V	2.0 GHz	0.27 dB	
	< 0.8 V	2.5 GHz	0.34 dB	
	< 0.3 V	3.2 GHz to 8 GHz	0.30 dB	
	0.3 V <	3.2 GHz to 8 GHz	0.31 dB	
	< 0.3 V	8 GHz to 12.5 GHz	0.37 dB	
	0.3 V <	8 GHz to 12.5 GHz	0.38 dB	
	< 0.3 V	12.5 GHz to 16 GHz	0.46 dB	
	0.3 V <	12.5 GHz to 16 GHz	0.42 dB	
	< 0.3 V	16 GHz to 18 GHz	0.39 dB	
	0.3 V <	16 GHz to 18 GHz	0.42 dB	
	< 0.3 V	18 GHz to 20 GHz	0.39 dB	
	0.3 V <	18 GHz to 20 GHz	0.40 dB	
	< 0.12 V	20 GHz to 23 GHz	0.55 dB	
	0.12 V <	20 GHz to 23 GHz	0.58 dB	
< 0.12 V	23 GHz to 25 GHz	0.57 dB		
0.12 V <	23 GHz to 25 GHz	0.61 dB		
< 0.12 V	25 GHz to 33 GHz	0.87 dB		
0.12 V <	25 GHz to 33 GHz	0.84 dB		
AC voltage (sine) in 50 Ω	4.4 mV to 5.6 V	1 Hz to 550 MHz	$3.3 \cdot 10^{-2} \cdot U$	$U$ = measured value Fluke 9500(B)+9530 (1 Hz to 3.2 GHz)  Fluke 9500(B)+9560
	4.4 mV to 3.4 V	550 MHz to 2.5 GHz	$6.3 \cdot 10^{-2} \cdot U$	
	4.4 mV to 2.2 V	2.5 GHz to 3.2 GHz	$1.1 \cdot 10^{-1} \cdot U$	
	4.4 mV to 2.2 V	3.2 GHz to 6 GHz	$1.4 \cdot 10^{-1} \cdot U$	
time marker – measuring devices	80 ms		$21 \cdot 10^{-9} \cdot t$	Fluke 9500(B)+9530 with Fluke 910R
	400 ms		$3.4 \cdot 10^{-9} \cdot t$	
DC resistance	50 Ω		44 mΩ	with Fluke-9500(B)+9530
	75 Ω		0.12 Ω	
	250 kΩ		0.98 kΩ	
	1 MΩ		0.92 kΩ	
DC voltage sources	0 V to ±5 V		$0.14 \cdot 10^{-3} \cdot U + 90 \mu\text{V}$	with Keithley 2000 $U$ = measured value  Self-measurement of the test specimen, smallest uncertainty related to MSO series, HiRes mode
	0 V to ±0.2 V		33 μV	
Noise RMS	20 MHz B		0.18 μV	BW= Bandwidth limitation Self-measurement of the test specimen
	175 MHz W		2.2 μV	
	0.2 GHz B 10 GHz W to		1.1 μV	
frequency measuring devices	12 kHz to 3.2 GHz		$0.27 \cdot 10^{-6} \cdot f$	with Fluke 9500 opt.100 $f$ = frequency
time difference (Delta Time Acc)	to 1 GHz		49 fs	Anritsu MG3694C
	> 1 GHz to 2 GHz		43 fs	
	> 2 GHz to 16 GHz		30 fs	

Valid from: 19.04.2024

Date of issue: 29.05.2024

Page 7 of 10

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Accreditation Certificate D-K-17602-01-00**

**On-site calibration**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
DC voltage sources	1 mV to 100 mV > 100 mV to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		$60 \cdot 10^{-6} \cdot U + 5.5 \mu\text{V}$ $35 \cdot 10^{-6} \cdot U + 20 \mu\text{V}$ $35 \cdot 10^{-6} \cdot U + 30 \mu\text{V}$ $54 \cdot 10^{-6} \cdot U + 0.3 \text{ mV}$ $55 \cdot 10^{-6} \cdot U + 3 \text{ mV}$	$U =$ measured value
measuring devices	1 mV to 330 mV > 330 mV to 3.3 V > 3.3 V to 33 V > 33 V to 330 V > 330 V to 1000 V		$21 \cdot 10^{-6} \cdot U + 3.0 \mu\text{V}$ $11 \cdot 10^{-6} \cdot U + 16 \mu\text{V}$ $12 \cdot 10^{-6} \cdot U + 0.16 \text{ mV}$ $18 \cdot 10^{-6} \cdot U + 2.0 \text{ mV}$ $20 \cdot 10^{-6} \cdot U + 5.0 \text{ mV}$	$U =$ measured value
DC current measured value	1 $\mu\text{A}$ to 330 $\mu\text{A}$ > 330 $\mu\text{A}$ to 3.3 mA > 3.3 mA to 33 mA > 33 mA to 330 mA > 330 mA to 1.1 A > 1.1 A to 3.0 A > 3.0 A to 11 A		$0.17 \cdot 10^{-3} \cdot I + 30 \mu\text{A}$ $0.12 \cdot 10^{-3} \cdot I + 0.1 \mu\text{A}$ $0.12 \cdot 10^{-3} \cdot I + 0.5 \mu\text{A}$ $0.12 \cdot 10^{-3} \cdot I + 6.0 \mu\text{A}$ $0.24 \cdot 10^{-3} \cdot I + 70 \mu\text{A}$ $0.45 \cdot 10^{-3} \cdot I + 0.2 \text{ mA}$ $0.58 \cdot 10^{-3} \cdot I + 1.0 \text{ mA}$	$I =$ measured value
current clamps	10 $\mu\text{A}$ to 165 mA > 165 mA to 1.65 A > 1.65 A to 5.5 A > 5.5 A to 15 A > 15 A to 55 A		$0.4 \cdot 10^{-3} \cdot I + 1.5 \mu\text{A}$ $0.4 \cdot 10^{-3} \cdot I + 20 \mu\text{A}$ $0.8 \cdot 10^{-3} \cdot I + 0.25 \text{ mA}$ $1.5 \cdot 10^{-3} \cdot I + 0.35 \text{ mA}$ $2 \cdot 10^{-3} \cdot I + 3 \text{ mA}$	coil with 5 windings
	10 mA to 165 mA > 165 mA to 1.65 A > 1.65 A to 16.5 A > 16.5 A to 55 A > 55 A to 150 A > 150 A to 550 A		$4 \cdot 10^{-3} \cdot I + 3 \mu\text{A}$ $4 \cdot 10^{-3} \cdot I + 15 \mu\text{A}$ $4 \cdot 10^{-3} \cdot I + 0.2 \text{ mA}$ $8 \cdot 10^{-3} \cdot I + 2.5 \text{ mA}$ $15 \cdot 10^{-3} \cdot I + 5 \text{ mA}$ $20 \cdot 10^{-3} \cdot I + 30 \text{ mA}$	coil with 5 windings
Rise time sources	40 ps to 25 ns	50 mV to 50 V	$4 \cdot 10^{-2} \cdot t_R + 4 \text{ ps}$	Periodic signals, pulse amplitudes
measuring devices and probes	40 ps to 1 ns 500 ps to 3 ns 1.5 ns to 25 ns	10 mV to 250 mV 0.25 V to 3 V 25 V und 50 V	$4 \cdot 10^{-2} \cdot t_R + 8 \text{ ps}$ $2 \cdot 10^{-2} \cdot t_R + 65 \text{ ps}$ $2 \cdot 10^{-2} \cdot t_R + 120 \text{ ps}$	$t_R =$ rise time
current clamps	1.5 ns to 20 ns 50 ns to 300 ns	0.5 A und 1A 5A	$3 \cdot 10^{-2} \cdot t_R + 200 \text{ ps}$ $3 \cdot 10^{-2} \cdot t_R$	

Valid from: 19.04.2024

Date of issue: 29.05.2024

Page 8 of 10

This document is a translation. The definitive version is the original German annex to the accreditation certificate.



**Annex to the Accreditation Certificate D-K-17602-01-00**

**On-site calibration**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Oscilloscopes with Oscilloscope calibrator DC voltage	0.0 V to 0.2 V 1 mV to 5 V 1 mV to 200 V	in 1 M $\Omega$ oder 50 $\Omega$ in 50 $\Omega$ In 1 M $\Omega$	25 $\mu$ V $0.19 \cdot 10^{-3} \cdot U + 19 \mu$ V $0.20 \cdot 10^{-3} \cdot U + 19 \mu$ V	with Fluke 9500  $U =$ measured value
Time Marker – measuring devices	80 ms 400 ms		$21 \cdot 10^{-9} \cdot t$ $3.4 \cdot 10^{-9} \cdot t$	$t =$ Messwert Fluke 9500/9530 with Fluke 910R
AC voltage (sine) in 50 $\Omega$	4.4 mV to 5.6 V 4.4 mV to 3.4 V 4.4 mV to 2.2 V 4.4 mV to 2.2 V	1 Hz to 550 MHz 550 MHz to 2.5 GHz 2.5 GHz to 3.2 GHz 3.2 GHz to 6 GHz	$3.3 \cdot 10^{-2} \cdot U$ $6.3 \cdot 10^{-2} \cdot U$ $1.1 \cdot 10^{-1} \cdot U$ $1.4 \cdot 10^{-1} \cdot U$	$U =$ measured value Fluke 9500/9530 (1 Hz to 3.2 GHz)  Fluke 9500/9560
DC resistance	50 $\Omega$ 75 $\Omega$ 250 k $\Omega$ 1 M $\Omega$		44 m $\Omega$ 0.12 $\Omega$ 0.98 k $\Omega$ 0.92 k $\Omega$	with Fluke 9500/9530
DC voltage sources	0 V to $\pm 5$ V  0 V to $\pm 0.2$ V		$0.14 \cdot 10^{-3} \cdot U + 90 \mu$ V  33 $\mu$ V	with Keithley 2000 $U =$ measured value  Self-measurement of the test specimen, smallest uncertainty related to MSO series, HiRes mode

Valid from: 19.04.2024

Date of issue: 29.05.2024

Page 9 of 10

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**On-site calibration**

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
frequency response	4.4 mV <sup>(4)</sup> to 5.6 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 3.4 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 3.4 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 2.2 V <sup>(4)</sup> 4.4 mV <sup>(4)</sup> to 2.2 V <sup>(4)</sup>	10 MHz to 100 MHz > 100 MHz to 550 MHz > 550 MHz to 1.1 GHz > 1.1 GHz to 2.5 GHz > 2.5 GHz to 3.2 GHz	0.22 dB 0.27 dB 0.35 dB 0.39 dB 0.48 dB	with Fluke 9500/9530 reference frequency f <sub>ref</sub> : 50 kHz to 10 MHz  Γ  ≤ 0.23 (VSWR ≤ 1.6) Peak-to-peak voltage of the incident wave set on the calibrator. The measured variable is the ratio of the RMS values of the input voltage measured by the DUT at the measuring frequency and the reference frequency. Reflection factor of the connected test specimen with $ \Gamma  = \frac{VSWR-1}{VSWR+1}$ with VSWR = standing wave ratio. reference impedance: 50 Ω
noise RMS	20 MHz BW 175 MHz BW 0.2 GHz to 10 GHz		0.18 μV 2.2 μV 1.1 μV	BW= Bandwidth limitation Self-measurement of the test specimen
frequency sources	12 kHz 3.2 GHz		0.27 · 10 <sup>-6</sup> · f	with Fluke 9500 opt.100 f = frequency

**Abbreviations used:**

- DIN Deutsches Institut für Normung e.V.
- EN Europäische Norm
- IEC International Electrotechnical Commission
- ISO International Organization for Standardization