

## Deutsche Akkreditierungsstelle

### Annex to the Partial Accreditation Certificate D-K-17616-01-02 according to DIN EN ISO/IEC 17025:2018

**Valid from:** **07.11.2022**

Date of issue: **14.02.2023**

This annex is a part of the accreditation certificate D-K-17616-01-00.

Holder of partial accreditation certificate:

**Thermo Electron LED GmbH**  
**Robert-Bosch-Straße 1, 63505 Langenselbold**

The calibration laboratory meets the minimal requirements of DIN EN ISO/IEC 17025:2018 and, if applicable, additional legal and normative requirements, including those in relevant sectoral schemes, in order to carry out the conformity assessment activities listed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

Calibration in the fields:

**Electrical quantities**

**DC and low frequency quantities**

- **DC resistance**
- **DC current**
- **DC voltage**
- **AC current**
- **AC voltage**

*This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.*

Abbreviations used: see last page

**Page 1 of 5**

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
DC voltage Measuring instruments	0V		2 µV	$U = \text{set value}$
	0,001 V to < 0,33 V		$37 \cdot 10^{-6} \cdot U + 2 \mu\text{V}$	
	0,33 V to < 3,3 V		$15 \cdot 10^{-6} \cdot U + 3 \mu\text{V}$	
	3,3 V to < 33 V		$17 \cdot 10^{-6} \cdot U + 25 \mu\text{V}$	
	33 V to < 330 V		$20 \cdot 10^{-6} \cdot U + 0,2 \text{ mV}$	
DC voltage Sources	330 V to 1000 V		$22 \cdot 10^{-6} \cdot U + 1,7 \text{ mV}$	$U = \text{measuring value}$
	0 V		2 µV	
	0,001 V to < 0,2 V		$5 \cdot 10^{-6} \cdot U + 2 \mu\text{V}$	
	0,2 V to < 2 V		$5 \cdot 10^{-6} \cdot U + 1 \mu\text{V}$	
	2 V to < 20 V		$5 \cdot 10^{-6} \cdot U + 5 \mu\text{V}$	
DC current Measuring instruments	20 V to < 200 V		$7 \cdot 10^{-6} \cdot U + 50 \mu\text{V}$	$I = \text{set value}$
	200 V to 1000 V		$7 \cdot 10^{-6} \cdot U + 0,6 \text{ mV}$	
	0 A		2 µA	
	100 µA to < 330 µA		$10 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	
	330 µA to < 3,3 mA		$30 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	
	3,3 mA to < 33 mA		$0,10 \cdot 10^{-3} \cdot I + 2 \mu\text{A}$	
	33 mA to < 330 mA		$0,12 \cdot 10^{-3} \cdot I + 5 \mu\text{A}$	
	330 mA to < 1,1 A		$0,25 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$	
DC current Sources	1,1 A to < 3 A		$0,45 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$	$I = \text{measuring value}$
	3 A to < 11 A		$0,6 \cdot 10^{-3} \cdot I + 0,6 \text{ mA}$	
	11 A to 20,5 A		$1,2 \cdot 10^{-3} \cdot I + 1,7 \text{ mA}$	
	0 A		2 µA	
	0,1mA to < 2 mA		$1 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	
DC resistance Measuring instruments	2 mA to < 20 mA		$4 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	$R = \text{set value}$
	20 mA to < 200 mA		$45 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$	
	0,2A to < 2 A		$0,21 \cdot 10^{-3} \cdot I + 20 \mu\text{A}$	
	2A to 20 A		$0,47 \cdot 10^{-3} \cdot I + 0,47 \text{ mA}$	
	0 Ω		0,5 mΩ	
DC resistance Measuring instruments	0,01 Ω to < 11 Ω		$35 \cdot 10^{-6} \cdot R + 1,6 \text{ m}\Omega$	$R = \text{set value}$
	11 Ω to < 110 Ω		$30 \cdot 10^{-6} \cdot R + 2,3 \text{ m}\Omega$	
	110 Ω to < 1,1 kΩ		$30 \cdot 10^{-6} \cdot R + 2,7 \text{ m}\Omega$	
	1,1 kΩ to < 11 kΩ		$35 \cdot 10^{-6} \cdot R + 25 \text{ m}\Omega$	
	11 kΩ to < 110 kΩ		$35 \cdot 10^{-6} \cdot R + 0,25 \Omega$	
	110 kΩ to < 1,1 MΩ		$40 \cdot 10^{-6} \cdot R + 2,5 \Omega$	
	1,1 MΩ to < 3,3 MΩ		$70 \cdot 10^{-6} \cdot R + 35 \Omega$	
	3,3 MΩ to < 11 MΩ		$0,16 \cdot 10^{-3} \cdot R + 60 \Omega$	
	11 MΩ to < 33 MΩ		$0,3 \cdot 10^{-3} \cdot R + 3 \text{ k}\Omega$	
	33 MΩ to < 110 MΩ		$0,6 \cdot 10^{-3} \cdot R + 3,5 \text{ k}\Omega$	
	110 MΩ to < 330 MΩ		$3,5 \cdot 10^{-3} \cdot R + 0,15 \text{ M}\Omega$	
	330 MΩ to 1,1 GΩ		$17 \cdot 10^{-3} \cdot R + 0,6 \text{ M}\Omega$	

Valid from: 07.11.2022

Date of issue: 14.02.2023

**Page 2 of 5**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks	
DC resistance Sources	0 Ω			0,5 mΩ	$R = \text{measuring value}$	
	0,1 mΩ	to		$20 \cdot 10^{-6} \cdot R + 0,015 \text{ mΩ}$		
	2 Ω	to		$15 \cdot 10^{-6} \cdot R + 0,02 \text{ mΩ}$		
	20 Ω	to		$10 \cdot 10^{-6} \cdot R + 0,06 \text{ mΩ}$		
	200 Ω	to		$10 \cdot 10^{-6} \cdot R + 0,6 \text{ mΩ}$		
	2 kΩ	to		$10 \cdot 10^{-6} \cdot R + 6 \text{ mΩ}$		
	20 kΩ	to		$10 \cdot 10^{-6} \cdot R + 60 \text{ mΩ}$		
	200 kΩ	to		$12 \cdot 10^{-6} \cdot R + 1,2 \text{ Ω}$		
	2 MΩ	to		$25 \cdot 10^{-6} \cdot R + 0,12 \text{ kΩ}$		
	20 MΩ	to		$0,15 \cdot 10^{-3} \cdot R + 12 \text{ kΩ}$		
AC voltage Measuring instruments	0,001 V		10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 500 kHz	0,75 · $10^{-3} \cdot U + 10 \text{ μV}$	$U = \text{set value}$	
	0,001 V to < 0,033 V			0,13 · $10^{-3} \cdot U + 10 \text{ μV}$		
	0,033 V			0,18 · $10^{-3} \cdot U + 10 \text{ μV}$		
	< 0,33 V			1,0 · $10^{-3} \cdot U + 9 \text{ μV}$		
	0,33 V			3,5 · $10^{-3} \cdot U + 18 \text{ μV}$		
	< 3,3 V			7,5 · $10^{-3} \cdot U + 75 \text{ μV}$		
	3,3 V		10 Hz to 45 Hz > 45 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 500 kHz	0,40 · $10^{-3} \cdot U + 10 \text{ μV}$		
	< 33 V			0,15 · $10^{-3} \cdot U + 12 \text{ μV}$		
	33 V			0,20 · $10^{-3} \cdot U + 12 \text{ μV}$		
	< 330 V			0,40 · $10^{-3} \cdot U + 12 \text{ μV}$		
	330 V			0,80 · $10^{-3} \cdot U + 45 \text{ μV}$		
	1000 V			2,0 · $10^{-3} \cdot U + 100 \text{ μV}$		

Valid from: 07.11.2022

Date of issue: 14.02.2023

**Page 3 of 5**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage Sources	0,001 V to < 0,2 V	10 Hz to 40 Hz > 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz	$0,18 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$ $0,13 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$ $0,13 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$ $0,15 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$ $0,37 \cdot 10^{-3} \cdot U + 12 \mu\text{V}$ $0,85 \cdot 10^{-3} \cdot U + 30 \mu\text{V}$	$U = \text{measuring value}$
	0,2 V to < 2 V	10 Hz to 40 Hz > 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$0,14 \cdot 10^{-3} \cdot U + 25 \mu\text{V}$ $0,11 \cdot 10^{-3} \cdot U + 25 \mu\text{V}$ $90 \cdot 10^{-6} \cdot U + 25 \mu\text{V}$ $0,13 \cdot 10^{-3} \cdot U + 25 \mu\text{V}$ $0,26 \cdot 10^{-3} \cdot U + 50 \mu\text{V}$ $0,66 \cdot 10^{-3} \cdot U + 0,25 \text{ mV}$ $3,5 \cdot 10^{-3} \cdot U + 2,9 \text{ mV}$ $12 \cdot 10^{-3} \cdot U + 30 \text{ mV}$	
	2 V to < 20 V	10 Hz to 40 Hz > 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 1 MHz	$0,14 \cdot 10^{-3} \cdot U + 0,25 \text{ mV}$ $0,11 \cdot 10^{-3} \cdot U + 0,25 \text{ mV}$ $90 \cdot 10^{-6} \cdot U + 0,25 \text{ mV}$ $0,13 \cdot 10^{-3} \cdot U + 0,25 \text{ mV}$ $0,26 \cdot 10^{-3} \cdot U + 0,50 \text{ mV}$ $0,66 \cdot 10^{-3} \cdot U + 2,5 \text{ mV}$ $3,5 \cdot 10^{-3} \cdot U + 29 \text{ mV}$ $12 \cdot 10^{-3} \cdot U + 0,30 \text{ V}$	
	20 V to < 200 V	10 Hz to 40 Hz > 40 Hz to 100 Hz > 100 Hz to 2 kHz > 2 kHz to 10 kHz > 10 kHz to 30 kHz > 30 kHz to 100 kHz	$0,14 \cdot 10^{-3} \cdot U + 2,3 \text{ mV}$ $0,11 \cdot 10^{-3} \cdot U + 2,3 \text{ mV}$ $95 \cdot 10^{-6} \cdot U + 2,3 \text{ mV}$ $0,13 \cdot 10^{-3} \cdot U + 2,3 \text{ mV}$ $0,26 \cdot 10^{-3} \cdot U + 5,0 \text{ mV}$ $0,67 \cdot 10^{-3} \cdot U + 25 \text{ mV}$	
	200 V to 1000 V	> 40 Hz to 10 kHz > 10 kHz to 30 kHz	$0,15 \cdot 10^{-3} \cdot U + 25 \text{ mV}$ $0,25 \cdot 10^{-3} \cdot U + 50 \text{ mV}$	

Valid from: 07.11.2022

Date of issue: 14.02.2023

**Page 4 of 5**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

**Annex to the Partial Accreditation Certificate D-K-17616-01-02**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC current Measuring instruments	0,029 mA to < 0,33 mA	10 Hz to 20 Hz	$1 \cdot 10^{-3} \cdot I + 1,5 \mu\text{A}$	$I = \text{measuring value}$
		> 20 Hz to 45 Hz	$0,6 \cdot 10^{-3} \cdot I + 1,5 \mu\text{A}$	
		> 45 Hz to 1 kHz	$0,45 \cdot 10^{-3} \cdot I + 1,5 \mu\text{A}$	
		> 1 kHz to 5 kHz	$1,9 \cdot 10^{-3} \cdot I + 1,5 \mu\text{A}$	
		> 5 kHz to 10 kHz	$7,5 \cdot 10^{-3} \cdot I + 1,5 \mu\text{A}$	
	0,33 mA to < 3,3 mA	10 Hz to 20 Hz	$2,2 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	
		> 20 Hz to 45 Hz	$1,3 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	
		> 45 Hz to 1 kHz	$1 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	
		> 1 kHz to 5 kHz	$2,2 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	
AC current Sources	3,3 mA to < 33 mA	> 5 kHz to 10 kHz	$6 \cdot 10^{-3} \cdot I + 1 \mu\text{A}$	$I = \text{measuring value}$
		10 Hz to 20 Hz	$2,1 \cdot 10^{-3} \cdot I + 2,5 \mu\text{A}$	
		> 20 Hz to 45 Hz	$1,1 \cdot 10^{-3} \cdot I + 2,5 \mu\text{A}$	
		> 45 Hz to 1 kHz	$0,5 \cdot 10^{-3} \cdot I + 2,5 \mu\text{A}$	
		> 1 kHz to 5 kHz	$1,0 \cdot 10^{-3} \cdot I + 2,4 \mu\text{A}$	
		> 5 kHz to 10 kHz	$2,5 \cdot 10^{-3} \cdot I + 3,5 \mu\text{A}$	
	33 mA to < 330 mA	10 Hz to 20 Hz	$2,1 \cdot 10^{-3} \cdot I + 25 \mu\text{A}$	
		> 20 Hz to 45 Hz	$1,1 \cdot 10^{-3} \cdot I + 25 \mu\text{A}$	
		> 45 Hz to 1 kHz	$0,5 \cdot 10^{-3} \cdot I + 25 \mu\text{A}$	
		> 1 kHz to 5 kHz	$1,2 \cdot 10^{-3} \cdot I + 60 \mu\text{A}$	
AC current Sources	0,33 A to < 1,1 A	> 5 kHz to 10 kHz	$2,4 \cdot 10^{-3} \cdot I + 0,12 \text{ mA}$	$I = \text{measuring value}$
		10 Hz to 45 Hz	$2,1 \cdot 10^{-3} \cdot I + 0,12 \text{ mA}$	
		> 45 Hz to 1 kHz	$0,6 \cdot 10^{-3} \cdot I + 0,12 \text{ mA}$	
		> 1 kHz to 5 kHz	$7 \cdot 10^{-3} \cdot I + 1,2 \text{ mA}$	
		> 5 kHz to 10 kHz	$30 \cdot 10^{-3} \cdot I + 6 \text{ mA}$	
	1,1 A to < 3 A	10 Hz to 45 Hz	$2,1 \cdot 10^{-3} \cdot I + 0,12 \text{ mA}$	
		> 45 Hz to 1 kHz	$0,7 \cdot 10^{-3} \cdot I + 0,13 \text{ mA}$	
		> 1 kHz to 5 kHz	$7 \cdot 10^{-3} \cdot I + 1,2 \text{ mA}$	
AC current Sources	3 A to < 11 A	> 5 kHz to 10 kHz	$29 \cdot 10^{-3} \cdot I + 5,9 \text{ mA}$	$I = \text{measuring value}$
		45 Hz to 100 Hz	$0,72 \cdot 10^{-3} \cdot I + 2,5 \text{ mA}$	
		> 100 Hz to 1 kHz	$1,2 \cdot 10^{-3} \cdot I + 2,5 \text{ mA}$	
	11 A to 20,5 A	> 1 kHz to 5 kHz	$35 \cdot 10^{-3} \cdot I + 2,5 \text{ mA}$	
		45 Hz to 100 Hz	$1,5 \cdot 10^{-3} \cdot I + 6 \text{ mA}$	
		> 100 Hz to 1 kHz	$1,8 \cdot 10^{-3} \cdot I + 6 \text{ mA}$	
		> 1 kHz to 5 kHz	$35 \cdot 10^{-3} \cdot I + 6 \text{ mA}$	
	2 A to 20 A	10 Hz to 2 kHz	$3 \cdot 10^{-3} \cdot I + 2,5 \text{ mA}$	

**Abbreviations used:**

CMC              Calibration and measurement capabilities

Valid from:      07.11.2022

Date of issue:    14.02.2023

**Page 5 of 5**

This document is a translation. The definitive version is the original German annex to the accreditation certificate.