

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-18255-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 27.03.2024 Date of issue: 27.03.2024

Holder of accreditation certificate:

DEPRAG SCHULZ GMBH u. CO. KG Carl-Schulz-Platz 1, 92224 Amberg

with the location

DEPRAG SCHULZ GMBH u. CO. KG Carl-Schulz-Platz 1, 92224 Amberg

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the general with the principles of DIN EN ISO 9001.

Calibration in the fields:

Mechanical quantities - Torque

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at https://www.dakks.de.



Annex to the Accreditation Certificate D-K-18255-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range			Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Torque Torque measurement devices	0.01 N·m	to	< 0.1 N·m	DIN 51309: 2022 case l	4 · 10 ⁻³	
	0.1 N·m	to	500 N∙m	DIN 51309:-2022	1 · 10 ⁻³	

Abbreviations used:

- CMC Calibration and measurement capabilities
- DIN Deutsches Institut für Normung e.V. German institute for standardization
- EN Europäische Norm European Standard
- IEC International Electrotechnical Commission
- ISO International Organization for Standardisation