



Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-21444-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 02.05.2024

Date of issue: 02.05.2024

Holder of accreditation certificate:

EP EHRLER PRÜFTECHNIK ENGINEERING GMBH
Wilhelm-Hachtel-Straße 8, 97996 Niederstetten

with the location

EP EHRLER PRÜFTECHNIK ENGINEERING GMBH
Wilhelm-Hachtel-Straße 8, 97996 Niederstetten

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Abbreviations used: see last page

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Page 1 of 6

Annex to the Accreditation Certificate D-K-21444-01-00

Calibration in the fields:

Mechanical Quantities

- Pressure ^{a)}

Fluid Quantities

- Gas flow rate ^{a)}
- Volume of flowing gases ^{a)}
- Mass of flowing gases ^{a)}

Thermodynamic Quantities

Temperature Quantities

- Direct reading thermometers ^{a)}

- Temperature transmitters, data loggers ^{a)}

Humidity Quantities

- Devices for relative humidity ^{a)}

^{a)} also on-site calibrations

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement | Remarks |
|---|--|--|---|--|
| Pressure Absolute pressure p_{abs} | 0.15 bar to 2 bar | DKD-R 6-1:2014 | 16 Pa + 1 · 10 ⁻⁴ · p_{abs} | Pressure medium: air or nitrogen |
| | > 2 bar to 10 bar | | 30 Pa + 2 · 10 ⁻⁴ · p_{abs} | |
| Differential pressure Δp | 0 hPa to 60 hPa | DKD-R 6-1:2014 | 0.3 Pa + 2 · 10 ⁻⁴ · Δp | Pressure medium: air or nitrogen at atmospheric pressure |
| | > 60 hPa to 600 hPa | | 3 Pa + 2 · 10 ⁻⁴ · Δp | |
| Negative and positive gauge pressure p_e | -0.85 bar to -0.6 bar | DKD-R 6-1:2014 | 27 Pa | Pressure medium: air or nitrogen |
| | > -0.6 bar to -0.06 bar | | 3 Pa + 2 · 10 ⁻⁴ · p_e | |
| | > -0.06 bar to 0.06 bar | | 0.3 Pa + 2 · 10 ⁻⁴ · p_e | |
| | > 0.06 bar to 0.6 bar | | 3 Pa + 2 · 10 ⁻⁴ · p_e | |
| | > 0.6 bar to 10 bar | | 40 Pa + 2 · 10 ⁻⁴ · p_e | |
| | > 10 bar to 30 bar | | 4 mbar + 2 · 10 ⁻⁴ · p_e | |
| | > 30 bar to 600 bar | | 40 mbar + 2 · 10 ⁻⁴ · p_e | |
| Fluid Quantities Volume flow rate Q_v or volume of flowing gases | 5.7 L/h to 80 L/h | 4-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | 4.2 · 10 ⁻³ · Q_v + 0.05 L/h | Calibration medium: atmospheric air at room temperature and atmospheric pressure or compressed air at room temperature up to max. 8 bar overpressure and max. 800 kg/h standards at atmospheric pressure (0.85 bar < $p_{N,abs}$ < 1.05 bar) |
| | 5.7 L/h to 80 m ³ /h | 7-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | 3.7 · 10 ⁻³ · Q_v + 0.065 L/h | |
| | 2 L/h to 21 m ³ /h | 16-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | 4.15 · 10 ⁻³ · Q_v + 0.066 L/h | |
| | 1.5 m ³ /h to 1 600 m ³ /h | 10-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | 5.0 · 10 ⁻³ · Q_v + 7 L/h | |
| | 2.5 m ³ /h to 4 000 m ³ /h | 13-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | 3.8 · 10 ⁻³ · Q_v + 2.8 L/h | |
| | 0.6 m ³ /h to 100 m ³ /h | Low pulsation rotary gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | 3.8 · 10 ⁻³ · Q_v | |
| | > 100 m ³ /h to 1 600 m ³ /h | | 4.1 · 10 ⁻³ · Q_v | |

Valid from: 02.05.2024

Date of issue: 02.05.2024

Page 2 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-21444-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement | Remarks |
|---|---|---|--|--|
| Volume flow rate Q_v or volume of flowing gases | 130 m ³ /h to 3 750 m ³ /h | Turbine gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $3.85 \cdot 10^{-3} \cdot Q_v + 0.27 \text{ m}^3/\text{h}$ | Calibration medium: atmospheric air at room temperature and atmospheric pressure or compressed air at room temperature up to max. 8 bar overpressure and max. 800 kg/h standards at atmospheric pressure (0.85 bar < $p_{N,abs}$ < 1.05 bar) |
| | 100 mL/h to 3 L/h | Laminar flow elements EP.LAB.VA.01.01: 2024-01, chap. 2.4.3 | $5.0 \cdot 10^{-3} \cdot Q_v + 2.4 \text{ mL/h}$ | |
| | > 3 L/h to 17 L/h | | $5.5 \cdot 10^{-3} \cdot Q_v + 15 \text{ mL/h}$ | |
| | > 17 L/h to 0.16 m ³ /h | | $5.35 \cdot 10^{-3} \cdot Q_v + 0.16 \text{ L/h}$ | |
| | > 0.16 m ³ /h to 0.68 m ³ /h | | $5.9 \cdot 10^{-3} \cdot Q_v + 0.65 \text{ L/h}$ | |
| | > 0.68 m ³ /h to 3 m ³ /h | | $6.1 \cdot 10^{-3} \cdot Q_v + 3 \text{ L/h}$ | |
| | > 3 m ³ /h to 5.5 m ³ /h | | $7.0 \cdot 10^{-3} \cdot Q_v + 5.2 \text{ L/h}$ | |
| | > 5.5 m ³ /h to 13 m ³ /h | | $6.0 \cdot 10^{-3} \cdot Q_v + 13 \text{ L/h}$ | |
| | 8.7 m ³ /h to 275 m ³ /h | | $6.5 \cdot 10^{-3} \cdot Q_v + 0.55 \text{ m}^3/\text{h}$ | |
| | > 275 m ³ /h to 1 848 m ³ /h | Prandtl probe EP.LAB.VA.01.01: 2024-01, chap. 2.4.4 | $5.3 \cdot 10^{-3} \cdot Q_v + 1.9 \text{ m}^3/\text{h}$ | |
| | > 1 848 m ³ /h to 10 000 m ³ /h | | $5.0 \cdot 10^{-3} \cdot Q_v + 10.0 \text{ m}^3/\text{h}$ | |
| Mass flow rate Q_m or mass of flowing gases | 6.3 g/h to 90 g/h | 4-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.0 \cdot 10^{-3} \cdot Q_m + 0.065 \text{ g/h}$ | Calibration medium: atmospheric air at room temperature and atmospheric pressure or compressed air at room temperature up to max. 8 bar overpressure and max. 800 kg/h standards at atmospheric pressure (0.85 bar < $p_{N,abs}$ < 1.05 bar) |
| | 6.3 g/h to 90 kg/h | 7-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.0 \cdot 10^{-3} \cdot Q_m + 0.09 \text{ g/h}$ | |
| | 2 g/h to 24 kg/h | 16-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.5 \cdot 10^{-3} \cdot Q_m + 0.1 \text{ g/h}$ | |
| | 1.7 kg/h to 1 800 kg/h | 10-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.0 \cdot 10^{-3} \cdot Q_m + 8 \text{ g/h}$ | |
| | 3 kg/h to 4 500 kg/h | 13-fold nozzle system EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.0 \cdot 10^{-3} \cdot Q_m + 4 \text{ g/h}$ | |
| | 0.7 kg/h to 120 kg/h | Low pulsation rotary gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $3.0 \cdot 10^{-3} \cdot Q_m$ | |
| | > 120 kg/h to 1 800 kg/h | | $3.3 \cdot 10^{-3} \cdot Q_m$ | |
| | 140 kg/h to 4 300 kg/h | Turbine gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $3.0 \cdot 10^{-3} \cdot Q_m + 0.35 \text{ kg/h}$ | |

Valid from: 02.05.2024

Date of issue: 02.05.2024

Page 3 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-21444-01-00

Permanent Laboratory

| Calibration and Measurement Capabilities (CMC) | | | | |
|--|-----------------------------|---|--|--|
| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement | Remarks |
| Mass flow rate Q_m or mass of flowing gases | 100 mg/h to 3 g/h | Laminar flow elements EP.LAB.VA.01.01: 2024-01, chap. 2.4.3 | $5.0 \cdot 10^{-3} \cdot Q_m + 2.6 \text{ mg/h}$ | Calibration medium: atmospheric air at room temperature and atmospheric pressure |
| | > 3 g/h to 17 g/h | | $5.0 \cdot 10^{-3} \cdot Q_m + 20 \text{ mg/h}$ | or |
| | > 17 g/h to 0.23 kg/h | | $4.9 \cdot 10^{-3} \cdot Q_m + 0.18 \text{ g/h}$ | compressed air at room temperature up to max. 8 bar overpressure and max. 800 kg/h |
| | > 0.23 kg/h to 0.93 kg/h | | $4.8 \cdot 10^{-3} \cdot Q_m + 1 \text{ g/h}$ | standards at atmospheric pressure ($0.85 \text{ bar} < p_{N,abs} < 1.05 \text{ bar}$) |
| | > 0.93 kg/h to 3.95 kg/h | | $5.6 \cdot 10^{-3} \cdot Q_m + 3.4 \text{ g/h}$ | |
| | > 3.95 kg/h to 7.1 kg/h | | $6.6 \cdot 10^{-3} \cdot Q_m + 6 \text{ g/h}$ | |
| | > 7.1 kg/h to 17.3 kg/h | | $6.0 \cdot 10^{-3} \cdot Q_m + 14 \text{ g/h}$ | |
| | 9.8 kg/h to 310 kg/h | | $7.0 \cdot 10^{-3} \cdot Q_m + 0.6 \text{ kg/h}$ | Test bench for fans with calibration medium: |
| | > 310 kg/h to 2 080 kg/h | | $4.8 \cdot 10^{-3} \cdot Q_m + 2.105 \text{ kg/h}$ | atmospheric air at room temperature and atmospheric pressure |
| | > 2 080 kg/h to 11 500 kg/h | Prandtl probe EP.LAB.VA.01.01: 2024-01, chap. 2.4.4 | $5.0 \cdot 10^{-3} \cdot Q_m + 10 \text{ kg/h}$ | |
| Temperature Quantities Direct reading thermometers and temperature transmitters with resistance sensor | -20 °C to 175 °C | DKD-R 5-1:2018 in thermostatic bath | 0.1 K | Comparison with standard resistance thermometer |
| Humidity Quantities Direct reading measuring devices for relative humidity with capacitive humidity sensors (no psychrometers) | 10 % to 90 % | DKD-R 5-8:2019 in humidity generator air temperature 21 °C to 25 °C | 1.5 % | Comparison with standard thermometer and 2 standard humidity sensors Measurement uncertainty expressed in relative humidity |

On-site Calibration

| Calibration and Measurement Capabilities (CMC) | | | | |
|--|-------------------------|------------------------------------|---|---|
| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement | Remarks |
| Pressure Absolute pressure p_{abs} | 0.15 bar to 2 bar | DKD-R 6-1:2014 | $18 \text{ Pa} + 1 \cdot 10^{-4} \cdot p_{abs}$ | Pressure medium: air or nitrogen |
| | > 2 bar to 10 bar | | $40 \text{ Pa} + 2.3 \cdot 10^{-4} \cdot p_{abs}$ | |
| Differential pressure Δp | 0 bar to 0.06 bar | DKD-R 6-1:2014 | $0.3 \text{ Pa} + 3 \cdot 10^{-4} \cdot \Delta p$ | Pressure medium: air or nitrogen at atmospheric pressure |
| | > 0.06 bar to 0.6 bar | | $3 \text{ Pa} + 3 \cdot 10^{-4} \cdot \Delta p$ | |
| Negative and positive gauge pressure p_e | -0.85 bar to -0.6 bar | DKD-R 6-1:2014 | 29 Pa | Pressure medium: air or nitrogen |
| | > -0.6 bar to -0.06 bar | | $3 \text{ Pa} + 3 \cdot 10^{-4} \cdot p_e$ | |
| | > -0.06 bar to 0.06 bar | | $0.3 \text{ Pa} + 3 \cdot 10^{-4} \cdot p_e$ | |
| | > 0.06 bar to 0.6 bar | | $3 \text{ Pa} + 3 \cdot 10^{-4} \cdot p_e$ | |
| | > 0.6 bar to 10 bar | | $50 \text{ Pa} + 2 \cdot 10^{-4} \cdot p_e$ | |
| | > 10 bar to 30 bar | | $4.3 \text{ mbar} + 2 \cdot 10^{-4} \cdot p_e$ | |
| | > 30 bar to 600 bar | | $42 \text{ mbar} + 2 \cdot 10^{-4} \cdot p_e$ | Pressure medium: oil |

Valid from: 02.05.2024

Date of issue: 02.05.2024

Page 4 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-21444-01-00

On-site Calibration

| Calibration and Measurement Capabilities (CMC) | | | | |
|--|--|--|---|---|
| Measurement quantity / Calibration item | Range | Measurement conditions / procedure | Expanded uncertainty of measurement | Remarks |
| Fluid Quantities Volume flow rate Q_v or volume of flowing gases | 5.7 L/h to 80 m ³ /h | Critically operated nozzles EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $4.07 \cdot 10^{-3} \cdot Q_v + 0.078 \text{ L/h}$ | Calibration medium: atmospheric air at room temperature and atmospheric pressure or compressed air at room temperature up to max. 8 bar overpressure, maximum flow rate within the limits of on-site availability standards at atmospheric pressure ($0.85 \text{ bar} < p_{N,\text{abs}} < 1.05 \text{ bar}$) |
| | 0.6 m ³ /h to 100 m ³ /h | Low pulsation rotary gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $4.56 \cdot 10^{-3} \cdot Q_v$ | |
| | > 100 m ³ /h to 1 600 m ³ /h | | $4.92 \cdot 10^{-3} \cdot Q_v$ | |
| | 130 m ³ /h to 3 750 m ³ /h | Turbine gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $4.62 \cdot 10^{-3} \cdot Q_v + 0.297 \text{ m}^3/\text{h}$ | |
| | 100 mL/h to 3 L/h | Laminar flow elements EP.LAB.VA.01.01: 2024-01, chap. 2.4.3 | $5.5 \cdot 10^{-3} \cdot Q_v + 2.88 \text{ mL/h}$ | |
| | > 3 L/h to 17 L/h | | $6.05 \cdot 10^{-3} \cdot Q_v + 18 \text{ mL/h}$ | |
| | > 17 L/h to 0.16 m ³ /h | | $5.885 \cdot 10^{-3} \cdot Q_v + 0.192 \text{ L/h}$ | |
| | > 0.16 m ³ /h to 0.68 m ³ /h | | $6.49 \cdot 10^{-3} \cdot Q_v + 0.78 \text{ L/h}$ | |
| | > 0.68 m ³ /h to 3 m ³ /h | | $6.71 \cdot 10^{-3} \cdot Q_v + 3.6 \text{ L/h}$ | |
| | > 3 m ³ /h to 5.5 m ³ /h | | $7.7 \cdot 10^{-3} \cdot Q_v + 6.24 \text{ L/h}$ | |
| | > 5.5 m ³ /h to 13 m ³ /h | | $6.6 \cdot 10^{-3} \cdot Q_v + 15 \text{ L/h}$ | |
| Mass flow rate Q_m or mass of flowing gases | 6.3 g/h to 90 kg/h | Critically operated nozzles EP.LAB.VA.01.01: 2024-01, chap. 2.4.1 | $3.3 \cdot 10^{-3} \cdot Q_m + 0.108 \text{ g/h}$ | |
| | 0.7 kg/h to 120 kg/h | Low pulsation rotary gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $3.3 \cdot 10^{-3} \cdot Q_m$ | |
| | > 120 kg/h to 1800 kg/h | | $3.63 \cdot 10^{-3} \cdot Q_m$ | |
| | 140 kg/h to 4 300 kg/h | Turbine gas meter EP.LAB.VA.01.01: 2024-01, chap. 2.4.2 | $3.3 \cdot 10^{-3} \cdot Q_m + 0.42 \text{ kg/h}$ | |
| | 100 mg/h to 3 g/h | Laminar flow elements EP.LAB.VA.01.01: 2024-01, chap. 2.4.3 | $5.5 \cdot 10^{-3} \cdot Q_m + 3.12 \text{ mg/h}$ | |
| | > 3 g/h to 17 g/h | | $5.5 \cdot 10^{-3} \cdot Q_m + 24 \text{ mg/h}$ | |
| | > 17 g/h to 0.23 kg/h | | $5.39 \cdot 10^{-3} \cdot Q_m + 0.216 \text{ g/h}$ | |
| | > 0.23 kg/h to 0.93 kg/h | | $5.28 \cdot 10^{-3} \cdot Q_m + 1.2 \text{ g/h}$ | |
| | > 0.93 kg/h to 3.95 kg/h | | $6.16 \cdot 10^{-3} \cdot Q_m + 4.08 \text{ g/h}$ | |
| | > 3.95 kg/h to 7.1 kg/h | | $7.26 \cdot 10^{-3} \cdot Q_m + 7.2 \text{ g/h}$ | |
| | > 7.1 kg/h to 15 kg/h | | $6.6 \cdot 10^{-3} \cdot Q_m + 16.8 \text{ g/h}$ | |
| Temperature Quantities Direct reading thermometers and temperature transmitters with resistance sensor | -20 °C to 175 °C | DKD-R 5-1:2018 in thermostatic bath | 0.1 K | Comparison with standard resistance thermometer |
| Humidity Quantities Direct reading measuring devices for relative humidity with capacitive humidity sensors (no psychrometers) | 10 % to 90 % | DKD-R 5-8:2019 in humidity generator air temperature 21 °C to 25 °C | 1.5 % | Comparison with standard thermometer and 2 standard humidity sensors Measurement uncertainty expressed in relative humidity |

Valid from: 02.05.2024

Date of issue: 02.05.2024

Page 5 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-21444-01-00

Abbreviations used:

| | |
|-----------|--|
| CMC | Calibration and measurement capabilities |
| DIN | Deutsches Institut für Normung e.V. – German institute for standardization |
| DKD-R | Calibration Guideline of Deutscher Kalibrierdienst (DKD), published by Physikalisch-Technische Bundesanstalt (PTB) |
| EN | Europäische Norm – European Standard |
| IEC | International Electrotechnical Commission |
| ISO | International Organization for Standardisation |
| EP.LAB.VA | In house method of EP EHRLER PRÜFTECHNIK ENGINEERING GMBH |

Valid from: 02.05.2024
Date of issue: 02.05.2024

Page 6 of 6

This document is a translation. The definitive version is the original German annex to the accreditation certificate.