

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-21453-01-00 according to DIN EN ISO/IEC 17025:2018

 Valid from:
 13.12.2021

 Date of issue:
 07.06.2024

Holder of accreditation certificate:

Stotz Feinmesstechnik GmbH Herman-Dreher-Straße 6, 70839 Gerlingen

with the location

Stotz Feinmesstechnik GmbH Kalibrierlaboratorium Herman-Dreher-Straße 6, 70839 Gerlingen

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

Calibrations in the areas:

Dimensional quantities

Length

- Diameter
- Form error

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at https://www.dakks.de.

Abbreviations used: see last page



Annex to the Accreditation Certificate D-K-21453-01-00

Permanent Laboratory

Calibration- and Measurement Capabilities (CMC)

Measurement quantity/ Calibration item	Range			Measurement conditions/procedure	Expanded uncertainty of measurement	Remarks
Length cylindrical standards adjusting rings, adjusting mandrels diameter	2 mm	to	230 mm	VDI/VDE/DGQ 2618, page 4.1:2006	0.9 μm + 13 · 10 ⁻⁶ · <i>d</i>	<i>d</i> = measured diameter
roundness deviation		to	40 μm		0.4 μm	diameter: 2 mm to 230 mm axial length: to 200 mm
straightness deviation of the surface line		to	40 µm		1.2 μm	axial length: to 200 mm
parallelism deviation of the surface lines		to	40 μm		1.3 μm + 1.0 · 10 ⁻⁶ · <i>d</i>	d = measured diameter 2 mm ≤ d ≤ 100 mm axial length: to 200mm
					1.4 μm	100 mm ≤ <i>d</i> ≤ 230 mm axial length: to 200 mm
taper standards and taper gauges diameter	2 mm	to	230 mm	VDI/VDE/DGQ 2618, page 4.12:2007	$1.1 \mu\text{m} + 13 \cdot 10^{-6} \cdot d$	<i>d</i> = measured diameter
diameter in the reference planes					2.4 μm + 31 μm · 10 ⁻³ · m / <i>l</i>	<i>l</i> = distance between the measuring levels in m
taper angle					0.001 ° + 0.13 ° · 10 ⁻³ · m / <i>l</i>	<i>l</i> = distance between the measuring levels in m
roundness deviation		to	40 µm		0.4 + 3 · 10 ⁻⁶ · <i>d</i>	<i>d</i> is the larger diameter
straightness deviation		to	40 µm		1.3 μm	



Annex to the Accreditation Certificate D-K-21453-01-00

Abbreviations used:

- CMC Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
- DIN Deutsches Institut für Normung e.V.
- VDE Verband der Elektrotechnik, Elektronik und Informationstechnik e.V.
- VDI Verein Deutscher Ingenieure e.V.