

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-21865-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 05.12.2022Date of issue: 05.12.2022

Holder of accreditation certificate:

Globalworx GmbH Klammsbosch 10, 77880 Sasbach

The calibration laboratory meets the minimal requirements of DIN EN ISO/IEC 17025:2018 and, if applicable, additional legal and normative requirements, including those in relevant sectoral schemes, in order to carry out the conformity assessment activities listed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

Calibration in the fields:

Thermodynamic quantities
Temperature quantities

- Temperature transmitters, data loggers

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at https://www.dakks.de.



Annex to the Accreditation Certificate D-K-21865-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded measurement of uncertainty	Remarks
Temperature Temperature transmitters and data loggers with resistance sensor	0°C to	25 °C	DKD-R 5-1:2018 In climatic chamber	0.42 K	Comparison with standard resistance thermometers

Abbreviations used:

CMC Calibration and measurement capabilities

DKD-R Calibration Guide of Deutscher Kalibrierdienst (DKD), published by the Physikalisch-

Technischen Bundesanstalt

Valid from: 05.12.2022 Date of issue: 05.12.2022